


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH CHAVAKKAD
3	Name of Equipment with Make, Model and Serial Number	Semi automatic biochemistry analyser Make - Robonik Model - Prietest touch S/N - AT0390412RBK
4	Equipment ID/ Barcode	0820706-#96136
5	Date of purchase/ Year of manufacture/Installation Date	01-06-2012
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	14-02-2024
9	Action taken	Checked the equipment and found contrast board and primary board are defective. Spare Enquired to OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	64760
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report, EOL from OEM attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found contrast board and primary board are defective. Equipment installed on 01/06/2012 aged up to 12 years and 4 months, oem informed that this machine is obsolete model, end of life letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Obsolete model.

Jr. Consultant (Biomedical)
NHM, Thrissur


Signature of JC BM (NHM)



Date


SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHAVAKKAD
Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 189877

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Thgh</u></p> <p>Address <u>Chavakkad</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>8139891595</u></p>	<p>Call Registration Date : <u>14/2/2024</u></p> <p>Caller ID : <u>96136</u></p> <p>Date of Visit : <u>15/2/2024</u></p> <p>Asset No. : <u>0820706</u></p> <p>EQPT Name : <u>Biochemistry analyser</u></p> <p>Manufacture <u>Robonix</u> Model : <u>pretest touch</u></p> <p>S. No. <u>AT0551010R.8</u> Dept. <u>laboratory</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not getting on, display issue.

Action Taken : checked the equipment and found contrast board and primary board are defective. enquired with spare oem

Completed Date : 15/2/24 Time : 9:40am Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>15/2/2024</u>	<u>9:30am</u>	<u>9:40am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Christy</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/2/2024</u></p> <p>Contact Number : <u>6238844347</u></p>	<p>Customer Name : <u>Praveen k p</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Date]</u></p> <p>Contact Number : <u>[Number]</u></p> <p>Designation : <u>Lab</u></p> <p>Hospital Seal : <u>THGH</u></p>
--	--



SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHAVAKKAD
[Signature]
THGH Chavakkad

Date : 27.06.2024

Dear Valued Customer,



Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0390412RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0390412RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**




Authorised Signatory

9:54 01:44

Voil LTE1 Voil 4G LTE2 71%

+919833696980



Date : 24.06.2024

Dear Valued Customer,

EOL Letter = Prietest Touch - AT0540315RBK.pdf

1 page · 89 kB · PDF

9:39 am

27 June 2024

You
Ethinte undo kayil


Pls share eol

12:03 pm

This message was deletea

3:38 pm

Forwarded



Date : 27.06.2024

Dear Valued Customer,

EOL Letter = Prietest Touch - AT0390412RBK.pdf

1 page · 89 kB · PDF

3:38 pm

1 July 2024

Message





Sub-INTENDENT
MALIK HEAD QUARTERS HOSPIT
CHAVAKKAD



MEV
Medical Engineering & Service
Private Limited
12th, North Anna Street,
Chennai - 600 026

CYRIX
CN: THQHCB63 L1 REGS: RT/ANB3102ABK
PRO : Biochemistry Analyzer
ID : 0820706
CAL.ON : 04/12/2023
CAL.DUR : 03/12/2024
ISO 9001:2008 Certified Co. Reg-324455943

CYRIX BIOLOGICAL EQUIPMENT MAINTENANCE
PROGRAMME GROUP
TOLL FREE NO: 1800 - 425-7699
BAR CODE - (8004-890615225)
0820706

CYRIX HEALTH CARE PVT.LTD
BEMP SERVICE PROVIDER
REVENTIVE MAINTENANCE SERVICE
PMS DONE ON: 03/12/2024
NEXT PMS DUE: 04/12/2024
PERFORMED BY: Chinnay



SUPERINTENDENT
TALUK HEAD QUARTERS HOSP
CHAVAKKAD

SUPERINTENDENT
PSHQ HEAD QUARTERS HOSPITAL
CHAVAKKAD



Product BIOCHEMISTRY ANALYSER IVD
Model PRIETEST TOUCH Voltage 115V-230V/ 60Hz-50Hz
Sr. No. AT0390412RBK Power 75VA
CE

Contrast Knob


RS-232
USB



SEMI AUTO BIOCHEMISTRY ANALYZER [ROBORIK]

		Received	Issued	Balance	Remarks
01/06/2012	Received from store	01		01	
	model : Roborik Pre-test Reus				
	SINO : AT0390412 RBK.				
	Above used transferred		from Stock	Reg volu. 3	page no: 29




SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHAVAKKAD