


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	PHC VENKIDANGU
3	Name of Equipment with Make, Model and Serial Number	Semi automatic biochemistry analyser Make - Robonik Model - Prietest touch S/N - AT2450814RBK
4	Equipment ID/ Barcode	0843670 #107273
5	Date of purchase/ Year of manufacture/Installation Date	Asset date: N/A (Actual installation date as per stock :09/03/2015)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	20/04/2024
9	Action taken	Checked the equipment and found contrast board and primary board are defective. Need to replace further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA



13	Asset Value	64760
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and EOL from OEM attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found contrast board and primary board are defective. Equipment installed on 09/03/2015 aged up to 09 years , oem informed that this machine is obsolete model, end of life letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory      #Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*obsolete model*

  
  
 Signature of JC BM (NHM)

Date *22/04/2024*.

  
 Signature of  
 Superintendent/Medical Officer (i/c)



**MEDICAL OFFICER IN CHARGE**  
 FAMILY HEALTH CENTRE  
 VENKITANGU, PADOOR P. O.  
 THRISUR - 680 524





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. :**  
222935

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>phc</u></p> <p>Address ..... <u>Venkitangy</u> <u>Thrissur</u></p> <p>Ph : .....</p>	<p>Call Registration Date : ..... <u>20/4/2024</u></p> <p>Caller ID : ..... <u>107273</u></p> <p>Date of Visit : ..... <u>22/04/2024</u></p> <p>Asset No. : ..... <u>0843670</u></p> <p>EQPT Name : ..... <u>Semi Auto biochemistry Analyser</u></p> <p>Manufacture <u>Robonik</u> Model : ..... <u>pridetest 2000</u></p> <p>S. No. <u>AT24508</u>   HRB Dept. ..... <u>Laboratory</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** display issue and not functioning properly

**Action Taken :** Checked the equipment and found contrast board and primary board are defective. Need to replace further checking.

Completed  Date : 22/4/24 Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>22/4/24</u>		

**Customer Remark** Completed  Pending

**Service Engineer Name :** Christy  
**Signature :** [Signature]  
**Date :** 22/4/24  
**Contact Number :** 6238844342

**Customer Name :** Vibisree PV  
**Signature :** [Signature]  
**Date :** 22-04-2024  
**Contact Number :** 9846850383  
**Designation :** Lab Technician  
**Hospital Seal :**



**MEDICAL OFFICER IN CHARGE**  
FAMILY HEALTH CENTRE  
VENKITANGU, PADOOR P.O.  
THRISUR - 680 524

**Date : 30th April 2024**

Dear Valued Customer,



Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT2450814RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT2450814RBK** as **End of Life (EOL) Notified.** (During from installation is 09 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**



Authorised Signatory



Forwarded




You

Photo

Pls share eol

10:58 am ✓✓

Forwarded



EOL Letter = Prietest Touch - AT2450814RBK.pdf

1 page • 89 kB • PDF

11:35 am

21 May 2024

Forwarded

Message









