



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 170280

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>GH ALAPPUZHA</u></p> <p>Address : <u>DIST: ALAPPUZHA</u></p> <p>Ph : <u>9847189787</u></p>	<p>Call Registration Date : <u>13/7/24</u></p> <p>Caller ID : <u>122932</u></p> <p>Date of Visit : <u>15/7/24</u></p> <p>Asset No. : <u>0410212</u></p> <p>EQPT Name : <u>Sterilized</u></p> <p>Manufacture : <u>NA</u> Model : <u>NA</u></p> <p>S. No. : <u>NA</u> Dept. : <u>Surgery OP</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Water leaking, hearing element is damaged.

Action Taken : checked the machine and found that hearing element and door are damaged. The water is leaking due to body is broken and rust condition wear and tear happened.

Completed Date : 15/7/24 Time : 2:00PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Midhunraj R</u>	<u>15/7/24</u>	<u>1:30PM</u>	<u>2:00PM</u>

Customer Remark Completed Pending
Machine is leaking and damaged condition


<p>Service Engineer Name : <u>Midhunraj R</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/7/24</u></p> <p>Contact Number : <u>9072522056</u></p>	<p>Customer Name : <u>Sathya</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/07/2024</u></p> <p>Contact Number : <u>9847189787</u></p> <p>Designation : <u>Grade II Nursing Officer</u></p> <p>Hospital Seal : <u>GENERAL HOSPITAL</u></p>
--	--

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	GH ALAPPUZHA
3	Name of Equipment with Make, Model and Serial Number	STERILIZER MAKE :- NA MODEL :- NA SN :- NA
4	Equipment ID & Barcode	122932 & 0410212
5	Date of purchase / Year of manufacture /Installation Date	20/02/2015
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	13/07/2024
9	Action taken	Checked the machine and found that heating element and door are damaged. The water is leaking due to body is broken and rust condition. Wear and tear happened.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended for Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS :- 3850/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that heating element and door are damaged. The water is leaking due to body is broken and rust condition. Wear and tear happened. The equipment was installed on 20/02/2015 and aged up to 8 years and 4 months. The machine is fully rusted and irreparable condition. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	MIDHUN RAJ R 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA

 15/12/24
Signature of JCBM (NHM)



Date

Seal


Signature of
Superintendent / Medical Officer (i/c)

Date	Name of items	Receipt	Total	Page No.	Vol	Part	Remarks
31/10/98	Electric sterilizer (old book P.No. 173)		4				
26/1/06	Received from store Page no. 17 Aidcell Volume II	1	5				
17/2005	Price Rs 3240/- (Eycop) Condemned sterilizer CG 14112/2005 Date 17/11/2005						
	As per order No. CG 9689/2008 SMU 14 21- 24.6.2008 condemned.						
14/8/10	Electric Sterilizer (small) page no 205 Cost Rs-950/- (CART OP)	1	3				
20/2/15	Rd from store (Small) S.R.M.C. Page No. 38 Cost. 1650/-	1	4				
24.12.2016	Condemned and Auctioned as per Order No B1 241 / 2013 GH Alappuzha dt 24.12.2016	3	1				

Carried over to p no ³²/~~1175~~ stock Regd
Vol II



For
OP Incharge
Verified
16/12/16
MINIRAS V
SNO
Sandhya
SUPERINTENDENT
GENERAL HOSPITAL
ALAPPUZHA



Handwritten notes in blue ink, oriented vertically and upside down relative to the page. The text includes:
KURIPASU
No. 1-10
Faz
Op Indrag
SUPERKUALITI
GENERAL H...

