



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 230373

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>District TB Centre</u>	Call Registration Date : <u>11-7-24</u>
Address <u>Thiruvananthapuram</u>	Caller ID : <u>122422</u>
<u>Kerala</u>	Date of Visit : <u>24-7-24</u>
Ph : <u>9446437171</u>	Asset No. : <u>011571</u>
	EQPT Name : <u>Centrifuge</u>
	Manufacture <u>Kensi</u> Model : <u>CGS</u>
	S. No. <u>18650</u> Dept. <u>Laboratory</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found Centrifuge motor, internal wiring and RPM controller defective. Need to replace these spares for further checking and working condition of the equipment.

Completed Date : 24/7/24 Time : 4:05 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Soosya Sudheep</u>	<u>24/7/24</u>	<u>3:30 PM</u>	<u>4:05 PM</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Soosya Sudheep</u>	Customer Name : <u>Aswathy</u>
Signature : <u>Soosya</u>	Signature : <u>Aswathy</u>
Date : <u>24/7/24</u>	Date : <u>24/7/24</u>
Contact Number : <u>7907406244</u>	Contact Number : <u>9656048820</u>
	Designation : <u>Lab Technician</u> (NTEP)
	Hospital Seal :




Aswathy
DISTRICT TB OFFICER
District TB Centre
Thiruvananthapuram

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	DISTRICT TB CENTRE THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name: CENTRIFUGE Make : KEMI Model : C6S Serial No : NA
4	Equipment ID/ Barcode	122422 / 0111571
5	Date of purchase/ Year of manufacture/Installation Date	01/05/2015
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11/07/2024
9	Action taken	Checked and found that the Motor defective and internal connection wires defective..Need to replace the spares for further checking and working condition of the equipment.Enquired spares from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	9651/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found that the Centrifuge motor and internal connection wires defective. The unit was installed on 01/05/2015 and covered upto 9+ years. The unit is very old and the spares are not available in the market. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	SOORYA SUDEEP 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Motor and internal connections were defective. Unit is old and spares are not available as per tender condition, this can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


25.07.2024
Signature of JC BM (NHM)

Date: 26/7/24




Signature of
Superintendent/Medical Officer
DISTRICT TB OFFICER
District TB Centre
Thiruvananthapuram (I/c)

18

CENTRIFUGE

0111571

Date of Receipt

5/20/25

units and
details

Received
quantity

01

Carried over

Page No. 88 from Volume No. I.

