


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC THRIPRAYAR
3	Name of Equipment with Make, Model and Serial Number	Nebuliser MAKE: Easycare MODEL: EC-7020 SN: NA
4	Equipment ID/ Barcode	0831435 # 116806
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date -Nil (As per the stock book installation date is 04-01-2015 stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	12-06-2024
9	Action taken	Checked the machine and found that Machine is not switching on. Found motor and piston are defective. need to replace the piston and motor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	1,353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on. Found motor and piston are defective. As per the stock book Equipment installed on 04-01-2015 aged up to 9 years 6 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation
17	Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM.

Nebulizer is about 9 years and quotation not received for this old model

Jr. Consultant (Biomedical 1)
NHM, Thrissur

[Signature]
Signature of JC BM (NHM)

[Signature]
Superintendent/Medical Officer (C)
SUPERINTENDENT
COMMUNITY HEALTH CENTRE
THRIPRAYAR
PO, PERINGOTTUKARA
THRISSUR - 680565

COMMUNITY HEALTH CENTRE, THRISSUR
PO, PERINGOTTUKARA
PIN: 680565
DATE: _____

Date _____



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
222759

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERS Approved Service Agency

Service Report

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : CHC</p> <p>Address : THRIPRAYAR THRISSUR.</p> <p>Ph : 04872397555</p>	<p>Call Registration Date : 12/06/2024</p> <p>Caller ID : 116806</p> <p>Date of Visit : 13/06/2024</p> <p>Asset No. : 0831435</p> <p>EQPT Name : Nebulizer</p> <p>Manufacture : Easycare Model : EC-7020</p> <p>S. No. : N/A. Dept. : OP</p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : equipment not getting on

Action Taken : Checked the machine and found that machine is not switching on. Found motor and piston are defective. Need to replace the piston and motor for further checking

Completed **Date** : 13/6/24 **Time** : 11:12 AM **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Christy	13/6/24	10:00 AM	11:12 AM

Customer Remark Completed Pending

Service Engineer Name : Christy
Signature :
Date : 13/6/24
Contact Number : 6238844349



Customer Name : Smija. C
Signature :
Date : 13/6/2024
Contact Number : 04872397555
Designation : NUR Sija thir
Hospital Seal :



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0831435**



