


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC KAKKAD
3	Name of Equipment with Make, Model and Serial Number	Weighing Machine Make : SAMSO Model : NA SN – NA
4	Equipment ID & Barcode	#120414 - 0841377
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date-20-11-2019 As per stock book installation date- 21-11-2019 (stock book attached for reference)
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01-07-2024
9	Action taken	Checked the machine and found board and sensor are faulty. need to replace sensor and board for further service.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	5000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked equipment found board and sensor damaged, In ticket master equipment installed on 20-11-2019 and As per stock book installation date- 21-11-2019 covered up to 4 years 8 months, Quotation not submitted since spares are not available in the market. So we are recommend equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spares not available



Signature of JC BM (NHM)

Medical Officer in-charge
PHC Kadukutty
P.O. Kattikuzhiatt - 680 308


Superintendent / Medical Officer (i/c)

Signature of seal

Date





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
222872

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>Phc</u></p> <p>Address : <u>Kakkad</u> <u>Cherissur</u></p> <p>Ph : <u>8111858359</u></p>	<p>Call Registration Date : <u>17/24</u></p> <p>Caller ID : <u>120414</u></p> <p>Date of Visit : <u>21/7/24</u></p> <p>Asset No. : <u>0841377</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>SAM&O</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>public health</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : display not turn on

Action Taken : checked the machine and found board and sensors are faulty. need to replace sensor and board for further service.


Completed Date : 21/7/24 Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jinto Joy</u>	<u>21/7/24</u>	<u>11:00am</u>	<u>12:00pm</u>

Customer Remark Completed Pending



Medical Officer In-charge
PHC Kadukutty
P.O, Kathikudam - 680 308

Service Engineer Name : Jinto Joy

Signature : [Signature]

Date : 21/7/24

Contact Number : 7893847125

Customer Name : Shereena P. I.

Signature : [Signature]

Date : 08/07/2024

Contact Number : 9446996235

Designation : JCPHN

Hospital Seal :

