


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC ELAVALLY
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: N/A Model: ZY-102W SN: 2106233411132
4	Equipment ID/ Barcode	0852928 #114200
5	Date of purchase/ Year of manufacture/Installation Date	28-07-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 30-05-2024
9	Action taken	Checked the machine and found that compressor motor, sieve bed and main board defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. Installation date is 28-07-2021. Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


No OEM support for spare



Moon
Signature of JC BM (NHM)



Date:

Signature of 
Superintendent/Medical Officer (i/c)
MEDICAL OFFICER IN CHARGE
GOVT. FAMILY HEALTH CENTRE ELAVALLY
ELAVALLY



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
222720

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>DHC ELAVALLY</u></p> <p>Address <u>ELAVALLY</u> <u>THRISSUR</u></p> <p>Ph : <u>9400164867</u></p>	<p>Call Registration Date : <u>30/5/24</u></p> <p>Caller ID : <u>114200</u></p> <p>Date of Visit : <u>31/05/24</u></p> <p>Asset No. : <u>0852928</u></p> <p>EQPT Name : <u>Oxygen Concentrator</u></p> <p>Manufacture _____ Model : <u>ZY-102W</u></p> <p>S. No. <u>2106233411132</u> Dept. <u>pharmacy</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Oxygen Concentrator not Switching on

Action Taken : Checked the machine and found that Compressor motor, Sieve bed and main board are defective. Need to replace these spares for further checking

Completed Date : 31/5/24 Time : 2:00pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SINTO JOY</u>	<u>31/5/24</u>	<u>12:00pm</u>	<u>2:00pm</u>

Customer Remark _____ Completed Pending

Service Engineer Name : SINTO JOY
Signature : [Signature]
Date : 31/5/24
Contact Number : 7907406353



Customer Name : [Signature]
Signature : [Signature]
Date : 31/5/24
Contact Number : _____
Designation : _____
Hospital Seal : _____

MEDICAL OFFICER IN CHARGE
GOVT: FAMILY HEALTH CENTRE ELAVALLY
ELAVALLY SOUTH, THRISSUR (DT)

FHC, Elavally
 PHC ELAVALLY, CHAVAKKAD
 ELAVALLY SOUTH, THRISSUR - 680511
 0487645117
 phcelavally@gmail.com

Stock Entry (Non Medicines) - MRC Report

Goods Receipt No. : KL00879/LS/00003/21-22 **Received Date** : 11/09/2021
Purchase Order No. : **Order Date** :
Invoice No. : KL008/FD/00428/21-22 **Invoice Date** : 28/07/2021
Discount Amount : 0.00 **Handling/Other Charges** : 0.00
Supplier Name : KMSCL FREE SUPPLY

Receipt No : KL00879/LS/00003/21-22 , Received Date : 11-Sep-2021 , Scheme: General Scheme

Print Date : 20/06/2024.

SL No	Manufacturer Name	Unit Rate	IGST (%) Other (%)	CGST (%) SGST (%)	Model / Serial No. Cont Person / No.	Warranty / AMC Date	Qty	Value
Item Details : E00038 / OXYGEN CONCENTRATOR 10 LTRS , - Unit : -								
1	TATA AUTOCOMP SYSTEM	0.00	0.00	0.00			4	0.00
Total Qty & Value :							4	0.00

Remarks:

Total Value : 0.00
Handling / Other Charges : 0.00
Discount Amount : 0.00
Total Invoice Amount : 0.00

Consignment Supplied by

Consignment Collect by

Generated by Shiny Thomas (PEN No :- 915946)



0852928-0822214- supply details enquiry

4 messages

Bijo T Joy <bijocyrix@gmail.com>

25 June 2024 at 10:22

To: KMSCL <kmsclbmecentral1@gmail.com>, bmetsr2020@gmail.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

kindly provide the supply details of below mentioned equipment.

hospital - phc elavally

equipment- oxygen concentrator

barcode - 0852928,0822214

caller id - 114200,115897

make - NA

model - :ZY-102W

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Bijo T Joy <bijocyrix@gmail.com>

25 June 2024 at 10:39

To: KMSCL <kmsclbmecentral1@gmail.com>, bmetsr2020@gmail.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

With Regards,


BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

WhatsApp Image 2024-06-25 at 10.39.04_800d7f8e.jpg
43K**KMSCL** <kmsclbmecentral1@gmail.com>

25 June 2024 at 11:48

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.

[Quoted text hidden]

--
Regards,**SALU U S****DISTRICT BIOMEDICAL ENGINEER****KMSCL****Contact No: +91 91889 10013****Bijo T Joy** <bijocyrix@gmail.com>

25 June 2024 at 12:35

To: KMSCL <kmsclbmecentral1@gmail.com>

Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Thanks for the information. Details are not available in the hospital stock book.

With Regards,

BILU T JOY

DISTRICT INCHARGE THRISSUR.

CYRIX HEALTHCARE PRIVATE LIMITED

30/641 B. Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia M:-8921870010 | E:- bijocynix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]



