


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH CHAVAKAD
3	Name of Equipment with Make, Model and Serial Number	ESR Analyser Make: Transasia Model: Ves Matic 20 SN: 202203
4	Equipment ID/ Barcode	0820723 # 116018
5	Date of purchase Year of manufacture/Installation Date	Ticket master installation date-27-01-2016 (Installation date in stock book -13-10-2015 stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 08-06-2024
9	Action taken	Checked the machine and found that printer kit and motoreducer cabl.translation are defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Printer replacement kit - 65205/- Motoreducer Cabl. Translation - 27255/- TOTAL - 92460/-

13	Asset Value	84000/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	110.2%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached and oem quotation attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered Printer kit and Motoreducer Cabl Translation are faulty. In Ticket master installation date is 27-01-2016. As per Stock book Installation date is 13-10-2015 and aged up to 8 years and 7 months. service charge is 110.2% of asset value. Both criteria for RBER met. As per tender clause 5.3.14.1 we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Quotation rate high


Consultant (Biomedical)
NHM, Thrissur



Signature of JC BM (NHM)



Date:


SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHAVAKKAD

Signature of
Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
222923

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Thq h</u></p> <p>Address <u>Chavakkad</u> <u>Thiruvissur.</u></p> <p>Ph :</p>	<p>Call Registration Date : <u>8/6/2024</u></p> <p>Caller ID : <u>116018</u></p> <p>Date of Visit : <u>10/6/2024</u></p> <p>Asset No. : <u>0820723</u></p> <p>EQPT Name : <u>ESR Analyser</u></p> <p>Manufacture <u>Transasia</u> Model : <u>Ves mark 20</u></p> <p>S. No. <u>202203</u> Dept. <u>LABORATORY</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : printer and equipment not functioning

Action Taken : Checked the machine and found that printer kit and motor/producer cable translation are defective, need to replace these spares for further checking.

Completed Date : 10/6/2024 Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>10/6/2024</u>		


Customer Remark Completed Pending

<p>Service Engineer Name : <u>Christy</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/6/2024</u></p> <p>Contact Number : <u>623884434</u></p>	 <p>Customer Name : <u>Soumya Mohan?</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/6/2024</u></p> <p>Contact Number : <u>9562212301</u></p> <p>Designation : <u>Lab technician</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--

**TALUK HEAD QUARTERS HOSPITAL
CHAVAKKAD**

E-S-R ANALYZER

13/10/2015				
		I physically founded	one	
13/10/2015	ESR Analyser	Vermatic	20	
	equipment name	:	ESR Analyser	
			Vermatic 20	
	price	:	210436/-	
	Barcode:		(8004 - 890615225)	
			0820723	
	As per tender - recommend the	clause	5.3.14/1	cyrix condemnation



QUOTATION

Bill To: CYRIX HEALTHCARE PVT LTD
 First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala,
 Pincode 682038 , Thrissur , Kerala 680506


Ref. No : 2425QUOTE001382
Date : Jul 2, 2024 12 22 PM
Validity : 30 days
TBM GSTIN : 27AAACTP2038C1ZT

Customer GSTIN No : 32AAFCC2499H2ZM

Ship to, Taluk Head Quarters Hospital,Chavakkad
 Hospital Road , Thrissur , Kerala 680506

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: VESMATIC 20 AUTOMATIC , PRODUCT CODE 10334 CE SR NO 2203

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same is / are given below, for approval.

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount	
1	166094	90278990	PRINTER REPLACEMENT KIT FOR VM 20(R30507750)	1	65205		18	65205	
2	160013	90278990	MOTOREDUCER. CABL.TRASLATION V30/V20 NEW (P30114060)	1	27255		18	27255	
Merchant Name :TRANSASIA BIO-MEDICALS LTD								Total (Rs)	92460
SCAN AND PAY WITH UPI QR Merchant Name: TRANSASIA BIO MEDICALS LTD 								Discount (Rs.) (-)	
								GST (Rs.)	16642.8
Rupees in words: One lakhs Nine thousand One hundred Two rupees and Eight paise only								Grand Total (Rs.)	109102.8

TERMS AND CONDITIONS :

- 1 Payment :** Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of "Transasia Bio-Medicals Ltd" to **Bank A/C no. : 00600330002538 IFSC code: HDFC0000060, UPI ID : transasiabiomedicalsltd@citibank.** You can also pay us through **UPI apps like BHIM/ Google Pay/ Paytm/PhonePe** etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
- 2 GST :** As applicable
- 3 Other Cost / Local duties :** Customer's responsibility
- 4 Delivery :** Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You
 Your faithfully,
For TRANSASIA BIO-MEDICALS LTD



RECEIVED & ACCEPTED
CUSTOMER SIGNATURE & STAMP:
CUSTOMER NAME: _____
CONTACT NO: _____

TRANSASIA BIO-MEDICALS LTD

Head Office: Transasia House B Chandivali Studio Road, Andheri (E), Mumbai 400 072 Tel: (022) 4030 9000 Fax: (022) 2857 3030 Mumbai Toll Free No: 1800 103 8226, WhatsApp No: 847006434 SMS No: 9212433444

www.transasia.co.in

Transasia Bio-Medicals Limited Company

Erba



Re: 116018-0820723 - thqh chavakad - for quotation

1 message

Tue, 2 Jul, 2024 at 12:42 pm

SANDEEP DAS M <s.das@transasia.co.in>
To: Bijo T Joy <bijocyrix@gmail.com>

Dear Mr. Bijo ,

Please find the attachment of Ves 20 Spare quote of THQH Chavakkad

From: Punam Nimbarte <n.punam@transasia.co.in>
Sent: Saturday, June 29, 2024 11:34 AM
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: transserve <transserve@transasia.co.in>
Subject: FW: 116018-0820723 - thqh chavakad - for quotation

Fyi..

Punam N

Executive – Customer Care

Transasia Bio-Medicals Pvt. Ltd.
Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai – 400072

From: Bijo T Joy [mailto:bijocyrix@gmail.com]
Sent: 27 June 2024 12:26
To: Punam Nimbarte
Cc: Anjitha T L; Sarang KM; Blesson Jose; transserve@transasia.co.in
Subject: Re: 116018-0820723 - thqh chavakad - for quotation

reminder

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com |

On Tue, 25 Jun 2024 at 15:23, Bijo T Joy <bijocyrix@gmail.com> wrote:

reminder

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | *

On Thu, 20 Jun 2024 at 13:18, Bijo T Joy <bijocyrix@gmail.com> wrote:

sn.no-2203

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | *

On Wed, 19 Jun 2024 at 15:01, Punam Nimbarte <n.punam@transasia.co.in> wrote:

Dear Customer,

Hope this email finds you well.

As Per Discussed-8921870010 ,please share the correct serial no.

Our team will review your concern and provide you with the resolution.

Appreciate your patience and continued support.

In case, if you wish to add more information, please feel free to respond to this email or connect us at

Toll Free: 1800 103 8226 / Whatsapp: 8451048434 / transserve@transasia.co.in

Punam N

Executive – Customer Care

Transasia Bio-Medicals Pvt. Ltd.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai – 400072

From: Punam Nimbarte [<mailto:n.punam@transasia.co.in>]

Sent: 19 June 2024 14:38

To: 'Bijo T Joy'

Cc: 'Anjitha T L'; 'Sarang KM'; 'Blesson Jose'; 'transserve@transasia.co.in'

Subject: RE: 116018-0820723 - thqh chavakad - for quotation

Dear Customer,

Hope this email finds you well.

This is to keep you informed that we acknowledge and need your cooperation in resolving your complaint at the earliest,

Please share the correct serial number of the instrument.

Our team will review your concern and provide you with the resolution.

Appreciate your patience and continued support.

In case, if you wish to add more information, please feel free to respond to this email or connect us at

Toll Free: 1800 103 8226 / Whatsapp: 8451048434 / transserve@transasia.co.in

Punam N

Executive – Customer Care

Transasia Bio-Medicals Pvt. Ltd.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai – 400072

From: Bijo T Joy [<mailto:bijocyrix@gmail.com>]

Sent: 19 June 2024 14:16

To: transserve@transasia.co.in

Cc: Anjitha T L; Sarang KM; Blesson Jose

Subject: Re: 116018-0820723 - thqh chavakad - for quotation

reminder

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

On Sat, 15 Jun 2024 at 15:16, Bijo T Joy <bijocyrix@gmail.com> wrote:

dear sir,

kindly provide the quotation for ESR analyzer

equipment - ESR analyzer

quotation for -motor,printer

make - transasia

model - Ves Matic 20

hospital - thqh chavakad

sr.no - 202203

ticket no- 116018

barcode - 0820723

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

