


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	DH KANHNGAD
3	Name of Equipment with Make, Model and Serial Number	BLOOD COLLECTION MONITOR MAKE: HICARE MODEL: HCM-14 S/N: 13033310
4	Equipment ID & Barcode	97593 & 1410212
5	Date of purchase / Year of manufacture / Installation Date	03-05-2013
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email / Toll free)	22-02-2024 (Toll free)
9	Action taken	Checked and found the main board defective. Need to replace main board for further checking of the complaint. Quotation enquired.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendation.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	79500
14	*Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/CYRIX(Attached or Not)	CYRIX SERVICE REPORT AND OEM MAIL PRINTOUT ATTACHED.
16	Reasons for recommending the equipment as BER	Checked and found main board defective. Equipment installed on 03-05-2013 and aged up to 10 years 9 month, life span of the equipment used. As per OEM this model is obsolete, spare not available. So we recommending the unit for condemnation as per tender clause 5.3.14.2.
17	Name & Signature of CYRIX Authority	ABHINAV O 

*Not mandatory #Based on the period of life and value as per the BER guidelines

*Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked the machine and found main board defective. Equipment installed on 03/05/2013. As per OEM mail reply this model has been discontinued and spares are not available. OEM mail printout attached. Recommending the equipment for BER.


Signature of CONSULTANT
Biomedical Engineering
Aranyakeralam (NHM)
1st Kanhangad

Date _____

 Seal


Signature of Superintendent/Medical Officer (i/c)




BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 186120

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>District Hospital</u></p> <p>Address <u>Kannangad</u> <u>Kasaragod</u></p> <p>Ph : <u>9747141402</u></p>	<p>Call Registration Date : <u>22.02.2024</u></p> <p>Caller ID : <u>97593</u></p> <p>Date of Visit : <u>23.02.2024</u></p> <p>Asset No. : <u>1410212</u></p> <p>EQPT Name : <u>Blood Collection Monitor</u></p> <p>Manufacture <u>HICARE</u> Model : <u>HCM-14</u></p> <p>S. No. <u>1303310</u> Dept. <u>Blood Bank</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working.

Action Taken : Checked the machine and found problem with machine main board. Need to replace the main board, its unable to repair locally. waiting for the spare quotation from the OEM.

Completed Date : 28.02.24 Time : 4.50pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abhinav</u>	<u>28.02.24</u>	<u>4.10pm</u>	<u>4.50pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Abhinav</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>28.02.2024</u></p> <p>Contact Number : <u>8943392989</u></p>	<p>Customer Name : <u>Arun Babu</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>28/2/2024</u></p> <p>Contact Number : <u>04672204333</u></p> <p>Designation :</p> <p>Hospital Seal :</p>
---	---

GOVT. DISTRICT HOSPITAL, KANHANGAD.

STOCK REGISTER OF INSTRUMENTS & EQUIPMENTS

Name of Article: **BLOOD COLLECTION MONITOR**

Date	Particulars of Receipt / Issue	Receipt	Issue	Balance	Cost	Initials	Name of Receiver	Designation	Signature of Receiver	Sub Stock Page No.
01-01-08	BF Page 83 Book Vol II	20304		1	75000					
13-02-09	R th from Dil Stone (CREM Blood Collection Monitor Model BCM-10)	1		2	71242		Ajith Kumar	BST Vc		22
16-05-11	R th from Dil Stone (Jove Automation B.C.M Model Jd-91, S.No. J10447)	1		3	41187/50		Ajith Kumar	BST Vc		
10/4/12	Condemned as per order DHO (H) Kijel dtd 4/4/12 (HL R th no 2013/04)		01	02						
03/5/13	R th from Dil Stone (SAC) HCM-14 (Hicare)	Supply								
	Sl no: 13033309 & 10	2		04	1,80,952/200		Ajith Kumar	BST IIC		
30/04/16	Received from Dil Stone (MSCL) Bismarck Model No D-601 Serial No: 201823080 Cost: 95,960/- Warranty 3Yr	1		05	95,960/-		Midhuna SR	LT Blood Bank		22
12/05/16	Condemned as per the order no 02-4563/16/Supdt. DH Kanhangad dated 12/5/16 [CREM Blood Collection Monitor, Midhuna]		1	04			Midhuna SR	LT Blood Bank		
7/13/2022	Transferred to unusable equipment Repulin (as per Cyax report) page no 8 Blood collection monitor Hicare. Sl no. 13033309.		01	03			1	LT Blood Bank		



Handwritten notes in blue ink at the bottom right of the page, including the name 'Midhuna SR' and some illegible text.



Handwritten signature
 സർവ്വകലാശാല
 ജില്ലാ ആരോഗ്യ
 കാണെക്കട
 കാസർഗോഡ്



Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY - Barcode 1410212 - Blood collection monitor - DH Kanhangad

2 messages

Abhinav O <abhinavcyrix@gmail.com>

Mon, Feb 26, 2024 at 1:15 PM

To: "customercarebct@gmail.com" <customercarebct@gmail.com>

Cc: Jayesh Krishnan <jayeshcyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir,

Please share the price and spare availability of Main board for Blood collection monitor. Machine details mentioning below.

Make : Hicare

Model : HCM - 14

SN : 1303310

Thanks and regards,

Abhinav O

Cyrix Healthcare pvt ltd

Kasaragod | Kerala

Mob :8943392989

Sreekumar <customercarebct@gmail.com>

Wed, Feb 28, 2024 at 10:40 AM

To: Abhinav O <abhinavcyrix@gmail.com>

Cc: Jayesh Krishnan <jayeshcyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>

Since this model has been discontinued, the spare for this blood collection monitor is not available.

Rgds,

Sreekumar T G

Senior Territory Officer (BCT)

HLL Lifecare Ltd,

Trivandrum

Mob: 7907856331

[Quoted text hidden]