


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	CHC PERINJANAM
3	Name of Equipment with Make, Model and Serial Number	ECG Machine Make : Edan Model : SE-1200 SN - SE120022209C0808D
4	Equipment ID & Barcode	#101213 - 0831807
5	Date of purchase / Year of manufacture /Installation Date	02-02-10
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-03-24
9	Action taken	Checked the machine and found main board ,battery and display are complaint. Need to replace the main board, battery and display for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	62000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and oem EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board, battery and display are complaint. Equipment installed on 02-02-10 and aged up to 14 YEARS and 3 months. oem informed that this machine is discontinued model and spare not available, oem letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


*Very old machine. EOL received.*



Date: .....




Signature of JC BM (NHM)  
Sr. Consultant (Biomedical)  
Thiruvananthapuram



**SUPERINTENDENT**  
Community Health Centre  
Perinjanam - 680 686

Superintendent / Medical Officer (I/c)  
Community Health Centre  
Perinjanam-680686



Signature of sea

*6/4/24*

Date



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. : 195058**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>CHC</u> .....</p> <p>Address ..... <u>Perinjaram</u> .....</p> <p>                  <u>Thrisser</u> .....</p> <p>Ph : ..... <u>9495085134</u> .....</p>	<p>Call Registration Date : ..... <u>14-03-2024</u> .....</p> <p>Caller ID : ..... <u>101213</u> .....</p> <p>Date of Visit : ..... <u>15-03-2024</u> .....</p> <p>Asset No. : ..... <u>0831807</u> .....</p> <p>EQPT Name : ..... <u>Eco machine</u> .....</p> <p>Manufacture ..... <u>Edan</u> ..... Model : <u>SE 1200</u></p> <p>S. No. <u>SG12002209.co</u> Dept. <u>Casualty</u></p>
---	--

**Service Classification : Breakdown Call**  **PMS**  **Calibration**  **Cust.Training**

**Problem Identified :** ..... Display issue, automatically turning off. .....

**Action Taken :** ..... checked the main and control main board, battery and display case complaint. Need to replace the main board, battery and display for further checking and spare required to O.Gm. .....

Completed  Date : 15-03-2024 Time : 11:00 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Bijo T Joy</u>	<u>15-03-2024</u>	<u>10:00 AM</u>	<u>11 AM</u>

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> <u>Bijo T Joy</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>15-03-2024</u></p> <p><b>Contact Number :</b> <u>8921890010</u></p>	<p><b>Customer Name :</b> <u>Phil George</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>15-03-2024</u></p> <p><b>Contact Number :</b> <u>9495085134</u></p> <p><b>Designation :</b> <u>SNO</u></p> <p><b>Hospital Seal :</b></p>
--	--



**SUPERINTENDENT**  
Community Health Centre  
Perinjaram - 680 686

**ASSISTANT SURGEON**  
Community Health Centre  
Perinjaram - 680686

28.03.2024

To,

M/s Cyrix Healthcare

First Floor, Petta Junction 30/641 B

Poonithura, Ernakulam- 682038

Sub: Reg. End of Life of ECG Machine

Dear Sir/Madam,

We would like to inform you that the equipment as per the below details is very old and the spare parts are not available. Hence we are unable to repair the equipment.

Equipment: ECG Machine

Model: SE1200

SN: SE120022209C0808D

Make: Edan Instruments Inc.

Thanking you

Sincerely



Anoop N

Manager

Sree Gokulam Healthcare Pvt Ltd



## For End of Life Letter

2 messages

Anjitha T L <purchasecyrix2@gmail.com>

Thu, Mar 28, 2024 at 2:21 PM

To: Anoop N <purchase@gokulamhealthcare.com>, service gokulamhealthcare <service@gokulamhealthcare.com>  
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Sarang KM <sarangcyrix@gmail.com>

Dear sir,

Kindly share EOL letter for EDAN ECG Machine

Hospital - CHC perinjanam  
Manufacture - Edan  
Model- SE-1200  
Sr.no- SE120022209C0808D  
Barcode - 0831807  
Caller id - 101213  
Installation date- 02-02-2010

Hospital - THQH OTTAPALAM  
Manufacture - Edan  
Model- SE1200  
Sr.no- SE120022209C0784D  
Installation date- 12.08.2014

--

Thanks & Regards

Anjitha T L  
Purchase Assistant  
PH:- 9947719111  
[purchasecyrix2@gmail.com](mailto:purchasecyrix2@gmail.com)  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Cyrix Healthcare Pvt Ltd**  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

service gokulamhealthcare <service@gokulamhealthcare.com>

Thu, Mar 28, 2024 at 4:09 PM


To: Anjitha T L <purchasecyrix2@gmail.com>  
Cc: "Cyrix Purchase Mgr." <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Sarang KM <sarangcyrix@gmail.com>


Dear Sir,

Kindly see the attached file

[Quoted text hidden]

### 2 attachments

 **IMG\_0001.pdf**  
38K

 **IMG\_0002.pdf**  
38K

# E.C.G. Machine.

70

		Rd	Id	Balance
29/1/10	Rd 1 ECG Machine 21/1/10. Sree Calcuttam HealthCare	1		1
20/8	2020 Received one finger tip pulse oximeter	1		2
23/4	2021 Received from the store	2		4
	working entry.			
22/3	2022 - C/O to the new equipment stock book Vol. 4 P. No. 31.			



