


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH CHALAKUDY
3	Name of Equipment with Make, Model and Serial Number	UPS Make: onlineups, bpe Model: NA SN: NA
4	Equipment ID/ Barcode	0820028 - #78143
5	Date of purchase/ Year of manufacture/Installation Date	30-05-2019 (stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 03-11-2023
9	Action taken	Checked the equipment and found that the control PCB, battery, transformer are defective. Need do replace these spare for further service.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	12-230 inverter PCB – 1298/- 12v-7 ah exide battery – 1139/- Transformer – 649 TOTAL – 3086/-


13	Asset Value	4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68.58%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that the control PCB, battery, transformer are defective, as per stock book Equipment installed on 30/05/19 and aged up to 4years and 5 months, Spare cost is 68.58% of asset value and both criteria met. As per tender clause 5.3.14. Ithis equipment recommend for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**


* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

spare cost is vely high


Signature of JC BM (NHM)

Date:




Superintendent
THQ Hospital Chalakudy
Signature of
Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 222889

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrlix.com | E-mail : bemp.kl@cyrlix.in

<p>Health Facility <u>THAH</u></p> <p>Address <u>Chalakkudy</u> <u>Thaissuvu</u></p> <p>Ph :</p>	<p>Call Registration Date : <u>3/11/23</u></p> <p>Caller ID : <u>78143</u></p> <p>Date of Visit : <u>4/11/23</u></p> <p>Asset No. : <u>0820028</u></p> <p>EQPT Name : <u>ups</u></p> <p>Manufacture <u>online ups</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>lab</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not switching on

Action Taken : The machine checked and found that Control PCB, battery and transformer are defective. need to replace of further service.

Completed Date : 4/11/23 Time : 3:30PM Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrlix Engineer	Date	Start Time	End Time
<u>Jinto</u>	<u>4/11/23</u>	<u>2:00PM</u>	<u>3:30PM</u>

Customer Remark Completed Pending

Service Engineer Name : Jinto Joy
Signature : [Signature]
Date : 4/11/23
Contact Number : 7593847125



Customer Name :

Signature : [Signature]

Date :

Contact Number :

Designation :

Hospital Seal :

Superintendent
THO Hospital Chalakkudy

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 27/01/2024

Contact : 8921950217

Email : backups@ionbackups.com

GSTIN: 32CGUPM9450L12N

600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12 -230 inverter PCB	1	1100.00	1100.00	18	1,298.00
2	7Ah Better y	1	890	890	28	1,139.00
3	12v transformer	1	550	550	18	649.00
PAYABLE AMOUNT						3086.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares, 28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN





barcode - 0820028 caller id - 78143-THQH CHALAKUDY

2 messages

Bijo T Joy <bijocyrix@gmail.com>
To: aionbackups12@gmail.com

Mon, 29 Jul, 2024 at 9:29 pm

dear sir,
it is informed that ups complaint was reported at THQH CHALAKUDY hospital. kindly attend and share the quotation
HOSPITAL - THQH CHALAKUDY
equipment - ups
barcode - 0820028
caller id - 78143
contact no - 8921870010

With Regards,

BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Bhagath Mohan <aionbackups12@gmail.com>
To: Bijo T Joy <bijocyrix@gmail.com>

Tue, 30 Jul, 2024 at 12:29 pm

Thanks & Regards

Bhagath Mohan
PH:-8921950217

[Quoted text hidden]

