


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH PUTHUKKAD
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: NA Model: ZY-10ZW SN: 2106233421046
4	Equipment ID/ Barcode	0822214- #115897
5	Date of purchase Year of manufacture/Installation Date	Ticket master installation date -N/A (As per hospital stock ledger equipment received on 27-08-2021-donated item)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 07-06-2024
9	Action taken	Checked the machine and found that compressor motor, sieve bed and main board are defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. As per hospital stock ledger equipment received on 27-08-2021. Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

No spare details available


 Jr. Consultant (Biomedical)
 NHM, Thrissur

Signature of JC BM (NHM)

Date:




 Signature of
 Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 186680

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kf@cyrix.in

<p>Health Facility <u>JHAH Puthurkad</u></p> <p>Address <u>Thississu</u></p> <p>Ph : _____</p>	<p>Call Registration Date : <u>7/6/2024</u></p> <p>Caller ID : <u>115897</u></p> <p>Date of Visit : <u>8/6/2024</u></p> <p>Asset No. : <u>0822214</u></p> <p>EQPT Name : <u>Oxygen Concentrator</u></p> <p>Manufacture <u>NA</u> Model : <u>ZY-10ZW</u></p> <p>S. No. <u>2106233421046</u> Dept. <u>Pain & Palliative</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : oxygen concentrator not working

Action Taken : Checked the machine and found that compressor motor, slave bed and main board are defective, need to replace these spares for further checking.

Completed Date : 8/6/2024 Time : _____ Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrrix Engineer	Date	Start Time	End Time

Customer Remark _____ Completed Pending

<p>Service Engineer Name : <u>Suha</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>8/6/2024</u></p> <p>Contact Number : <u>9821890010</u></p>	<p>Customer Name : <u>Rinimol Davis</u></p> <p>Signature : _____</p> <p>Date : _____</p> <p>Contact Number : <u>9961152546</u></p> <p>Designation : <u>Secondary Palliative Nurse</u></p> <p>Hospital Seal : _____</p>
--	--



8/6/2024
 9961152546
 Secondary Palliative Nurse



0822214-115897-thqh pudukad- oxygen concentrator

3 messages

Bijo T Joy <bijocyrix@gmail.com>

Sat, 29 Jun, 2024 at 12:46 pm

To: KMSCL <kmsclbmecentral1@gmail.com>, bmetsr2020@gmail.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

kindly provide below supply details of below mentioned oxygen concentrator.

hospital - thqh pudukad

equipment - oxygen concentrator

make - NA

model- ZY-10ZW

barcode - 0822214

caller id - 115897

SN- 2106233421046

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

KMSCL <kmsclbmecentral1@gmail.com>

Sat, 29 Jun, 2024 at 12:56 pm

To: Bijo T Joy <bijocyrix@gmail.com>

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.

[Quoted text hidden]

-

Regards,

SALU U S

DISTRICT BIOMEDICAL ENGINEER

KMSCL

Contact No: +91 91889 10013

Bijo T Joy <bijocyrix@gmail.com>

Thu, 1 Aug, 2024 at 4:31 pm

To: KMSCL <kmsclbmecentral1@gmail.com>

dear mam,

We checked the stock details at the hospital. As this machine has a donated item, we could not find the supply details at the hospital.

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]



ಇಸುಲುವುಲಿ ಸುಲಾ ಲುಯಿ

To Date : 31-Jul-2024

From Date : 01-Jan-2019
Item Name : OXYGEN CONCENTRATOR 10 LTRS - - - (E00038)

Sl. No.	Transaction Date	Issue / Receipt No.	WH Issue No./Date	Particulars	Batch.No	Mfg.Date	Exp.Date	Manufacturer/Supplier	Unit Rate	Received Qty	Issued Qty	Balance Qty
1	19-07-2021	KL00804/SP/00010/21-22	/	Received from Stock Entry	K5BW MODEL	07-07-2021	06-05-2026	district zoology teachers association	79,000.0000	1	0	
2	20-07-2021	KL00804/SP/00330/21-22		Issued to Ward : INCHARGE - CASUALTY	K5BW MODEL	07-07-2021	06-05-2026	district zoology teachers association	79,000.0000	0	1	
3	05-08-2021	KL00804/SP/00013/21-22	/	Received from Stock Entry	0	01-07-2021	31-12-2031	kmsci free supply	0.0000	5	0	
4	18-08-2021	KL00804/RW/00007/21-22	/	Return from Ward	0	01-07-2021	31-12-2031	kmsci free supply	0.0000	3	0	
5	18-08-2021	KL00804/SP/00451/21-22		Issued to Ward : INCHARGE - PHARMACY	0	01-07-2021	31-12-2031	kmsci free supply	0.0000	0	3	
6	20-08-2021	KL00804/SP/00486/21-22		Issued to Ward : INCHARGE - MALE WARD	0	01-07-2021	31-12-2031	kmsci free supply	0.0000	0	3	
7	27-08-2021	KL00804/SP/00439/21-22		Issued to Ward : INCHARGE - CASUALTY	0	01-07-2021	31-12-2031	kmsci free supply	0.0000	0	2	

Print Date: 31-07-2024

Print Time: 01:14:29 PM

