


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	CHC PUTHENCHIRA
3	Name of Equipment with Make, Model and Serial Number	UPS Make: APC Model: BX600 SN: E21637012452
4	Equipment ID/ Barcode	0830815 #117553
5	Date of purchase/ Year of manufacture/Installation Date	As per ticket master installation date is 28-01-2021 (Actual installation date is 05-12-2016 , stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 17-06-2024
9	Action taken	Checked the equipment and found that the control PCB, battery, transformer are defective. Need do replace these spare for further service.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	12-230 inverter PCB – 1298/- 12v-7 ah exide battery – 1139/- Transformer – 649/- <b>TOTAL – 3086/-</b>

13	Asset Value	4500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68.58%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that the control PCB, battery, transformer are defective, as per stock book Equipment installed on 05/12/16 and aged up to 7 years and 6 months, repairing cost is 68.58% of asset value and both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Quotation rate is high*

  
**Jr. Consultant (Biomedical)**  
 NHM

Signature of JC BM (NHM)

Date:

*18/7/24*



*Signature of*  
**Superintendent/Medical Officer (i/c)**

Pin- 680 682, puthenchirachc@gmail.com





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 226160

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p><b>Health Facility</b> ... C.H.C. Puthenchira</p> <p><b>Address</b> ... Puthenchira</p> <p><b>Ph :</b> .....</p>	<p><b>Call Registration Date :</b> 17/6/2024</p> <p><b>Caller ID :</b> 117553</p> <p><b>Date of Visit :</b> 19/6/24</p> <p><b>Asset No. :</b> 083.0815</p> <p><b>EQPT Name :</b> UPS</p> <p><b>Manufacture</b> ... APC <b>Model :</b> BX600</p> <p><b>S. No.</b> F216376245L <b>Dept.</b> Lab</p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Equipment not switching on

**Action Taken :** Checked the equipment and found that control PCB, Battery, Transformer are defective. Need to replace these spares for further checking

**Completed**  **Date :** 19/6/24 **Time :** ..... **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Ramdas	19/6/24	10:00	11:20 AM

**Customer Remark** Completed  Pending   
 ഏറ്റവും നല്ല സേവനം നൽകിയിരിക്കുന്നു. സഹായം നൽകിയിരിക്കുന്നു. പൂർണ്ണമായി.  
 നാമുടമസ്ഥർ: ആരോഗ്യ കേന്ദ്രം, പുത്തൻചിറ  
 Pin- 680 682, puthenchirachc@gmail.com

**Service Engineer Name :** Ramdas  
**Signature :** [Signature]  
**Date :** 19/6/24  
**Contact Number :** 974520163

**Customer Name :** Kajifha P.N.  
**Signature :** [Signature]  
**Date :** [Date]  
**Contact Number :** 949674 5834  
**Designation :** Lab Technician GR I  
**Hospital Seal :** [Seal]

[Signature]  
19/7/24



## AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 02/07/2024

Contact: 8921950217

Email: [aionbackups12@gmail.com](mailto:aionbackups12@gmail.com)

GSTIN: 32CGUPM9450L1ZN

### 600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT. LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12-230 INVERTER PCB	1	1100.00	1100.00	18	1298.00
2	12v-7Ah EXIDE BATTERY	1	890.00	890.00	28	1139.00
3	TRANSFORMER	1	550.00	550.00	18	649.00
<b>PAYABLE AMOUNT</b>						<b>3086.00</b>

#### Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN







---

**CHC PUTHENCHIRA - barcode - 0830815 caller id - 117553**

2 messages

---

**Bijo T Joy** <bijocyrix@gmail.com>  
To: aionbackups12@gmail.com

Mon, 29 Jul, 2024 at 9:25 pm

dear sir,

it is inform you that ups complaint reported at CHC PUTHENCHIRA hospital. kindly attend and share the quotation

HOSPITAL - CHC PUTHENCHIRA

equipment - ups

barcode - 0830815

caller id - 117553

contact no - 8921870010

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

---

**Bhagath Mohan** <aionbackups12@gmail.com>  
To: Bijo T Joy <bijocyrix@gmail.com>

Tue, 30 Jul, 2024 at 1:10 pm

Received your mail.

**Thanks & Regards**

**Bhagath Mohan**  
**PH:-8921950217**

[Quoted text hidden]





