


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC CHERPU
3	Name of Equipment with Make, Model and Serial Number	Equipment: Refrigerator Make – LG Model - GL-195NH5 S/N - NA
4	Equipment ID/ Barcode	#118092 – 0830638
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - NA. (As per stock book installation date is 26/02/2011, stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	19-06-2024
9	Action taken	Checked the refrigerator and found refrigerator not cooling. found that compressor, power board, gasket, timer and thermostat faulty. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	Door gasket -2,700/- Compressor & gas charging - 7,500/- Timer - 850 /- Thermostat - 750/- Power board - 1500/- TOTAL - 13,300/-

13	Asset Value	19,990/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.55%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and Quotation attached
16	Reasons for recommending the equipment as BER	Checked the refrigerator and found refrigerator not cooling. found that compressor, power board, gasket, timer and thermostat faulty. As per stock book equipment installed on 26-02-2011 and aged up to 13 years 03 month, Repair cost is 66.55% both criteria for RBER met. As per tender clause 5.3.14.1 we are recommending equipment for condemnation.
17	Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Spare cost is high.

Jr. Consultant (Biomedical)
NHM, Thr


Signature of JC BM (NHM)

Date




Signature of Superintendent/Medical Officer (I/c)
SUPERINTENDENT
COMMUNITY HEALTH CENTRE
CHERPU-680 561



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 206242

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kt@cyrix.in

Health Facility : <u>GHC</u>	Call Registration Date : <u>19/06/2024</u>
Address : <u>CHERPU</u>	Caller ID : <u>118092</u>
<u>THRISSUR</u>	Date of Visit : <u>19/06/2024</u>
Ph :	Asset No. : <u>0830638</u>
	EQPT Name : <u>REFRIGERATOR</u>
	Manufacture : <u>LG</u> Model : <u>GL-195NHS</u>
	S. No. : <u>NA</u> Dept. : <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT GETTING ON

Action Taken : CHECKED THE EQUIPMENT & FOUND REFRIGERATOR NOT COOLING & FOUND THAT COMPRESSOR, POWER BOARD GASKET, THERMIST & THERMOSTAT FAULTY. NEED TO REPLACE THIS FOR FURTHER CHECKING.

Completed Date : 19/06/24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMBADY V S</u>	<u>19/06/24</u>	<u>11:00am</u>	<u>12:00pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>AMBADY V S</u>	Customer Name :
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>19/06/24</u>	Date :
Contact Number : <u>7907996506</u>	Contact Number :
	Designation :
	Hospital Seal :

SUPERINTENDENT
COMMUNITY HEALTH CENTRE
CHERPU-680 581

Cool House

Deto
Opposite SBI BANK, Kodakara Thrissur
☎ 9895171425

Quotation

To,
Cyrix Healthcare Pvt Ltd
Kochi, Ernaakulam
☎ 9847299500

Quotation# Quote-459
Date: 17-07-2024

Dear Sir/Mam,

Thank you for your valuable inquiry. We are pleased to quote as below:


#	DESCRIPTION	QTY	PRICE	TOTAL
1	Compressor with Gas charging	1 Nos	₹ 7,500.00	₹ 7,500.00
2	Timer	1 Nos	₹ 850.00	₹ 850.00
3	Thermostat	1 Nos	₹ 750.00	₹ 750.00
4	Power Board	1	₹ 1,500.00	₹ 1,500.00
5	Door Gasket	1 Nos	₹ 2,700.00	₹ 2,700.00
GRAND TOTAL				₹ 13,300.00

We hope you find our offer to be in line with your requirement.

For, COOL HOUSE

AUTHORIZED SIGNATURE
Kodakara - 680 684
Refrigeration & Air Conditioning
Cool House

BIOMEDICAL EQUIPMENT MAINTENANCE
 PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
0830638



CYRIX
 HEALTHCARE
 CN: CMC005 N
 PRO: Respiration
 ID: 02048
 CAL ON: 04/09/13
 CAL DUE: 01/09/14
 (800) 8601 7669 Certified Co. Ph. 9744555073

