


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC THIRUVILWAMALA
3	Name of Equipment with Make, Model and Serial Number	Weighing Machine Make : SAMSO Model : NA Sr.No- NA
4	Equipment ID & Barcode	88437 # 0831971
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date – NA (As per stock book installation date is 15-02-2020 stock book attached for your reference)
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	02-01-2024
9	Action taken	Checked the machine and found board and sensor are faulty. Need to replace the board and sensor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	1600/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report
16	Reasons for recommending the equipment as BER	Checked equipment found board and sensor damaged, equipment installed on 15-02-2020 and covered up to 3 years 11 months, Quotation not submitted since spares are not available in the market. and we recommend equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

weighing machine's board and sensor defective.

*Jr. Consultant Biomed
NHM, Thrissur*

[Signature]
Signature of JC BM (NHM)

Superintendent / Medical Officer (i/c)	Signature of seal	Date
<i>[Signature]</i> SUPERINTENDENT COMMUNITY HEALTH CENTRE THIRUVILWAMALA THRISSUR - 680 588		



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 225116

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC Thiruvilwamala</u></p> <p>Address <u>Thiruvilwamala,</u> <u>Thrissur</u></p> <p>Ph : <u>04884282002</u></p>	<p>Call Registration Date : <u>02/01/2024</u></p> <p>Caller ID : <u>88437</u></p> <p>Date of Visit : <u>03/01/2024</u></p> <p>Asset No. : <u>0831971</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>Sanso</u> Model : <u>Na</u></p> <p>S. No. <u>Na</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment is not working

Action Taken : Checked the equipment and found board and sensor are faulty. Need to replace the board and sensor for further checking.


Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin AR</u>	<u>31/1/2024</u>	<u>10:00am</u>	<u>11:00am</u>

Customer Remark Completed Pending



Service Engineer Name : Aswin AR

Signature :

Date : 31/1/2024

Contact Number : 8589965265

Customer Name :

Signature :

Date :


Contact Number :


Designation : **SUPERINTENDENT**

Hospital Seal : **COMMUNITY HEALTH CENTRE**
THIRUVILWAMALA
THRISUR - 680 588

WEIGHING MACHINE

Date	Items	Rtd	Issued	Balance	Sps
15/2/20	Taken from old stock Beronitwe			1	
	Book page no: 11 (Rd - 17/6/08)			NIL	
	Issued to use		1	NIL	
6/4/21	Trans to unscrutable register			NIL	Reli
	Page no: 27 (1)				
17/10/22	As per order NO: 850/2021 Dated				
	17/10/2022 condemned as weighing			NIL	New
	Machine. Rd. 25/2/2011 (1)				




 SUPERINTENDENT
 COMMUNITY HEALTH CENTRE
 THIRUVILWAMALA
 THIRISSUR - 680 588



COMMUNITY HEALTH CENTRE
Date.....
THIRUVILWAMALA

[Signature]
SUPERINTENDENT
COMMUNITY HEALTH CENTRE
THIRUVILWAMALA
THRISSUR - 680 588