


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC VALLACHIRA
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: N/A Model: ZY-102W SN: NA
4	Equipment ID/ Barcode	0843273 # 120653
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date – NA (As per stock book installation date - 14-09-2021, stock book for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 02-07-2024
9	Action taken	Checked the machine and found that compressor motor, sieve bed and main board defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. As per the stock book Installation date is 14-09-2021. Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Supplier details unavailable


Jr. Consultant (Biomedical)
NHM, Thrissur


Signature of JC BM (NHM)

Date:


Date.....
Superintendent/Medical Officer (i/c)
NHM, Thrissur


Signature of
Superintendent/Medical Officer (i/c)
NHM, Thrissur

കുടുംബാരോഗ്യ കേന്ദ്രം, വല്ലൂരി



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 206310

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <i>PNC</i></p> <p>Address <i>VALLACHIRA</i></p> <p style="padding-left: 150px;"><i>THRISSUR</i></p> <p>Ph : <i>8547512651</i></p>	<p>Call Registration Date : <i>02/07/24</i></p> <p>Caller ID : <i>120653</i></p> <p>Date of Visit : <i>02/07/24</i></p> <p>Asset No. : <i>0843273</i></p> <p>EQPT Name : <i>OXYGEN CONCENTRATOR</i></p> <p>Manufacture <i>NA</i> Model : <i>NA</i></p> <p>S. No. <i>NA</i> Dept. <i>P&P</i></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *MACHINE NOT SWITCHING ON*

Action Taken : *CHECKED THE MACHINE & FOUND THAT COMPRESSOR MOTOR SIEVE BED & MAIN BOARD DEFECTIVE, NEED TO REPLACE THESE SPARES FOR FURTHER CHECKING.*

Completed Date : *02/07/24* Time : *11:30 AM* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>AMBADY VS</i>	<i>02/07/24</i>	<i>10:00 AM</i>	<i>11:30 AM</i>

Customer Remark Completed Pending

<p>Service Engineer Name : <i>AMBADY VS</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>02/07/24</i></p> <p>Contact Number : <i>7907996506</i></p>	<p>Customer Name : <i>ASINI. A.G</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>02/07/24</i></p> <p>Contact Number : <i>8547512651</i></p> <p>Designation : <i>Pharmacist</i></p> <p>Hospital Seal : STORES CUSTODIAN FHC VALLACHIRA</p>
---	--



Phc vallachira -0843273-120653

5 messages

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>, bmetsr2020@gmail.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 3 Aug, 2024 at 2:22 pm

Dear sir,
kindly provide below supply details of below mentioned oxygen concentrator.

hospital - phc vallachira
equipment - oxygen concentrator
make - NA

model- ZY-10ZW
barcode - 0843273
caller id - 120653
SN- NA

With Regards,

BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

KMSCL <kmsclbmecentral1@gmail.com>
To: Bijo T Joy <bijocyrix@gmail.com>
Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 3 Aug, 2024 at 2:29 pm

Dear Bijo,

Kindly share the images of the equipment.

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>
Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 3 Aug, 2024 at 2:40 pm

[Quoted text hidden]

KMSCL <kmsclbmecentral1@gmail.com>
To: Bijo T Joy <bijocyrix@gmail.com>
Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 3 Aug, 2024 at 2:35 pm

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>
To: KMSCL <kmsclbmecentral1@gmail.com>
Cc: bmetsr2020@gmail.com, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sat, 3 Aug, 2024 at 2:43 pm

dear mam,
We checked the stock details at the hospital. As this machine has a donated item, we could not find the supply details at the hospital.

With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]



CENTRA MEDICAL EQUIPMENT MAINTENANCE
PROGRAM (MEMBER)
TOLL FREE NO: 1-800-425-7669
BAR CODE - (8004-890615225)
0843273

Model No. _____
Date of Issue No. 24th Oct 2001
210624341055



சென்ட்ரல் மெடிக்கல் எக்ஸ்சேன்ஸ்
Date.....
சென்ட்ரல் மெடிக்கல் எக்ஸ்சேன்ஸ்