



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 222613

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|--------------------------------------|---|
| Health Facility : <u>CIHCH Ranni</u> | Call Registration Date : <u>8/7/2024</u> |
| Address : <u>Parthanamthitta</u> | Caller ID : <u>121738</u> |
| <u>Pin - 689672</u> | Date of Visit : <u>9/7/2024</u> |
| Ph : <u>9188522985</u> | Asset No. : <u>0351230</u> |
| | EQPT Name : <u>Nebuliser</u> |
| | Manufacture <u>Life care</u> Model : <u>CN 03LC</u> |
| | S. No. <u>CN03LC20230901</u> Dept. <u>Casualty</u> |

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine Found that problem compressor motor with piston complaint

Completed Date : 9/7/2024 Time : 2:30pm Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|--------------|-----------|-------------|-----------|
| 1. <u>NA</u> | <u>NA</u> | <u>NA</u> | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|-----------------------|-----------------|---------------|---------------|
| <u>Sumesh Michael</u> | <u>9/7/2024</u> | <u>1:00pm</u> | <u>2:30pm</u> |

Customer Remark Completed Pending



Service Engineer Name : Sumesh Michael
Signature : [Signature]
Date : 9/7/2024
Contact Number : 9544367904

Customer Name : [Name]
Signature : [Signature]
Date : 9/7/2024
Contact Number : 9188522985
Designation : Superintendent
Hospital Seal : M.C.C.M. Taluk Head Quarters Hospital RANNI


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|--|--|
| 1 | Name of District | PATHANAMTHITTA |
| 2 | Name of Hospital | THQH RANNI |
| 3 | Name of Equipment with Make, Model and Serial Number | NAME- NEBULISER MAKE- Life care MODEL- CN 03 LC SL.NO- CNO3LC20230901 |
| 4 | Equipment ID & Barcode | 121738 & 0351230 |
| 5 | Date of purchase / Year of manufacture /Installation Date | 11/11/2023 |
| 6 | Warranty details (Yes/No) | NO WARRANTY |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | NO AMC/CAMC |
| 8 | Date of breakdown (Date of registration of complaint through email/ Toll free) | 08/07/2024 |
| 9 | Action taken | CHECKED AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE. |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | FULLY DAMAGED |
| 11 | Recommendations for repair (required service details) | NOT RECOMMENDING FOR REPAIR |
| 12 | Cost of spares (specify parts and cost) | NA |



| | | |
|----|--|---|
| 13 | Asset Value | Rs. 1478 /- |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | NA |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | CYRIX SERVICE REPORT ATTACHED |
| 16 | Reasons for recommending the equipment as BER | CHECK AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE. QUOTATION NOT SUBMITTED SINCE SPARES ARE NOT AVAILABLE IN THE MARKET.SO RECOMMENDING THE EQUIPMENT FOR RBER. |
| 17 | Name & Signature of CYRIX Authority | Sumesh Michael  |

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

The equipment used is casually. There will be at least 100 nebulisation per day. The equipment is completely beyond repair. As per stock entry the equipment was purchased on 11.11.2023. But as per BER guide line up to 9999/- should be used for at least 3 years. BER is recommended depending on the current equipment condition

Signature of JC BM (NHM) ^{Shyles} 30/1/24
CONSULTANT (BIOMEDICAL)
 NATIONAL HEALTH MISSION
 PATHANAMTHITTA

Date



Signature of ^{June 22/2/24}
 Superintendent **M.C.M. Taluk Head Quarters Hospital.**
 Medical Officer (i/c) **RANNI**

4 Nos Consumed as per orders

NI-1326/2019/Supt incm TITC/110 20.11.2019

Cr-0 (P) 23/1/95/144 fwp/Dtd. 29.3.1995

Balance

2 Nos

Balance

1/1/2023 Transferred to unrecivable Register
Page - 04

1

11/11/23 Removed from store

2

4

consumed due to per

order No A1-16411/2023/Supt

TITC/Renoi dated 23/2/23

→

1

3

2/12/23 Tr out to unrecivable Register
P-04 (DOR 15/12/18)

01



Ward Name : 7392 / OP/CASUALTY, Scheme : KASP

Confirmation Report

Issue Note(MIN) No : KL00310/LS/00461/23-24, Issue Date : 11/11/2023

With reference to Issue No. KL00310/LS/00461/23-24 dated 11/11/2023. I am forwarding here with the following items. Please acknowledge and return the original copy duly signed.

Print Date : 15/11/2023.

| S No. | Item Code | Item Name | Unit | Supplier Name | Page No | Volume | Rate | Qty | Value |
|-----------------|-----------|--------------------------------|------|---------------------------------|---------|--------|----------|-----|----------|
| 1 | D22248 | NEBULISER, | | LIFECARE | | | 1,478.40 | 2 | 2,956.80 |
| 2 | S27281 | NEBULIZATION MASK -ADULT, null | null | ROMSONS SCIENTIFIC&SURGICALS | | | 95.20 | 10 | 952.00 |
| Total Value ₹ : | | | | | | | | | 3,908.80 |

Remarks:

Institution Incharge

Generated by Ranni THQH (PEN No :- N/A)

Jolly Johns

Store Custodian
PHARMACIST STORE KEEPER
MCCM. THQH - RANNI

Consignment Collect by



1 for BER
Joyce
20/11/24




SUPERINTENDENT
M.C.C.M. Taluk Head Quarters Hospital,
RANNI

COMPRESSOR NEBULIZER

MODEL NO. : CN-03 LC
 INPUT SUPPLY : 230V/50HZ, 80VA
 BATCH NO. : CN03LC20230901
 MFG. DATE : SEPT. 2023

MARKETED BY: ANK LIFECARE PRIVATE LIMITED
 KOTHAYITA ROAD, OFF NH BYE PASS, NEAR THUKAVU JN.
 YENNALA P.O., KOTHAI-692 028 KERALA
 EMAIL ID : mail@anklifecare.com
 CUSTOMER CARE HELPLINE NO. 1800 1212 666
 COUNTRY OF ORIGIN : INDIA

MANUFACTURED BY: RANNEY MEDICAL INDIA PVT LTD
 2/27 MC STY INDUSTRIAL AREA, KANKAMMALA DEPT-110001
 ARTWORK NO. - IM/ART/PL/B/CN03LC
 MFG. LIC. NO. MFG/IND/2023/000284



HEALTH
 RANNEY
 HOSPITAL

Pooja
22/1/23
 SUPERINTENDENT
 M.C.C.M. Taluk Head Quarters Hospital
 RANNEY