



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 228274

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : THD H</p> <p>Address : Thuvavoor Alappuzha</p> <p>Ph : 9495779968</p>	<p>Call Registration Date : 6/7/2024</p> <p>Caller ID : 121460</p> <p>Date of Visit : 15/7/2024</p> <p>Asset No. : 0421312</p> <p>EQPT Name : Nebulizer</p> <p>Manufacture : Postura Model : NA</p> <p>S. No. : NA Dept. : Casualty</p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Nebulizer is not getting on

Action Taken : Regarding the inspection of the equipment found that the compressor motor with piston defective. Need to replace this spares for further checking and working condition of Equipment.

Completed Date : 15/7/2024 Time : 10:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Naseef	15/7/24	9:00AM	10:00AM

Customer Remark : Not working

Completed Pending

<p>Service Engineer Name : Naseef</p> <p>Signature : <i>Naseef</i></p> <p>Date : 15/7/24</p> <p>Contact Number : 9526744524</p>	<p>Customer Name : <i>Atanaw</i></p> <p>Signature : <i>Atanaw</i></p> <p>Date :</p> <p>Contact Number :</p> <p>Designation : Medical Officer in Charge</p> <p>Hospital Seal : Taluk Hospital, Thuvavoor Alappuzha - Pin: 688532</p>
---	---



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH THURAVOOR
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- PERFEXA MODEL :- BR-CN116 SN :- NA
4	Equipment ID & Barcode	121460 & 0421312
5	Date of purchase / Year of manufacture /Installation Date	27/05/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	06/07/2024
9	Action taken	Regarding the inspection of the equipment found that the compressor motor with piston defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS :- 1353 /-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached.
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Equipment installed on 27/05/2014 and covered up to 10 + years. Quotation is not submitted since spares are not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NASEEF K N <i>Naseef</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

[Signature]
18/07/24
JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA

Signature of JC BM (NHM)

For Condemnation

[Signature]

Medical Officer in Charge

Signature *[Signature]*
Taluk Hospital, Thuravoor
Superintendent / Medical Officer (i/c)
Alappuzha - Pin: 688532

Date





Atkanan

Medical Officer in Charge
Taluk Hospital, Thiruvavoor
Alappuzha- Pin: 688532