



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223228

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility DMH</p> <p>Address Peroor Kada Thiruvananthapuram</p> <p>Ph : 8547307521</p>	<p>Call Registration Date : 02-07-2024</p> <p>Caller ID : 120599</p> <p>Date of Visit : 08-07-2024</p> <p>Asset No. : 0113639</p> <p>EQPT Name : Weighing Machine</p> <p>Manufacture Local purchase Model : NA</p> <p>S. No. NA Dept. Pediatric O.P.</p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found board, load sensor and battery defective. Need to replace these spores for further checking and working condition of the equipment

Completed Date : 08/07/24 Time : 12:00 pm Spare Required


Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	NA	NA	NA
2.				
3.				

Cyrux Engineer	Date	Start Time	End Time
AKW Manohar	08/07/24	11:00 AM	12:00 PM

Customer Remark Pending

<p>Service Engineer Name : AKW Manohar</p> <p>Signature : <i>[Signature]</i></p> <p>Date : 08/07/2024</p> <p>Contact Number : 8921568757</p>	<p>Customer Name : Anjali S.S</p> <p>Signature : <i>[Signature]</i></p> <p>Date : 8/7/24</p> <p>Contact Number : 8547307521</p> <p>Designation : Senior Nursing Officer</p> <p>Hospital Seal :</p>
--	--




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	DISTRICT MODEL HOSPITAL PEROORKADA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Weighing Machine Make : SALTER Model : NA Serial No : NA
4	Equipment ID/ Barcode	120599/0113639
5	Date of purchase/ Year of manufacture/Installation Date	06/05/2011
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	02/07/2024
9	Action taken	Checked and found that the mainboard and load sensor defective. Need to replace these spare for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	6725/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found that the mainboard and load sensor defective. The unit was installed on 06/05/2011 and covered upto 13+ years. Since the spares are not available in the market, recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	AKHIL MANOHAR 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Main board and load sensor defective. So as per tender condition, this can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


26.07.2024
Signature of JC BM (NHM)

ok

Date: 26/07/24



Dr. B. UNNIKRISHNAN MBBS, LLB
Reg. No. 21986
Asst. Director, Kerala State Health Services
Deputy Superintendent
Govt. District Model Hospital
Peroorokada, Thiruvananthapuram

Signature of
Superintendent/Medical Officer (i/c)

