



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223748

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <i>TH Malayinkeezh</i>.....</p> <p>Address <i>Malayinkeezh</i>.....</p> <p>..... <i>Thiruvananthapuram, Kerala</i>.....</p> <p>Ph : <i>75 10 23 59 72</i>.....</p>	<p>Call Registration Date : <i>03/07/24</i>.....</p> <p>Caller ID : <i>120443</i>.....</p> <p>Date of Visit : <i>03/07/24</i>.....</p> <p>Asset No. : <i>0121870</i>.....</p> <p>EQPT Name : <i>BP Apparatus Mercury</i>.....</p> <p>Manufacture <i>Danamon</i>..... Model : <i>DeLuxe</i>.....</p> <p>S. No. <i>NR</i>..... Dept. <i>OP</i>.....</p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *checked and found mercury spillage. Identified mercury tank and glass rod defective.*.....

Action Taken : *Need to replace mercury tank and glass rod. Should refill the mercury to check the further working condition of the machine.*.....

Completed Date : *03/07/24* Time : *11:00 am* Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <i>NR</i>	<i>NR</i>	<i>NR</i>	<i>NR</i>
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Mohammed Syed M</i>	<i>03/07/24</i>	<i>10:00 am</i>	<i>11:00 am</i>

Customer Remark Completed Pending

<p>Service Engineer Name : <i>Mohammed Syed M</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>03/07/24</i></p> <p>Contact Number : <i>7293499129</i></p>	<p>Customer Name : <i>Sreenivasan. KC</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>3/7/2024</i></p> <p>Contact Number : <i>56235972</i></p> <p>Designation : <i>[Signature]</i></p> <p>Hospital Seal : <i>[Seal]</i></p>
---	--




Superintendent:
[Signature]
Tarak Headquarters Hospital
Malavinkal

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sr No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	TH MALAYINKEEZH
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Make : Diamond Model : Deluxe Serial No : NA
4	Equipment ID/ Barcode	120443/0121870
5	Date of purchase/ Year of manufacture/Installation Date	08/08/2017
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	02/07/2024(Toll free)
9	Action taken	Checked and found mercury spillage. Identified Mercury tank and glass rod defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found mercury spillage. Identified mercury tank and glass rod defective. The unit is installed on 08/08/2017 and covered upto 6+ years. Since the mercury is not available in the market. Recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	MOHAMMED SIYAD 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Mercury spillage, mercury tank and glass rod defective. So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


20.07.2024.
Signature of JC BM (NHM)

Date:




Signature of
Superintendent Medical Officer (i/c)
Taluk Headquarters Hospital
Malayinkil

8

REGISTER OF

(For watching the disbursement of monies drawn

Undisbursed balance


B. P. APPARATHU

Date	Annual serial number	Description of bill or cheque originally cashed	Amount of the bill or cheque	Amount undisbursed ^a
1	2	3	Rs. Balance	5
		Received		
26/1/16		Office establishment bill found on physical verification.	2000 - 2	
2/13/17		Received from stock Office travelling allowance bill (Bill NO 61969-2017-18) Page NO. 15)	2000 - 4	
8/8/17		Received from stock Cash stock page no 15)		5
		Deduct amount paid		
10/1/2018		Received from stock Closed purchase book Closing balance page no 11		7
17/1/2018		Received from stock LCD BP Apparathu P. NO 19	(RM)	8
17/1/2018		Received from stock Details of balance Donated two BP Apparathu P. NO 45	(RM)	10
11/1/2019		Received from stock condemnation condemned order no (5/1/2018 2/1/2018)	(RM)	8

^aAs soon as a bill is drawn, columns (1) to (4) should be filled in. Column (4) should be attested by the drawing officer when he signs the bill. After the bill is cashed, the amount of the day should be entered in column (5). Column (5) should be attested by the drawing officer at the end of the day.

4/12/19

14.000 50 sub
Date NO 7



Superintendent,
Tatalu Headquarters Hospital,
Malayankil

CYRIX
 HEALTHCARE SYSTEMS
 S/N: 238415
 CN: TBM009
 PRO : B.P Apparatus
 ID : 0121870
 CAL.ON : 20 12 2023
 CAL.DUE : 20 12 2024
 TEL: 001 800 425 7669
 FAX: 001 800 425 7669

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
 TOLL FREE NO: 1800 - 425-7669
 BAR CODE - (8004-890815225)
 0121870

S ATTENDANCE
 (NHM)

NOT WORK



TALUK HEADQUARTERS HOSPITAL
 MALAYINKIL
 PIN - 696 571.
 THIRUVANANTHAPURAM

[Signature]
 Superintendent
 Taluk Headquarters Hospital
 Malayinkil