



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
214882

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC Vallicode</u></p> <p>Address <u>Vallicode</u> <u>Pathanamthitta Pin: 689648</u></p> <p>Ph : <u>7510765394</u></p>	<p>Call Registration Date : <u>01/07/2024</u></p> <p>Caller ID : <u>120354</u></p> <p>Date of Visit : <u>03/07/2024</u></p> <p>Asset No. : <u>0340399</u></p> <p>EQPT Name : <u>Sterilizer</u></p> <p>Manufacture <u>NA</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Lab.</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found that body leakage gasket faulty. Safety valve and steam release valve faulty. Door and body handle faulty. Heating element and ceramic coil combitor with cable faulty machine fully damaged.

Completed Date : 03/07/2024 Time : 12:30 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sijo Zachariah</u>	<u>03/07/2024</u>	<u>09:00am</u>	<u>12:30pm</u>

Customer Remark Completed Pending



Medical Officer in-charge
Primary Health Centre
Vallicode

Service Engineer Name : Sijo Zachariah
Signature : [Signature]
Date : 03/07/2024
Contact Number : 8113954221


Customer Name : Sulekha V.S.
Signature : [Signature] 31/24
Date :
Contact Number : 7510765393
Designation : Nursing officer grade II
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PATHANAMTHITTA
2	Name of Hospital	PHC VALLICODE
3	Name of Equipment with Make, Model and Serial Number	Name- Sterilizer Make- Na Model- Na Sl.no- Na
4	Equipment ID & Barcode	120354 & 0340399
5	Date of purchase / Year of manufacture / Installation Date	04/10/2019
6	Warranty details (Yes/No)	No Warranty
7	*AMC/CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/Toll free)	01/07/2024
9	Action taken	Checked and found that body leakage, gasket faulty, safety valve and steam release valve faulty, door & body handle faulty, Element and Ceramic coil connector with cable faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Quotation Not Available


13	Asset Value	Rs. 7563.64/-
14	#PercentagevalueofthecostofspareswithrespecttoCostofPurchase/Asset Value	Not Available
15	Abstract of Service Report provided bythe OEM/ Authorized Service Provider/CYRIX(AttachedorNot)	Cyrix service report
16	Reasonsfor recommendingtheequipment as BER	Checked and found that body leakage, gasket faulty, safety valve and steam release valve faulty, door & body handle faulty, Element and Ceramic coil connector with cable faulty and the machine body fully damaged due to fully burning issue. The machine was aged up to 5 years and repairing of the machine was potential harms to patients and operational efficiency. So, the machine recommended for BER
17	Name&SignatureofCYRIXAuthority	Sijo Zachariah 

*Not mandatory #BasedontheperiodoflifeandvalueaspertheBERguidelines

*Attach Photograph


RemarksandRecommendationsofJuniorConsultant(Biomedical)NHM:


Physically verified on 19/01/2024. BER is recommended depending on the current equipment condition.


30/1/24
Signature of JCBM (NHM)
CONSULTANT (BIOMEDICAL)
NATIONAL HEALTH MISSION
PATHANAMTHITTA

Date 15/1/24

Seal




Signature of
Superintendent/Medical Officer (i/c)

AUTOCLAVE

Date	Description	Recd	Issue	Balanc	Part
4/10/19	Received from Vijay Dental Suppliers /mc Road Adoor as part of Kayakalp / United fund Invoice No: 528 / dated 4/10/19 / RqG Ri. 6356 / Tax 18%, CGST 5.72, SGST 5.72 KFC 1% (63.56) Total 7563.64 ₹				
4/10/19	Issued to Staff Nurse	01	01	01	01
			01	NIL	01



[Signature]
 Medical Officer in-charge
 Primary Health Centre
 Vallicode.

PRIMARY HEALTH CENTRE VALLICODE

Vallicode P O Pathanamthitta Phone No 04682356338

Email phcvallikode@gmail.com

No. 229/24


Date 15/07/2024


TO




Cyril

Sub: Autoclave machine complaint

This is to certify that, the Autoclave machine using at this institution having Barcode : 0340399 was examined by me and not found any serial no or make and model, imprinted on the machine nor entered in stock book:




Sajan Babu
Medical Officer In-charge
Primary Health Centre
Vallicode.



[Signature]
 Medical Officer in charge
 Primary Health Centre
 Vallicode.

