



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
230115

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>THRUH</u>	Call Registration Date : <u>01/07/2024</u>
Address : <u>North Paravur</u>	Caller ID : <u>120239</u>
<u>ELM</u>	Date of Visit : <u>03/07/2024</u>
Ph : <u>9744828306</u>	Asset No. : <u>0722969</u>
	EQPT Name : <u>Nebuliser</u>
	Manufacture <u>Life Line</u> Model : <u>Alpha</u>
	S. No. <u>NA</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Fumes not getting

Action Taken : Checked and found compressor motor with piston defective. Need to replace these parts for further checking and ensuring condition of the equipment.

Completed Date : 3/7/24 Time : 11:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Motor</u>	1		
2. <u>Piston</u>	1		
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sebn Siby</u>	<u>3/7/24</u>	<u>10:00am</u>	<u>11:30am</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Sebn Siby</u>	Customer Name : <u>SHIPRENTENDENT HEADQUARTERS HOSPITAL</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>3/7/24</u>	Date : <u>3/7/24</u>
Contact Number : <u>9746751738</u>	Contact Number : <u>9744828306</u>
	Designation : <u>SNO</u>
	Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	THQH NORTH PARAVOOR ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Nebuliser Make : Life-Line Model : Alpha SN : Na
4	Equipment ID & Barcode	120239 & 0722969
5	Date of purchase / Year of manufacture /Installation Date	03-09-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01-07-2024 (Toll free)
9	Action taken	Checked and found compressor motor with piston defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1353
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Machine installed on 03-09-2018 and covered up to 5+ years. Quotation not attached since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Sebin

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

unable to repair reported compressor and piston defective
Installed on 2018.

Submitted BER after 30 days.

Devathy
Consultant BNE
24/7/20
Signature of JC BM (NHM)
Bio-Medical Consultant (NHM)
National Health Mission
Aroyyakkulam Erumam

Date

Seal

Signature of
Superintendent / Medical Officer (i/c)



At
SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
NORTH PARAVUR

STOCK REGISTER

NEBULIZER

B/F from page no. 78
147

Date	Invoice/Voucher No.	Received from or issued to	Cost	Receipt	Issue	Balance	Remarks	Unit
		Tr in from Page No: 78				(11+5) C16)		
		Being unserviceable						
		Tr out to unserviceable Register Page no 35			5			
		Condemned as per order no: 1791 dt - 21/5/17 Supt TIA&H N-Paravur Ekm			4	12		
3/9/18		Rd from Store (Donation)		1		13		
3/11/18		Rd from Store (Donation)		2		15		
		Tr out to unserviceable register P. no. 45			6			
		Condemned as per order no: C-723/19 Supt TIA&H N-Paravur dated 20/5/19			6	9	11/3/18-1 5/3/18-1 19/5/18-4 3/9/18-1 3/11/18-2	
17/1/2020		Tr out to mmu (page no 40)		1		8	3/11/18-1	Stg
2/5/2021		Tr out to covid isolation (page no 22)		1		7	3/11/18-1	majid
1/10/21		Tr in from CSLTC			1	8	26/3/21	
		6 no: Tr out to unserviceable register p no 61						
		Condemned 6 no: as per order no Supt/TIA&H N-Paravur dt 3/12/22					0721488-105 Neb 0721494-105 gel 2-736/22/	
		Balance				2	3/9/18 26/3/21	



SEPARATE INDEPENDENT
TALUK HEAD QUARTERS HOSPITAL
NORTH PAPAVUR

