



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 183564

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>THOH Pulinkunnu</u>	Call Registration Date : <u>24/2/24</u>
Address <u>THOH Pulinkunnu, Alappuzha</u>	Caller ID : <u>98049</u>
Ph : <u>944788443</u>	Date of Visit : <u>26/2/24</u>
	Asset No. : <u>0421184</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture <u>life care</u> Model : <u>Piston compressor</u>
	S. No. <u>N/A</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Fumes not getting

Action Taken : checked and found compressor motor with piston defective. Need to replace these spares for further checking working condition of the equipment.

Completed Date : 11/3/24 Time : 12:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Vishnu Mohan</u>	<u>11/3/24</u>	<u>11:00am</u>	<u>12:30pm</u>

Customer Remark Completed Pending


Service Engineer Name : <u>Vishnu Mohan</u>	Customer Name : <u>Srejanand N's</u>
Signature :	Signature :
Date : <u>11/3/24</u>	Date : <u>11/3/24</u>
Contact Number : <u>9072522084</u>	Contact Number : <u>944788443</u>
	Designation : <u>NSG-officer</u>
	Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH PULINCUNNU
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- LIFE CARE MODEL :-PISTON - COMPRESSOR SN :- NA
4	Equipment ID & Barcode	98049 & 0421184
5	Date of purchase / Year of manufacture /Installation Date	27/10/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24/02/2024
9	Action taken	Regarding the inspection of the equipment found that compressor motor with piston defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS:- 1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Equipment installed on 27/10/2018 and covered up to 5 years and 4 months. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	VISHNU MOHAN NAIR 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


12/03/24
JYOTHISH THOMAS
Signature of JCBM (NHM)
NHM ALM


piston irrevocably damaged

Date

11/3/24



Seal


Signature of
Superintendent / Medical Officer (i/c)

THQ HOSPITAL
PULINCUNNU

webulizer

glaska Receipt Book

33

24/10 To - in books F.W.
Inst No. 100 No. 100 page no. 100

24/10 Received from store inst no. 100
Vol. No. IV G page no. 292. Rs 1300/-

24/10 Received from store
(Parva Books) (2 books out)

25/10 Received from the store
Cost Rs 1930/- page no IV G Inst 369

25/10/15 As per order no - B-1000
(2015) 54444 Dated 1/10/2015
Cancelled.

25/10 Receipt from the store
Mathew Job
State Stores Verification Officer
Directorate of Health Services
Trivandrum, Kerala

25/11/17 Received from the store
Invoice no 309 Reli 2952-38/-
from K.S.T Ekro
(Inst. Page Vol. no IV G
page no 25

25/11/18 Rd from store
Page no. main stock page no 25 (vol. 59)

SUPERINTENDENT
THE HOSPITAL
PULINCUNNU

25/11/18 Rd from store
Page no 26 Sastharama Regula I G
Kannaya medical Aluva 6 km
Invoice no 1977 dt. 25/11/18 Rs. 1030/-



