



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 229581

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : District Hospital	Call Registration Date : 07-06-2024
Address : Aluva, Ernakulam (Dt)	Caller ID : 115862
Ph : 09961169527	Date of Visit : 08-06-2024
	Asset No. : 0711233
	EQPT Name : Nebuliser
	Manufacture : Fruechek Model : N8001
	S. No. 680011801126 Dept. Special Ward

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Nebuliser not working

Action Taken : checked the machine and found that motor and piston are faulty

Completed Date : 08-06-24 Time : 4:00PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Motor	1 No.		
2. Piston	1 No.		
3.			

Cyrix Engineer	Date	Start Time	End Time
Jibin Thomas George	08-06-24	3:00PM	4:00PM

Customer Remark Completed Pending


Service Engineer Name : Jibin Thomas	Customer Name : Di Smiji George
Signature : <i>Jibin Thomas</i>	Signature : <i>Smiji George</i>
Date : 08-06-2024	Date :
Contact Number : 9047958051	Contact Number :
	Designation : SUPERINTENDENT
	Hospital Seal : DISTRICT HOSPITAL ALUVA

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	DISTRICT HOSPITAL, ALUVA
3	Name of Equipment with Make, Model and Serial Number	Equipment: NEBULISER Make : TRUCHECK Model : N8001 Sr.no :6800118011126
4	Equipment ID & Barcode	115862 & 0711233
5	Date of purchase / Year of manufacture /Installation Date	21-11-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07-06-2024 (Toll Free)
9	Action taken	Checked the machine and found that motor and piston are faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs.1353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that motor and piston are defective. Equipment installed on 21-11-2017 and the machine outlived 7+ years. Quotation not submitted since the spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Installation date is 25/5/17 as per asset, on stock it is 21/11/17. Motor and piston faulty reported and reported can available spares. Not-attached spare enquiry details. Submitted after 30 days.

Devaly Consultant
24/7/24
Signature of JC BM (NHM)




Superintendent / Medical Officer (i/c)

Signature of seal

Date

STOCK REGISTER

Reg. No. _____

Page No. _____

ITEM Electricals

Date/No	Received/issued to whom	Opening Balance	Receipt	Total	Issue	Balance	Initials of receiver	Remarks
6.1.18	Ad from store		1			1		
	Remund from the store		1 (R-1945)			2		
12/4/19	TR INV FROM CH ward Rs 2200/- Hmc vol 5 CH ward main B.R for Instruments & Furniture Page no: 32		1			3		
14/2/21	To out to consular chgo hod. Date 25				2	2		
15/1/22	Out to consular chgo hod. Date 25				1	1		
	Condemned as per order 466/22 dt 14/1/22				1	0		



