



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
191209

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC Theikodithanam</u></p> <p>Address <u>Theikodithanam</u> <u>K. H. H. Nagar</u></p> <p>Ph : <u>9544752133</u></p>	<p>Call Registration Date : <u>29/6/2024</u></p> <p>Caller ID : <u>120116</u></p> <p>Date of Visit : <u>2/7/2024</u></p> <p>Asset No. : <u>0540476</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>Resdy Mist</u> Model : <u>Piston</u></p> <p>S. No. <u>28295</u> Dept. <u>OP</u></p>
-----------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Fumes not coming.

Action Taken : The compressor & motor is damaged. Need to replace the motor & compressor in order to check the working condition of the equipment.

Completed Date : 2/7/2024 Time : 10:20am Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Sreeju Suresh</u>	<u>2/7/2024</u>	<u>10:05am</u>	<u>10:20am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Sreeju Suresh</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>2/7/2024</u></p> <p>Contact Number : <u>7902854749</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Signature]</u></p> <p>Contact Number : <u>9809030717</u></p> <p>Designation : <u>N/O</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
-------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------



Elizabeth

MEDICAL OFFICER IN CHARGE
FHC TRICKODITHANAM - 686 104

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**


Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC THRIKODITHANAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Nebulizer Make:Ready Mist Model: Piston SN: 028295
4	Equipment ID & Barcode	120116&0540476
5	Date of purchase / Year of manufacture /Installation Date	26/04/2008
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29/06/2024 (Toll free)
9	Action taken	Checked and found out the problem .with compressor and motor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A

30
10/04/2024
MEDICAL OFFICER
PHC THRIKODITHANAM



13	Asset Value	Rs. 1353/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with compressor and motor. Machine installed in 26/04/2008 and covered up to 16 + years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sreerju suresh 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset value : Rs 1353
 cost of spare : Reported that spare not available
 life period = (26/4/2008) 16 years.
 As per the RRER guidelines recommended as BER

CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION

Signature of JC BM (NHM)

Date

Seal

Signature of 
 Superintendent Medical Officer (u/c)

MEDICAL OFFICER IN CHARGE
FHC THIRKODITHANAM - 686 104



1800 cot

19.12.05 Rd from store
 19.12.05 IN use

28/9/2020 shifted to DCC Kurnurphoam
 (St. Xavier's church)

3/9/2021 shifted to Amara sub-centre

Medical Officer in Charge
 Primary Health Centre
 Thirudithanam
 Kottamuri, PO

Received from DCC Kurnurphoam
 Shashi Bala JHA

Received	Issued	Balance
12	12	12
	12	NIL
	(5)	7
	1	6
5		11
	in use	

Nebuliser -

26.4.08 Received from hi store.

26.4.08 IN use Reckitt's Para.
 1353/—

7/4/18 Received from store
 nebuliser

(webmart - small size)
 (stock in new block)
 Amara

Transferred to
~~1300~~ Substock register 2
 page no 26

Compressor Nebulizer

Model No.: ReadyMist
Input : 230V/50Hz
Current : 0.8A
Serial No.: 028295
Lot No.: 27/04/07

Mode of operation-continuous
No AP/APG IPX0



Elisabeth
MEDICAL OFFICER IN CHARGE
MC THRICKODITHANAM - 686 104





Elizabeth
MEDICAL OFFICER IN CHARGE
FHC THIRICKODITHANAM - 686 104



Elizabeth
MEDICAL OFFICER IN CHARGE
FHC THRICKODITHANAM - 686 104



A circular purple stamp with the text "FAMILY HEALTH CENTRE" at the top and "THRICKODITHANAM - 686 104" at the bottom. In the center of the stamp is a crest featuring a lion and a unicorn, which is the emblem of the Government of Kerala.