



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
228279

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>THQH</u></p> <p>Address : <u>Cherthala</u> <u>Alappusha</u></p> <p>Ph : <u>9400864340</u></p>	<p>Call Registration Date : <u>28/6/2024</u></p> <p>Caller ID : <u>119791</u></p> <p>Date of Visit : <u>18/7/24</u></p> <p>Asset No. : <u>0420025</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>NIDEK</u> Model : <u>4235CE</u></p> <p>S. No. <u>NA</u> Dept. <u>Casualty</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Nebulizer is not getting on

Action Taken : Regarding the inspection of the equipment found that the compressor motor with piston defective. Need to replace this spares for further checking and working condition of Equipment.

Completed Date : 18/7/24 Time : 11:00 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Naseet</u>	<u>18/7/24</u>	<u>10:00 am</u>	<u>11:00 am</u>

Customer Remark
Not working
Sreedevi's (SNO)

Completed Pending

<p>Service Engineer Name : <u>Naseet</u></p> <p>Signature : <u>Naseet</u></p> <p>Date : <u>18/7/24</u></p> <p>Contact Number : <u>9526 744524</u></p>	<p>Customer Name :</p> <p>Signature :</p> <p>Date :</p> <p>Contact Number :</p> <p>Designation :</p> <p>Hospital Seal :</p>
---	---




SUPERINTENDENT
VALUK HEAD QUARTERS HOSPITAL
CHERTHALA

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- NIDEK MODEL :- 4235CE SN :- NA
4	Equipment ID & Barcode	119791 & 0420025
5	Date of purchase / Year of manufacture /Installation Date	21/08/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	28/06/2024
9	Action taken	Regarding the inspection of the equipment found that the compressor motor with piston defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS :- 1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached.
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Equipment installed on 21/08/2018 and covered up to 5 + years. Quotation is not submitted since spares is not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NASEEF K N 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Recommended for BER



18/07/24


Signature of JC BM (NHM)

For condemnation

Date




18/7
Signature of
Superintendent / Medical Officer (i/c)

**SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA**

Thaluk Head Quarters Hospital, Cherthala

302

21-08-18 purchased from kattungal trades Bill No 231 dt

1.	Bp Apparatus Mercury type	5 Nos	2880.00
2.	Accu check strip 50's pack	2 Nos	750.00

21-08-18 purchased from kattungal trades Bill No 234 dt

1.	Accu check Machine	4 Nos	999.00
2.	Accu check strip 50's	8 Nos	750.00
3.	Stethoscope	5 Nos	500.00

21-08-18 purchased from kattungal trades Bill No 235 dt

1.	Nebuliser	2 Nos	2950.00
----	-----------	-------	---------

21-08-18 purchased from kattungal trades Bill No 233 dt

1.	Bp Apparatus mercury type	5 Nos	2680.00
2.	Accu check Machine	1 No	999.00


21-08-18 purchased from devalekshmi Agencies, Mayiltham,


1.	2 x 1 pvc Reducer	3 Nos	17.00
2.	1 1/2 pvc pipe	25	30.00
3.	1 1/2 Geot Van	3 No	34.00
4.	1 1/2 m/n	3 No	5.00
5.	Solvent 100ml	1	54.00
6.	Bottle	3 Nos	9.00
7.	1 1/2 pvc tee	3 Nos	3.

22-8-2018 Received from Kinniya Community Pharmacy, Cherthala
 1. No. 21/9/2, dt 22.08.2018.

1. Cap Ye Gut. 10 Nos. 9.31




SUPERINTENDENT
THALUK HEAD QUARTERS HOSPITAL
CHERTHALA
 Sreedev's (SVO)



RA
PSK



Sreedhar
Sreedhar's (SNO)

[Signature]
SUPERINTENDENT
TAKUR HEAD QUARTERS HOSPITAL
CHERTHALA