



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
231312

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>T.H.Q.H</u></p> <p>Address <u>Cherthala</u></p> <p>..... <u>Alappusha</u></p> <p>Ph : <u>9400864340</u></p>	<p>Call Registration Date : <u>23/6/2024</u></p> <p>Caller ID : <u>119790</u></p> <p>Date of Visit : <u>18/7/24</u></p> <p>Asset No. : <u>04-20027</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Optima</u> Model : <u>END1-WD</u></p> <p>S. No. <u>007382</u> Dept. <u>Casualty</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Nebulizer is not getting on

Action Taken : Regarding the inspection of the equipment found that the compressor motor with piston defective. Need to replace this spare for further checking and working condition of equipment.

Completed Date : 18/7/24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nasuf</u>	<u>18/7/24</u>	<u>11:00am</u>	<u>12:00pm</u>

Customer Remark Not working See serial P n/o Sup

Completed Pending



Service Engineer Name : Nasuf
Signature : [Signature]
Date : 18/7/24
Contact Number : 9526741524

Customer Name :
Signature :
Date :
Contact Number :
Designation :
Hospital Seal :

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- OPTIMA MODEL :- CN01-WD SN :- 007382
4	Equipment ID & Barcode	119790 & 0420027
5	Date of purchase / Year of manufacture /Installation Date	21/08/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	28/06/2024
9	Action taken	Regarding the inspection of the equipment found that the compressor motor with piston defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS :- 1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached.
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Equipment installed on 21/08/2018 and covered up to 5 + years. Quotation is not submitted since spares is not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NASEEF K N <i>Naseef</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

[Signature] 18/07/24

JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 NHM ALAPPUZHA

Signature of JC BM (NHM)

For condemnation

Date



[Signature] 18/7

Signature of
 Superintendent / Medical Officer (i/c)

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA

Thaluk Head Quarters Hospital, Cherthala

302

21-08-18 purchased from kallungal trades Bill No 231 dt

1.	Bp Apparatus Mercury type	5 Nos	2880.00
2	Accu check strip 50's pack	2 Nos	750.00

21-08-18 purchased from kallungal trades Bill No 234 dt

1.	Accu check Machine	4 Nos	999.00
2	Accu check strip 50's	8 Nos	750.00
3	Stethoscope	5 Nos	500.00

21-08-18 purchased from kallungal trades Bill No 235 dt

1.	Nebuliser	2 Nos	2950.00
----	-----------	-------	---------

21-08-18 purchased from kallungal trades Bill No 233 dt

1.	Bp Apparatus mercury type	5 Nos	2680.00
2	Accu check Machine	1 No	999.00

21-08-18 purchased from devalekshmi Agencies, magalthara,

1.	2 x 1 pvc Reducer	3 Nos	17.00
2	1" pvc pipe	25	30.00
3	1" G-eot Van	3 No	34.00
4	1" m/n	3 No	5.00
5	Solvent 100ml	1	54.00
6	Pin Hole	3 Nos	9.00
7	1" pvc tee	3 Nos	3.

22-8-2018 Received from Kinnari Community Pharmacy, Cherthala
 1. No. 2/9/2, dt 22.08.2018.




PSO
 SUPERINTENDENT
 CHERTHALA
 THALUK HEAD QUARTERS HOSPITAL

Sreedevi (SNO)

10 Nos.

9.31



PA / PSE
 18/7
 Sreedan
 Sreedan's (SNO)

ATTENDENT
 OPTIMA