



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 230269

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THOH, Kayamkulam	Call Registration Date : 26/06/2024
Address : Alappuzha	Caller ID : 119462
Kerala	Date of Visit : 27/06/2024
Ph : 9656058520	Asset No. : 0421051
	EQPT Name : BP apparatus
	Manufacture : Anitha Model : Etkometer
	S. No. 395135 Dept. Medical ward

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mercury leaking and Mercury tank Found broken.

Action Taken : checked the machine and found that mercury was lost, Mercury tank, BP bladder and hinges are defective. Needs to replace these spares for further checking and working condition of the equipment.

Completed Date : 27/06/24 Time : 12:00Pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Mxidula-G	27/06/24	11:15am	12:00Pm

Customer Remark Completed Pending

Service Engineer Name : Mxidula-G
Signature : *[Signature]*
Date : 27/06/2024
Contact Number : 8590207163

Customer Name : Mumthas
Signature : *[Signature]*
Date : 27/6/24
Contact Number : 9656058520
Designation : Nursing officer
Hospital Seal :



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH KAYAMKULAM
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS MAKE:- ANITHA INDUSTRIES MODEL:- ELKOMETER SN:- 395135
4	Equipment ID & Barcode	119462 & 0421051
5	Date of purchase / Year of manufacture /Installation Date	23/03/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	26/06/2024
9	Action taken	Checked the machine and found that mercury was lost. Mercury Tank, BP Bladder and Hinges are Defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS:-1323
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report attached.
16	Reasons for recommending the equipment as BER	Checked the machine and found that mercury was lost. Mercury Tank, BP Bladder and Hinges are Defective. The equipment was installed on 23/03/2018, aged up to 6+ years. Quotation not submitted since spares are not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	MRIDULA G

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

[Signature]
23/07/24

Signature of JC BM (NHM)

Date



[Signature]
Signature of
Superintendent / Medical Officer
Govt. Taluk Hospital
Kayamkulam

121

B.P. Apparatus

Dd Balance

6/1/06	originally received		1	1	2015
3/1/06	in cont to female wd (N.B) in form mail			2	31
24/9/07	Rfd to store			3	13-8
10/1/2010	Rfd from D.D.		1		31
14/5/2010	Bell no 12 dt 14/5/2010 Rdfm store		1	4	
3/12/13	Rd		1	5	
	Three BP apparatus condemned on 23.2.2013 as per order no GOCMS) 63/2013/H&FWD	3		3-3=0	
6/16	Rd		1	2+1=3	
	103 per order no: B. 1636/2015/Supt T&K / Dtd			3-1=2	
18/7/2016	1 Condemned date			2+1=3	
1/2/17	Received [68041574]	1		4	
23/3/18	rd from store	1			
	S104A (Page num 53)				
2/11/18	Received from store				

Pad Lock

25/2/06	Rd from HDG Stock			2	
24/7/2010	Rd from store (Page no. 157 L-P book Cost Rs 74)		1	3	
	Two Pad lock condemned on date 23.2.2013 as per order no GOCMS) 63/2013/H&FWD	2		3-2=1	
24/3/15	Rd from store Lock & key (Lp 223 Volume - 8) Rs. 60	1		2	
31/1/2020	Rd from store		4	Rs. 300	
15/2/22	It was to unusable Repetition paper - 15 (Date of Recy) 24/3/16, 24/7/16	2	2		

[Handwritten Signature]
 Superintendent
 Govt. Taluk Hospital
 Kayamkulam





Superintendent
Govt. Taluk Hospital
Kayamkulam