



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 223746

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

550

Health Facility : FHE Kottikal	Call Registration Date : 22/06/24
Address : Kottikal	Caller ID : 118489
Thiruvananthapuram, Kerala	Date of Visit : 22/06/24
Ph : 9400 09 71 66	Asset No. : 0140268
	EQPT Name : Weighing Machine
	Manufacture Rossmax Model : K.B.10.1
	S. No. R/S. No. B/Loc. No. Dept. : OP

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : checked and found board and load sensor defective

Action Taken : Need to replace board and load sensor & check the further working condition of the machine.

Completed Date : 22/06/24 Time : 01:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. NR	NR	NR	NR
3.			

Cyrix Engineer	Date	Start Time	End Time
Mohammed Sijadh M	22/06/24	12:30 pm	01:00 pm

Customer Remark Completed Pending

Service Engineer Name: Mohammed Sijadh M
Signature: [Signature]
Date: 22/06/24
Contact Number: 7293499127

Customer Name: [Signature]
Signature: [Signature]
Date: 22/06/24
Contact Number: 8131972485
Designation: [Signature]
Hospital Seal: [Seal]

MEDICAL OFFICE CENTRAL
FAMILY HEALTH CENTRE, KOTTU
KOTTUKAL P.O. PUNNAKULAM
THIRUVANANTHAPURAM - 695 50


550

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	PHC KOTTUKAL
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Weighing Machine Make : Rossmax Model : WB101 Serial No : NA
4	Equipment ID/ Barcode	118489/0140268
5	Date of purchase/ Year of manufacture/Installation Date	03/04/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	21/06/2024(Toll free)
9	Action taken	Checked and found board and load sensor defective.Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	850/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found board and load sensor defective. The unit is installed on 03/04/2021 and covered upto 3+ years. Since the spares are not available in the market. Recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	MOHAMMED SIYAD 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Board and load sensor defective.
So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram




18.07.2024

Signature of JC BM (NHM)

Date:




Signature of ~~for~~ MOI/c
Superintendent/Medical Officer, (i/c) IN CHARGE
FAMILY HEALTH CENTRE, KOTTUKAL
KOTTUKAL P.O., PUNNAKKULAM
THIRUVANANTHAPURAM - 695 501


Weight Machine

Date	Received from Pharmacy	Receipts	Issued to	Balance	Remarks
11/18	Weight Machine (Digital)	2	1 (OP)	1	2 (NOLocore)
6/8/18	(Mahaveer) (1953-19)		1 (OP)	NIL	6/8/18
15/5/21		c/o to CR person - 7		1	
2/1/21	Weighting machine (Rosemax)	1			18/11
2/1/21	(c/o to UR person - 7)		1 (OP)	NIL	Order made done on 2/6/22
3/4/21	Weight machine (Duro)	1		1	Weight machine handed over Pharmacist
15/5/21	(Rosemax)		1 (OP)	NIL	C & F form & scanner image report in
25/4/2023	Received from Store Rs 1400/-	1		2	


Signature
Dr. K. Raj
for MOHC.



MOHC, P.O. PUNNAKKULAM
THIRUVANANTHAPURAM - 695 501

 **CYRIL** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0140268

rossmax
Model No. WB101
Max: 180kg / 397lb
3"AAA" Batteries
Manufacturer: Rossmax Swiss GmbH,
Framstrasse 16, CH-9442, Berneck, Switzerland
LOT: 195P02



Handwritten signature
Pa-Ligi Raj
for MO etc
MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE, KOTTUKAL
KOTTUKAL P.O., PUNNAKKULAM
THIRUVANANTHAPURAM - 695 501