



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	CHC EZHIKKARA ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Laboratory Microscope Make : Labomed Model : ATC2000 SN : Na
4	Equipment ID & Barcode	117829 & 0732859
5	Date of purchase / Year of manufacture /Installation Date	14-08-2009
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18-06-2024 (Toll free)
9	Action taken	Checked and found eye piece, prisam, objective lens and stage assembly defective. Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 34440 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and EOL letter attached
16	Reasons for recommending the equipment as BER	Checked and found eye piece, prisam, objective lens and stage assembly defective. Machine installed on 14-08-2009 and outlived 14+years. Enquired the spare with oem. The Oem informed that spares are not available and this model is obsolete As per the tender clause 5.3.14.2 we recommending the equipment for condmination.
17	Name & Signature of CYRIX Authority	Sebin

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


As per asset installed date is 18/8/2009. mfg universal agencies reported spare unavailabl. Report Attached.  
 Lense, eyepiece, prisam, objective Lense and stage assembly defective reported. Submitted BER after 30 days.

*Sevin*  
 Consultant BME  
 28/7/24  
 Signature of JC BM (NHM)

Chief Medical Officer (NHM)  
 National Health Commission (NHM)  
 Arakkulam

Date

*Noted*  
*Dr. Vinod Paulose*  
 30/7/2024  
 Signature of  
 Superintendent / Medical Officer (i/c)



പ്രധാനപ്പെട്ട ഓഫീസർ ഇൻചാർജ്ജ്  
 യാമുപുരം നോളി കേന്ദ്രം  
 എഴുതിക്കുക

# UNIVERSAL AGENCIES

Vision for excellence

ISO 9001 : 2008

PALAKKAL ANGADI, THRISSUR - 680 001, KERALA  
0487 - 2427038, 2427031, Mob : 93870 211122, 8943341569  
Email : unilage@gmail.com, unilage.sales@gmail.com



UA-L. 009/2024

03/04/2024

DISTRICT INCHARGE  
KLBEMMP-ERNAKULAM  
CYRIX HEALTH CARE PVT. LTD

Dear Sir

**Sub: Service of Microscope Model ATC 2000**

With reference to above and we wish to inform you that the further repair is not possible on seeing the condition of the equipment. Also due to the non availability of the spare parts in the market as this is an old model. Hence we would recommend buying a new machine. Sorry for the in convenience caused to you in this regards.

Thanking You


*Noted*  
*Jee*  
*Dr. Vinod Paulose*  
*30/7/2024*



Yours Faithfully,  
For Universal Agencies,

*[Signature]*  
Manager.

മുദ്രിപ്പിച്ചത് 30-07-2024  
മുദ്രിപ്പിച്ചത് 30-07-2024  
മുദ്രിപ്പിച്ചത് 30-07-2024




LAB REGISTER

Date	Received from	Received from	Quantity	Unit	Value	Particulars	Amount
------	---------------	---------------	----------	------	-------	-------------	--------

MICROSCOPE

11	From DMS (Monocular)		1				
	From DMS (Monocular)		1				
2/8/00	From DTSE (Binocular MLW)		1		3		
14/1/00	To SMCH, Coimbatore (Li. no. C.T. 22382/2000 Dated: 1.11.2000) Monocular		1		2	Lab	
							(1 Not in working order)
18/2/09	Binocular microscope (From DHS. Tum) Labomed DTC 2000		1		3	Lab	4740/-
2/8/09	To THRU, A. Parur. (MLW) (Li. no. SA 329/2000/DTC/EDW/29.07.09)		1		2	Lab	
15/1/12	As per request of SIS, issued to THRU paravur		1		1		
21/2/17	From TU Biolabomed Binocular Microscope (Model no. LX400 with LED illumination Rs 46758/-)		1		1		

True copy  
Dr. Vijay Prasad  
Principal  
Government College  
Mannargudi, Tirunelveli  
611302





Noted  
 Dr. Vinod Paul  
 30/7/2024  
 മലയാളം കേന്ദ്രം  
 മലയാളം കേന്ദ്രം  
 ഏകീകരണം

