



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223693

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Taluk Hospital</u></p> <p>Address <u>Vithura</u></p> <p>..... <u>Trivandrum</u></p> <p>Ph : <u>8281464462</u></p>	<p>Call Registration Date : <u>15/06/2024</u></p> <p>Caller ID : <u>117458</u></p> <p>Date of Visit : <u>20/06/2024</u></p> <p>Asset No. : <u>0121536</u></p> <p>EQPT Name : <u>Oxygen Concentrator</u></p> <p>Manufacture <u>Longfian</u>..... Model : <u>Jay-1.0</u>.....</p> <p>S. No. <u>MZJ.10D.56294</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working

Action Taken : checked and found sieve bed and solenoid valve defective - need to replace these spares for further checking and working condition of the equipment - Enquired spares from OEM

Completed Date : 20/06/2024 Time : 11:00 AM Spare Required

Spare Replaced Requested


Sl. No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>Na</u>	<u>Na</u>	<u>Na</u>	<u>Na</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Midhun - BM</u>	<u>20/06/2024</u>	<u>9:30 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Midhun - BM
Signature : [Signature]
Date : 20/06/2024
Contact Number : 9074167987

Customer Name : Ashwathy . A . S .
Signature : [Signature]
Date : 20/06/24
Contact Number : 93066 08947
Designation : Staff Nurse
Hospital Seal : [Seal]




Medical Officer (In Charge)
Taluk Hospital, Vithura

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH VITHURA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name :Oxygen concentrator Make : Longfian Model : JAY 10 Serial No : NA
4	Equipment ID/ Barcode	117458/0121536
5	Date of purchase/ Year of manufacture/Installation Date	14/08/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	15/06/2024
9	Action taken	Checked and found sieve bed and solenoid valve defective. Need to replace these spares for further checking and working condition of the unit. Enquired spares from OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Sieve Bed (2) – 13361.89/- Solinoid valve – 1670.24/- Freight cost – 12526.77/- Total–27558.89/-

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	103.99%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found sieve bed and solenoid valve defective. The repairing cost is more than the asset value, which is 103.9%. The unit was donated to the institution at covid time. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Sieve bed and solenoid valve damaged. Since repair cost is high. it can be BER

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


18-07-2024
Signature of JC BM (NHM)

Date:




Signature of
Superintendent/Medical Officer (i/e)

Medical Officer (In Charge)
Taluk Hospital, Vithura

133

Dg. Conservator

Rd. near Barun

14/12/21


Revenue for work

cost

1

[Signature]

Medical Officer (in Charge)
Taluk Hospital, Vithura



Taluk Hospital, Vithura



[Handwritten Signature]
 Medical Officer in Charge
 Taluk Hospital, Vithura



REGARDING OXYGEN CONCENTRATOR SPARE QUOTATION (117458-0121536)

1 message

Kasyep P V <kasyepcyrix@gmail.com>
To: long-fian@longfian.com

Tue, 9 Jul 2024 at 1:05 am

Dear Sir,
I'm reaching out to enquire about a quotation for the parts for the Oxygen concentrator.
Make : Longfian

Required quotation for;
1. seive bed - 2 nos
2. Solenoid valve

We look forward to hearing from you soon.

Thanks & Regards,

KASYEP PV
District Incharge , Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com



Spare part for JAY-10

5 messages

Gavin <wendy@longfian.com>
Reply to: Gavin <wendy@longfian.com>
To: kasyepcyrix <kasyepcyrix@gmail.com>

Tue, 9 Jul 2024 at 7:21 am

Hi Kasyep,

Good day.

We get your email for the spare parts.

1. Sieve beds 372-φ88 2nos Total price USD160.

2. Solenoid valve 1pcs price 20USD.

The above prices do not include shipping cost.

Best regards

Gavin Liu

Global Marketing Manager

LONGFIAN SCITECH CO.,LTD

Add:2F&3F,East Section,Building 12,Power valley pioneer park, NO.369 Huiyang street,071051,Baoding Hebei Province,China

T: +86 312 5909505

M: +86 137 3166 5258

Email: wendy@longfian.com

Web: www.longfian.com

Kasyep P V <kasyepcyrix@gmail.com>
To: Gavin <wendy@longfian.com>
Cc: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Wed, 10 Jul 2024 at 8:06 am

Dear sir,

We are here from Kerala , INDIA

Please let us know the shipping charges to here

Thanks & Regards,

KASYEP PV

District Incharge ,Trivandrum

CYRIX HEALTHCARE Pvt.Ltd

KLBEMP

Ph : +91 7593847134

kasyepcyrix@gmail.com

On Tue, 9 Jul 2024, 7:21 am Gavin, <wendy@longfian.com> wrote:

[Quoted text hidden]

Gavin <wendy@longfian.com>
Reply to: Gavin <wendy@longfian.com>
To: Kasyep pv <kasyepcyrix@gmail.com>
Cc: dm2.klbemp <dm2.klbemp@cyrix.in>, zm1.klbemp <zm1.klbemp@cyrix.in>

Wed, 10 Jul 2024 at 8:55 am

can you show me the address details?

Best regards

Gavin Liu

Global Marketing Manager

LONGFIAN SCITECH CO.,LTD

Add:2F&3F,East Section,Building 12,Power valley pioneer park, NO.369 Huiyang street,071051,Baoding Hebei Province,China

T: +86 312 5909505

M: +86 137 3166 5258

Email: wendy@longfian.com

Web: www.longfian.com

----- Original Message -----

From: Kasyep P V<kasyepcyrix@gmail.com>

Date: 2024-07-10 10:37

To: Gavin<wendy@longfian.com>

Cc: DM Abijith-KLBEMP<dm2.klbemp@cyrix.in>,Afsal Yunuze<zm1.klbemp@cyrix.in>

Subject: Re: Spare part for JAY-10

[Quoted text hidden]

Kasyep P V <kasyepcyrix@gmail.com>

Wed, 10 Jul 2024 at 9:14 am

To: wendy@longfian.com

Cc: dm2.klbemp@cyrix.in, zm1.klbemp@cyrix.in

Sir,

Address mentioned below;

CYRIX HEALTHCARE PVT LTD 1ST FLOOR, 30641B, KOCHI DANUSHKODI ROAD, POONITHURA, ERANAKULAM-682038

State Name : Kerala, Code: 32

[Quoted text hidden]

Gavin <wendy@longfian.com>

Wed, 10 Jul 2024 at 10:48 am

Reply to: Gavin <wendy@longfian.com>

To: Kasyep pv <kasyepcyrix@gmail.com>

Cc: dm2.klbemp@cyrix.in <dm2.klbemp@cyrix.in>, zm1.klbemp@cyrix.in <zm1.klbemp@cyrix.in>

freight cost 150USD, without clearance duty,

Best regards

Gavin Liu

Global Marketing Manager

LONGFIAN SCITECH CO.,LTD

Add:2F&3F,East Section,Building 12,Power valley pioneer park, NO.369 Huiyang street,071051,Baoding Hebei Province,China

T: +86 312 5909505

M: +86 137 3166 5258

Email: wendy@longfian.com

Web: www.longfian.com

----- Original Message -----

From: Kasyep P V<kasyepcyrix@gmail.com>

Date: 2024-07-10 11:44

To: wendy@longfian.com

Cc: dm2.klbemp@cyrix.in,zm1.klbemp@cyrix.in

Subject: Re: Re: Spare part for JAY-10

[Quoted text hidden]