



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
227845

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : THQH</p> <p>Address : Fort Kochi Emakulam</p> <p>Ph : 9946516354</p>	<p>Call Registration Date : 15/06/2024</p> <p>Caller ID : 117371</p> <p>Date of Visit : 29/07/2024</p> <p>Asset No. : 0750735</p> <p>EQPT Name : Nebuliser</p> <p>Manufacture : Omron Model : NE-C106</p> <p>S. No. 2021070117496 Dept. MM Ward</p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment fumes not getting

Action Taken : checked the equipment and found that Compressor motor with Proban Defective. Need to Replace these spares for further checking and working condition of the equipment.

Completed Date : 29/07/2024 Time : 2:00 PM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	NA	NA	NA
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
Jithu S. Kumar	29/07/2024	12:00 PM	2:00 PM

Customer Remark : NIL. Completed Pending

<p>Service Engineer Name : Jithu S. Kumar</p> <p>Signature : </p> <p>Date : 29/07/2024</p> <p>Contact Number : 7012355513</p>	<p>Customer Name : Jagiya</p> <p>Signature : </p> <p>Date : 29/07/24</p> <p>Contact Number : 9946516354</p> <p>Designation : Nursing Staff</p> <p>Hospital Seal :</p>
---	---



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	THQH FORTKOCHI ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Nebuliser Make : Omron Model : NE-C106 SN : 202107011748V
4	Equipment ID & Barcode	117371 & 0750735
5	Date of purchase / Year of manufacture /Installation Date	01-10-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15-06-2024 (Toll free)
9	Action taken	Checked and found compressor motor with piston defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1480/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Machine installed on 01-10-2017 and covered up to 6+ years. Quotation not attached since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Jithin.

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor & piston defective repaired by Cyrix. As per the asset value met BER, and spares are available reported. Spare enquiry report not attached. Submitted BER after 30 days.

Devathy
 Consultant BME
 Signature of JC BM (NHM)
 24/7/24
 Bio-Medical
 National Health
 Aranyakaram



Date

Seal

[Signature]
 Signature
 Superintendent
 Medical Officer (i/c)



Filter



BIO MEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMEP)

CYRIX

MAINTENANCE SERVICES

TOLL FREE NO. 1999 - 915 1999

BAR CODE - (8004-890615225)

0750735



Handwritten in blue ink: **WTFN**

Handwritten in blue ink: **سولیا**
Handwritten in blue ink: **میں**
Handwritten in blue ink: **میں**
Handwritten in blue ink: **میں**



OMRON COMPRESSOR NEBULIZER
MODEL: NE-C106
RATING: 230V~ 50Hz 110VA

SN 202107011748V

2021-07-01

OMRON HEALTHCARE Co., Ltd
KYOTO, 617-0002 JAPAN

EC REP

OMRON HEALTHCARE EUROPE BV
Scorpius 13, 2132 LR Honseloord, THE NETHERLANDS
Asia Pacific HQ: OMRON HEALTHCARE SINGAPORE PTE LTD

MADE IN VIETNAM

CE
121 0197

MD



Handwritten signature and text in Malayalam script.