


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC KUMBIDI
3	Name of Equipment with Make, Model and Serial Number	UPS MAKE : NA MODEL : NA SN : NA
4	Equipment ID/ Barcode	0942752 # 72346
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date = Nil (As per stock book installation date is 30/07/2019, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	02/10/2023 (Toll free)
9	Action taken	Checked the machine found not switching on. Problem with 12-230 inverter pcb, 7ah battery, 12v transformer damaged. Enquire for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	12-230 inverter board=1298/- 7ah battery = 1139/- 12v transformer = 649/- TOTAL = 3086/-

13	Asset Value	RS.4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found not switching on. Problem with 12-230 inverter pcb, 7ah battery, 12v transformer damaged. Enquire for quotation. As per stock book Equipment installed on 30-07-2019 aged up to 4 years 2 months. Repairing cost is 68.5% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Equipment installed on 30-7-2019 aged up to 4 years. Quotation submitted by the third party. Repair cost is 68.5% of the asset value. So both criteria of RBER met. Hence recommended for BER.

Signature of JC BM (NHM)

Asha
Consultant (Biomedical Engg)
NHM (Arogyakaram), Palakkad

Date:



Signature of
Superintendent/ Medical Officer

Asha
Medical Officer
Family Health Centre
Kumbidi, Palakkad (Dis)
Pin - 679553



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 217562

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC KUMBIDI</u></p> <p>Address <u>PALAKKAD</u></p> <p>Ph : <u>0466 0933900</u></p>	<p>Call Registration Date : <u>02/10/2023</u></p> <p>Caller ID : <u>72346</u></p> <p>Date of Visit : <u>03/10/2023</u></p> <p>Asset No. : <u>0942752</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture <u>NA</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>LAB</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the machine found that UPS not switching ON. Problem with 12-230 Inverter PCB, 7Ah Battery, 12V transformer damaged. Enquire for quotation.

Completed Date : 03/10/2023 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Tejas</u>	<u>03/10/2023</u>	<u>11:00 AM</u>	<u>12:00 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Tejas

Signature : [Signature]

Date : 03/10/23

Contact Number : 854809842



Customer Name : [Signature]

Signature : [Signature]

Date : 03/10/2023

Contact Number : 0466 - 2933900

Designation : Medical Officer

Hospital Seal :

Medical Officer
 Family Health Centre
 Kumbidi, Palakkad (Dist)
 Pin - 679553

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 27/07/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

600VA ONLINE UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12-230 INVERTER PCB	1	1100.00	1100.00	18	1298.00
2	7Ah BATTERY	1	890.00	890.00	28	1139.00
3	TRANSFORMER	1	550.00	550.00	18	649.00
PAYABLE AMOUNT						3086.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,
BHAGATH MOHAN





REQUIRED SERVICE PROPOSAL 0942752 # 72346 UPS @ PHC KUMBIDI , PALAKKAD

2 messages

Mon, 29 Jul, 2024 at 10:07 am

Akhilesh P <akhileshcyrix@gmail.com>

To: aionbackups12@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Please share the price and service availability of UPS @ PHC KUMBIDI, PALAKKAD

Details are furnished below

Hospital Name : PHC KUMBIDI ,PALAKKAD

Department : LABORATORY

Equipment Name : 600 VA UPS

Make : CYBER POWER

Model : NA

SN : NA

Thanks & Regards

AKHILESH P

DISTRICT INCHARGE - PALAKKAD

Bemp-Kerala

Mobile : +91 7593847154

Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD

30/641B, Petta, Punithura

Ernakulam- 682038

www.cyrixhealthcare.com

Bhagath Mohan <aionbackups12@gmail.com>

To: Akhilesh P <akhileshcyrix@gmail.com>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Thu, 1 Aug, 2024 at 9:45 am

Received your mail.

Thanks & Regards

Bhagath Mohan

PH:-8921950217

[Quoted text hidden]


Date:
Family Health Centre
Palakkad, Kumbidi • 679553



Medical Officer
Family Health Centre
Palakkad (Dist)
Kumbidi • 679553

SEMI AUTO BIOCHEMISTRY ANALYSER WITH UPS

Date		Receipt	Issued to	main stock
30/7/19.	Received from the store	1	Lab	1
30/7/19	Issued to lab	1	1	1
30/9/19	Received from store <u>UPS</u>	1	1	1
30/9/19	Issued to Job	1	1	1




Medical Officer
 Family Health-Centre
 Kumbidi, Palakkad (Dist)
 Pin - 679553