


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	DH WADAKKANCHERRY
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Make: Devilbiss Model: Airforce 1 SN:00011103
4	Equipment ID/ Barcode	0812022 #124179
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - NA Stock book installation date- 02-02-2017 (stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 19-07-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Equipment installed on 02-02-17 aged up to 7 years 5 month. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spall not available



Nam

Signature of JC BM (NHM)

Date: _____



K.P.

Signature of
Superintendent Medical Officer (c)
DISTRICT HOSPITAL, WADAKKANCHERY
KUMARANELLUR P.O. THRISUR-680 590



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 225128

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>DH Wadakkancherry</u></p> <p>Address <u>Wadakkancherry,</u> <u>Thaissua</u></p> <p>Ph : <u>9961810237</u></p>	<p>Call Registration Date : <u>19/07/2024</u></p> <p>Caller ID : <u>124179</u></p> <p>Date of Visit : <u>20/07/2024</u></p> <p>Asset No. : <u>0812022</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Devilbiss</u> Model : <u>AirForce</u></p> <p>S. No. <u>00011103</u> Dept. <u>KHRW Payward</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment is not working

Action Taken : checked the machine and found that machine is not working. found Motor and Piston are defective. Need to replace these parts for further checking.

Completed Date : 20-07-24 Time : Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin. AR</u>	<u>20/7/2024</u>	<u>12:00 PM</u>	<u>1:30 PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Aswin. AR</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/7/2024</u></p> <p>Contact Number : <u>8589965265</u></p>	<p>Customer Name : <u>Shaharban</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/7/2024</u></p> <p>Contact Number : <u>9539649635</u></p> <p>Designation :</p> <p>Hospital Seal :</p>
--	--



SUPERINTENDENT

REF MQ5900UK Air Force One
30 min ON / 30 min OFF

Zhongshan Globalcare Medical
Technology Co., Ltd

Other details refer to Instruction Manual

SN 00011103 230V - 50Hz 1A

CE 0123

Made in China

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0812022

