


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	CHC MATTATHUR
3	Name of Equipment with Make, Model and Serial Number	HEAMATOLOGY ANALYZER Make : Prokan Model :PE-6000 Sr.No- A60061700100007EH
4	Equipment ID & Barcode	#117702 – 0830473
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date -12-02-17 Actual installation date – 16-06-2017 (stock book attached for reference)
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18-06-2024
9	Action taken	Checked the machine and found 3 way valve,2way valve, CCU board, WBC chamber, RBC chamber and LCD display are complaint. Need to replace these for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	3way valve – 36900/- 2way valve – 36900/- Ccu board – 59500/- WBC chamber- 32250/- RBC chamber – 32350/- LCD display -30000/- GST - 42084/- <b>TOTAL - 275884/-</b>

13	Asset Value	199083/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	138.63%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report, and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found 3way valve, 2way valve, CCU board, WBC chamber, RBC chamber and LCD display are complaint, As per stock book Equipment installed on 16/06/2017 and aged up to 7 years, Spare cost is 138.63% of asset value and Both criteria for RBER met. As per tender clause 5.3.14.1 we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Spare cost is very high*

**Jr. Consultant (Biomedical)**  
NHM, Thrissur



Signature of JC BM (NHM)



സൂപ്രണ്ട്  
സാമൂഹിക ആരോഗ്യ കേന്ദ്രം  
മുതലൂർ, പിൻ-680 699

Signature of seal

Date 29/7/24



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 224306

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CHC</u></p> <p>Address : <u>Mattathur</u> <u>Thrishur</u></p> <p>Ph : <u>8907186408</u></p>	<p>Call Registration Date : <u>18/6/24</u></p> <p>Caller ID : <u>117702</u></p> <p>Date of Visit : <u>19/6/24</u></p> <p>Asset No. : <u>0830473</u></p> <p>EQPT Name : <u>Haematology analyser</u></p> <p>Manufacture : <u>prokan</u> Model : <u>PE-6000</u></p> <p>S. No. <u>AG0061700100007EM</u> Dept. <u>Laboratory</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** display not turn on

**Action Taken :** checked the machine and found 3 way valve, 2 way valve, CCU board, WBC chamber, RBC chamber and LCD display are complaint need to replace these for further checking and spare enquired to oem.

Completed  Date : 19/6/24 Time : 5:00PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. 3 way valve	1		
2. 2 way valve	1		
3. CCU board	1		
WBC chamber	1		
RBC chamber	1		
LCD display	1		

Cyrix Engineer	Date	Start Time	End Time
<u>Jinto Joy</u>	<u>19/6/24</u>	<u>1:00PM</u>	<u>5:00PM</u>

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> <u>Jinto Joy</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>19/6/24</u></p> <p><b>Contact Number :</b> <u>7593841125</u></p>	<p><b>Customer Name :</b> <u>Shri M. J. Mathayyan</u></p> <p><b>Signature :</b> </p> <p><b>Date :</b> <u>19/6/24</u></p> <p><b>Contact Number :</b> <u>9995648159</u></p> <p><b>Designation :</b> <u>Lab Technican</u></p> <p><b>Hospital Seal :</b> </p>
--	---



# Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)




Certificate No.: 20903/A0001/UK/En  
ISO 9001:2015  
CERTIFIED COMPANY

Proforma Invoice No.: ADPL/S/RD/20240719/24-25				Date :19-07-2024		
To, Cyrix Healthcare Pvt. Ltd. Ernakulam, First Floor, Petta Junction 30/641B, Poonithura, kerala				Ref. : By Mail.		
				Date :19-07-2024		
				CHC MATTATHUR		
Sr. No	Description	Part No.	Price Each Ex- Work	Qty	Total (In Rs.)	
A	(Instrument Model: Hematology Analyzer) (Model : PE-6000) (S/No. A60061700100007EH10)					
	<u>Spares/Consumables Procurement Details:</u>					
	(i)	3-Way Valve	PE00-006	18,450.00	2	36,900.00
	(ii)	2-Way Valve	PE00-005	18,450.00	2	36,900.00
	(iii)	CCU Board	PE60-010	59,500.00	1	59,500.00
	(iv)	WBC Chamber	PE60-003	35,250.00	1	35,250.00
	(v)	RBC Chamber	PE60-002	35,250.00	1	35,250.00
	(vi)	LCD Display	PE60-026	30,000.00	1	30,000.00
<b>Total</b>					<b>2,33,800.00</b>	
(+ GST @ 18%					<b>42,084.00</b>	
<b>Total: Two Lakh Seventy Five Thousand Eight Hundred Eighty Four Only.</b>					<b>2,75,884.00</b>	

Terms & Conditions : { Validity: 90 Days} {Delivery : 3-4 Weeks} {Payment Terms : 100% Advance }

### Bank Account Details :-

Name of the Firm : Aspen Diagnostic Pvt. Ltd.  
Bank Name : ICICI Bank Limited  
Account No. : 016005003239  
Account Type : Current Account  
MICR Code : 110229029  
IFSC Code : ICIC0000160

Aspen Diagnostics Pvt. Ltd.  
  
19/7/2024  
Rajaram Krishnamurthy  
National Service Manager





**0830473-117702- chc mattathur- for quotation**

3 messages

**Bijo T Joy** <bijocyrix@gmail.com>  
To: service support <servicesupport@rdgc.in>  
Cc: binto@rdgc.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Fri, 19 Jul, 2024 at 10:44 am

Dear sir,  
kindly provide the quotation for below mentioned Haematology Analyzer's spares.

hospital - CHC MATTATHUR  
equipment - heamatology analyzer  
quotation need - ccu board  
wbc chamber  
display  
rbc chamber  
3-way valve (2nos)  
2 way valve (2nos)

barcode - 0830473  
caller id - 117702  
sr.no - A60061700100007EH

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

**service support** <servicesupport@rdgc.in>  
To: Bijo T Joy <bijocyrix@gmail.com>  
Cc: binto@rdgc.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Fri, 19 Jul, 2024 at 1:30 pm

Dear sir,

Please share CCU Board Picture

Thanks & Regards

Hari Ram

[Quoted text hidden]

**service support** <servicesupport@rdgc.in>  
To: Bijo T Joy <bijocyrix@gmail.com>  
Cc: binto@rdgc.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Fri, 19 Jul, 2024 at 6:04 pm

Dear sir,

Please find the PI & do the needful at the earliest.

Kindly make the payment & share the details for billing.

Thanks & Regards

Hari Ram

**From:** Bijo T Joy [mailto:[bijocyraj@gmail.com](mailto:bijocyraj@gmail.com)]  
**Sent:** 19 July 2024 10:45 AM  
**To:** service support  
**Cc:** [amto@eddc.in](mailto:amto@eddc.in); Blesson Jose; DM Sarang-KLBEMP  
**Subject:** 0830473-117702- chc mattathur- for quotation

Dear sir,

[Quoted text hidden]

# Hematology Analyser & UPS

DATE	No: and Date of Invoice	From whom received	Receipt	Issued
16/6/2017	RDPL/374/17-18 13/6/2017	store	1	1
20/9/2022	Invoice no: DM/1016/22-23 Date: 12/9/2022	store	01	01
	order no: 50/568/2022-23/53 date: 31/8/2022	(BCAIS)		
		(NTHM, Tender no: KNSCC/EP/T 311/53/21 Signed order no: NTHM/4220/EP2/2022 Installation 12/10/22 Handy upto 12/11/23)		

