


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	CHC NEDUVA
3	Name of Equipment with Make, Model and Serial Number	EQUIPMENT: CENTRIFUGE Make: REMI Model: R4C Serial Number: CFLC-3142
4	Equipment ID/ Barcode	TICKET ID: 122891 BARCODE NO :1030749
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date: 09/11/2016 As per stock book installation date: 11-09--2012
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of Purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	13/06/2024
9	Action taken	Checked the machine & found problem with Motor . Need to replace motor for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	NOT WORKING
11	Recommendations for repair(required service details)	No recommendation
12	Cost of spares (specify parts and cost)	MOTOR: 6443/-

13	Asset Value	9651/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report & OEM Quotation
16	Reasons for recommending the equipment as BER	<p>Checked the machine and found Motor defective . In ticket master installation date is 09-11-2016, As per stock book the equipment was installed on 11/09/2012. Aged up to 11 Years and 10 months. Repairing cost is 66% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.</p>
17	Name & Signature of CYRIX Authority	GOKUL KT 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Verified with the damage of motor.
Beyond economic repair.*


Jr. Consultant Biomedical Engineer
NHM Malappuram
Signature of JC BM (NHM)



Date: 31/07/24


Signature of Superintendent Medical Officer (i/c)

സാമൂഹികാരോഗ്യ സെക്ഷൻ
മെട്രിക് ഓഫീസ്
കോഡ്: 676 319



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 219042

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrinx.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CHC Neduvu</u></p> <p>Address : <u>Neduvu</u> <u>Malappuram</u></p> <p>Ph : <u>7592072991</u></p>	<p>Call Registration Date : <u>13/07/24</u></p> <p>Caller ID : <u>122891</u></p> <p>Date of Visit : <u>15/07/24</u></p> <p>Asset No. : <u>1030769</u></p> <p>EQPT Name : <u>Centrifuge</u></p> <p>Manufacture <u>Remi</u> Model : <u>R5C</u></p> <p>S. No. <u>CFLC-3152</u> Dept. <u>Laboratory</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working

Action Taken : checked the machine and found problem with motor. Need to replace the motor for further checking

Completed Date : 15/7/24 Time : 1:30 Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrinx Engineer	Date	Start Time	End Time
<u>Alzal Rahman</u>	<u>15/7/24</u>	<u>12:00PM</u>	<u>1:30PM</u>

Customer Remark : _____ Completed Pending

<p>Service Engineer Name : <u>Alzal Rahman</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/7/24</u></p> <p>Contact Number : <u>9517312845</u></p>	<p>Customer Name : <u>deekshini Babu S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/7/24</u></p> <p>Contact Number : <u>7592072991</u></p> <p>Designation : <u>Lab Tech GI-I</u></p> <p>Hospital Seal : _____</p>
---	--

REMI SALES & ENGINEERING LIMITED



46/1109, Puthiyadath Road, Vytilla P.O, Ernakulam Kochi-682019
 Phone : 0484-482019, Mob: 9821409547 Email id: itsservice@remi.com, service@remisales.com

Corporate Off : Remi House, 3rd Floor, 11 Cama Inst. Estate, Goregaon (East) Mumbai 400 085, INDIA
 • Brand: +91 22 4058 8888 SALES SUPPORT NO.8879448882 • Email id : sales@remisales.com

QUOTATION CUM PROFORMA INVOICE

Q/N :80

Completed No.	Customer Name & Address:					Quotation No	PRRTRD		
Date	15/7/2024	CYRUS HEALTHCARE PVT LIM					Date	15/7/2024	
ENGINEER NAME	Vign	DISTRICT INCHARGE ,CALCUT					REF	Phone	
REF.	Phone	989114288					DATE	15/7/2024	
Email	amb@cyrus@cyrus.com	Gokul							
Model no.	B-4C	SR NO. /MFG DT.							
Sr No.	Description	Part code	MSM/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (D+E)
	ELECTRIC MOTOR (RT)	ML25000140	850300	5460	1	5460	18%	983	6443
								TOTAL (D)	6443
	GAS CHARGEING	0	0	0	0	0	18%	0	0
	SERVICE CHARGES	0	0	0	0	0	18%	0	0
								TOTAL (E)	6443
Rspees in Words:	Rs: Six thousand Four hundred forty three only .								6443

Terms & Conditions:

1. Spares are suitable only to use with REMI instruments & not for any other make.
2. We will accept return of spares only if they are unused and not damaged. At our discretion, the returned spare will be replaced. The difference in prices, if any will have to be borne by you.
3. To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
4. 100% Advance along with your order confirmation.
5. Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from receipt of cash falling which REMI will not be held responsible for any discrepancy.
6. Courier/ transportation charges are payable at destination & not included in our price.
7. Validity: The quotation is valid for 30 days from the date of this quotation / Proforma Invoice.
8. Subject to Ernakulam Jurisdiction only.

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD



BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- A/C NO: 35699154671
- IFSC CODE: SBIN0013223
- GST NO: 32AAACR0421M1Z5
- Bank Name: STATE BANK OF INDIA
- MICR CODE: 682002053
- PAN NO: AAACR0421M
- BRANCH: ELAMAKKARA
- SWIFT CODE: SBININ88395
- CIN NO. U31100MH198PL022314



Jr. Consultant Biomedical Engineer
 NHM Malappuram

മെഡിക്കൽ ഓഫീസർ
 സാമൂഹ്യാരോഗ്യ മന്ത്രം, മെട്രോ
 മെട്രോപ്പോളിറ്റൻ ഫി.ഒ - 678 119



1030749 Need price quotation for Remi R4c Motor

2 messages

Mon, 15 Jul, 2024 at 1:06 pm

Gokul K T <gokulcyrix@gmail.com>
To: instservicekochi@remigroup.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, kochi@remigroup.com

Dear Sir, Madam

Can you send the Price quotation for Centrifuge Motor Make:Remi , model :R4C, Serial number :Na . spare image attached below

Barcorde:1030749
Ticket id:122891
CEBTRIFUGE
CHC NEDUVA

GOKUL KT
CYRIX HEALTHCARE PVT Ltd
DISTRICT INCHARGE
9809314280
MALAPPURAM

Mon, 15 Jul, 2024 at 1:16 pm

instservicekochi <instservicekochi@remigroup.com>
To: Gokul K T <gokulcyrix@gmail.com>
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, kochi@remigroup.com

Dear Sir,

Greetings from REMI GROUP.

As requested attached quote for your reference. **PLEASE RELEASE THE PAYEMENT IN ADVANCE .**

Regards,

REMI KOCHI

8921409347

[Quoted text hidden]



മെഡിക്കൽ ഓഫീസർ
സാമൂഹ്യരോഗ്യ കേന്ദ്രം, മെട്രൂവ
മെട്രൂട്ടി പി.ഒ - 676 319

**Jr. Consultant Biomedical Engineer
NHM Malappuram**

CENTRIFUGE


Date	From whom to Receive or to Issue	Receipt	Issue	Balance	Remarks	Sig
2-5-2016	Brought forward from lab stocking Vol. II page no. 19			2	Received from Store 12.7.98 Crotek Centrifuge ID No 67102958 Received on 11.9.2012 Model R-4C Remi - Company. Rate: 10315/- Id. No. 67102959	QB
31-3-2022	Annual physical verification found 2 Centrifuge (Crotek and Remi) working good condition					QB
31-3-2023	Annual physical verification found 2 Centrifuge (Crotek & Remi) working good condition					QB
31-3-2024	Annual physical verification found 2 Centrifuge (Crotek & Remi) working good condition					QB




മെഡിക്കൽ ഓഫീസർ
 സാമൂഹ്യാരോഗ്യ കോശം, മംഗലൂർ
 മെട്രിക്സ് റി.ഒ. - 676 319

Jr. Consultant Biomedical Engineer
 NHM Malappuram



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP) 

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 
1030749

REMI CENTRIFUGE CE

CAT. NO./DATE :	R-4C	3/12
SERIAL NO :	CEIC-31175	
R.P.M :	4400	AMPS.: 0.25
SUPPLY :	220 /230 V, 50Hz 1ØAC	

REMI ELEKTROTECHNIK LIMITED
 (INSTRUMENTS DIVISION)
 VASA - 401 208 (INDIA)


Jr. Consultant Biomedical Engineer
NHM Malappuram




മാധ്യമങ്ങൾ ഓഫീസർ
സാമൂഹ്യസേവ്യ കേന്ദ്രം, നെടുവ
ചെട്ടിപ്പുഴ, പി.ഒ - 676 119