


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	THQH OTTAPALAM
3	Name of Equipment with Make, Model and Serial Number	DIALYSIS MACHINE Make : B BRAUN Model: DIALOG+ SN: 106340
4	Equipment ID/ Barcode	0921333 #114897
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil (as per stock book installation date 21-07-2013, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03-06-2024 toll free
9	Action taken	Checked the machine Found not switching on. Problem with smps, pressure sensor, motor for gear pump, gear pump ep, fpe, fpa, membrane for balance chamber, lx800 mother board, compact flash card lgb, analog board, digital board version 6, pressure monitoring module, retainer tube green, heparin pump defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Switched mode power supply mc2 = 189332/- Pressure sensor = 28332/- Motor for gear pump = 39363/- Gear pump ep = 64988/- Gear pump fpe = 50413/- Gear pump fpa = 48007/- Membrane for balance chamber = 9323/- Lx800 mother board = 115895/- Compact flash card lgb = 12986/- Analog board = 60230/- Digital board version 6 = 82308/- Pressure monitoring module = 50463/- Retainer tube green = 1405/- Heparin pump = 90123/- Total = 843168/- GST 18% = 151770/- Total amount = 994938/-


13	Asset Value	RS. 506667/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	196%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine Found not switching on. Problem with smps, pressure sensor, motor for gear pump, gear pump ep, fpe, fpa, membrane for balance chamber, lx800 mother board , compact flash card lgb, analog board, digital board version 6, pressure monitoring module, retainer tube green, heparin pump defective Enquired OEM for quotation. equipment installation date as per stock book 21-07-2013 aged up to 10 years 11 months. Repairing cost is 196% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 Equipment installed on 21-7-2013 aged up to 10 years. Quotation attached. ~~the~~ The spare cost exceeds the actual value of machine's cost. Both criteria of RBER met. Hence recommended for BER.

  
 Signature of JC BM (Consultant, Biomedical Engg) NHM (Arogyakeralam), Palakkad  
 26/07/24

Date: \_\_\_\_\_



  
 Superintendent  
 Taluk Head Quarters Hospital  
 Ottapalam  
 Signature of Superintendent/Medical Officer (i/c)





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 207768

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>THQH</u> .....</p> <p>Address ..... <u>Ottapalam</u> .....</p> <p>..... <u>Palakkad</u> .....</p> <p>Ph : ..... <u>8590109226</u> .....</p>	<p>Call Registration Date : <u>03-06-2024</u> .....</p> <p>Caller ID : <u>114897</u> .....</p> <p>Date of Visit : <u>05-06-2024</u> .....</p> <p>Asset No. : <u>0921333</u> .....</p> <p>EQPT Name : <u>Dialysis Machine</u> .....</p> <p>Manufacture <u>B. Braun</u> Model : <u>Dialog plus</u> .....</p> <p>S. No. <u>106340</u> Dept. <u>Dialysis unit</u> .....</p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working properly .....

**Action Taken :** checked the machine found not switching on. Problem with smps, pressure sensor, motor for gear pump, gear pump c.p.tipe, fixa membrane for balance chamber, 1x800 motor board, compact flash card, i.g.b, analog board, digital board version 6, pressure monitoring module. Retainer tube given. hepania pump defective. Enquired OEM for quotation

Completed  Date : 05-6-2024 Time : 1:15 p.m. Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Midhuryak</u>	<u>05-6-24</u>	<u>11:30 am</u>	<u>1:15 pm</u>

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> <u>Midhuryak</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>05-6-24</u></p> <p><b>Contact Number :</b> <u>9072522081</u></p>	<p><b>Customer Name :</b> <u>ROSEY DAS GEORGE</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>5/6/2024</u></p> <p><b>Contact Number :</b> <u>9446488071</u></p> <p><b>Designation :</b> <u>SENIOR NURSING OFFICER</u></p> <p><b>Hospital Seal :</b> <u>Taluk Head Quarters Hospital Ottapalam</u></p>
--	--



**Superintendent**  
Taluk Head Quarters Hospital  
Ottapalam





### Service report

Service ETS Request No. <b>49963</b>
Date of Visit : <b>11-7-2024</b>

Installation

PMS / TSC

Break Down

Customer Name : <b>Taluk Head quarters Hospital</b>		Service Request Category:	
Customer Address : <b>Ottapalam</b>		Installation	
Contact Number :		User Training checklist signed:	<input type="checkbox"/> Y <input type="checkbox"/> N
Division BA <input checked="" type="checkbox"/> HC <input type="checkbox"/> AE <input type="checkbox"/>		Installation checklist signed:	<input type="checkbox"/> Y <input type="checkbox"/> N
Service Place Onsite <input checked="" type="checkbox"/> Workshop <input type="checkbox"/> Office <input type="checkbox"/>		Warranty End Date:	(YYYY / MM / DD)
Model : <b>Dialog + EVO</b>		TSC cum PMS	
Serial Number : <b>106340</b>		If Yes PM KIT replacement date:	(YYYY / MM / DD)
Software version : <b>9.12</b> Working hours:		User Training:	<input type="checkbox"/> Y <input type="checkbox"/> N
Observations :		User Training checklist signed:	<input type="checkbox"/> Y <input type="checkbox"/> N
RO TDS: (0-10PPM Acceptable limits) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Breakdown call	
Neutral to Earth Voltage: (0-2 Volt)		Software:	<input type="checkbox"/> Y <input type="checkbox"/> N
Use of UPS / Stabilizer: (recommended 3KVA per machine in Dialog model) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Electrical Hardware :	<input type="checkbox"/> Y <input type="checkbox"/> N
Bicarb mixer : <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Mechanical Hardware:	<input type="checkbox"/> Y <input type="checkbox"/> N
Reported Problem: <b>machine not switching on</b>		Trend backup: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	Service Type
Observed Problem:		50% Citric Disinfection after each Therapy: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	Warranty / Extended warranty: <input type="checkbox"/>
Service Resolution: <b>check the machine need to replace mention spare - SMPS-1, Pressure sensor-2, F.P. Gear Pump-motor-1, Retainer tube green-1, Gear Pump RPE, FPA-1 membran for balance chamber-1, Digital board-1, Analog board-1, PA/PV module-1, Mother board-1, CF card-1, Heparin pump-1</b>		50% Citric Disinfection after end of day: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	CAMC : <input type="checkbox"/>
Following spareparts are :		To be replaced : <input type="checkbox"/> Y <input type="checkbox"/> N	
Replaced : <input type="checkbox"/> Y <input type="checkbox"/> N		To be quoted: <input type="checkbox"/> Y <input type="checkbox"/> N	

SN	Article number	Article Name	Quantity
		SMPS, 1+800 mother board -1	
		Analog board-1, digital board-1	
		Gear pump- RPE, FPA, FP - 1 each	
		Pressure sensor-1, PA/PV modul-1	

Date (YYYY / MM / DD)	Start time (Hrs : Min)	End time (Hrs : Min)	Labour Hours Spent: (Hrs : Min)

This is to certify that above mentioned spareparts of B.Braun machine are consumed / to be consumed on our in hospital for satisfactory working of machine.

Status of Service request:	Completed: <input type="checkbox"/> Y <input type="checkbox"/> N	Under Observation: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	Followup Date : (YYYY / MM / DD)
Contact person:	Signature with Date & Time	Stamp	
<b>Nimisha Asokan</b>	<b>11/7/24</b>		
Customer remark / Feedback:			
Engineer name & Signature: <b>D. J.</b>	Engineer Contact No.:		



To,  
Cyrix Healthcare  
C/O Ottapalam

**Sub: - Quotation for Spare Parts**

**Our Ref.BB/BA-W-220/24 Date: 17.07.2024**

Dear Sir/ Madam,

Greetings from BBraun Medical India Pvt. Ltd.

Kindly find our most competitive rate for Spare part for Hemodialysis machine.

Art. No.	Description	Qty	Unit price ( Excl. Tax)	Total Price ( Excl. Tax)
3461076	Switch mode power supplyMc2	1	189,332	189,332
3456124A	Pressure Sensor	2	14,166	28,332
34560718	Motor for gear pump	1	39,363	39,363
3461097	Gear pump Ep	1	64,988	64,988
34562331	Gear pump Fpe	1	50,413	50,413
34562330	Gear pump Fpa	1	48,007	48,007
3451615A	Membrane for balance chamber	1	9,323	9,323
34560610	LX800 motherboard	1	115,895	115,895
3456088A	Compact flash card 1 GB	1	12,986	12,986
34560734	Analog board	1	60,230	60,230
34562303	Digital board, version 6	1	82,308	82,308
34561641	Pressure monitoring module	1	50,463	50,463
3456077A	Retainer tubes (green)	1	1,405	1,405
3451432D	HEPARIN PUMP COMPACT	1	90,123	90,123
			<b>Total:</b>	<b>843,168</b>
			<b>GST: 18%</b>	<b>151,770</b>
			<b>Total Amount:</b>	<b>994,938</b>





**Purchase Order-KL 24 25 101031 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0921333 -THQH  
OTTAPALAM**

11 messages

Anjitha T L <purchasecyrix2@gmail.com>

Fri, 5 Jul, 2024 at 11:55 am

To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>  
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir,

Please find the attached Purchase Order and do the needful. Kindly send a confirmation mail with details of material delivery (Docket details) along with the invoice Copy mentioning our PO No. (For reference) with Seal and Signature.

EDD as per payment term: 10/07/24

*For Queries and Clarifications*

1. Technical: [dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in)
2. Accounts: [paymentcyrix@gmail.com](mailto:paymentcyrix@gmail.com)
3. Logistics: [logistics@cyrix.in](mailto:logistics@cyrix.in)
4. PO Clarification: [ampurchase@cyrix.in/purchasemgr@cyrix.in](mailto:ampurchase@cyrix.in/purchasemgr@cyrix.in)
5. Delivery address: **Cyrix Healthcare PVT LTD**

**Methanam Road  
Door No: 312/a Eloor North,  
Kerala – 683501- India  
Nikesh V  
Mob: 9072522014**

**@ Cyrix Accounts**

Payment Term: 100% adv

Thanks & Regards

Anjitha T L



Sr. Purchase Executive  
PH:- 9947719111  
[purchasocyrix2@gmail.com](mailto:purchasocyrix2@gmail.com)  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Cyrix Healthcare Pvt Ltd**  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Mon, 8 Jul, 2024 at 11:07 am

**Babu Murugesan** <[babu.murugesan@bbraun.com](mailto:babu.murugesan@bbraun.com)>  
To: Anjitha T L <[purchasocyrix2@gmail.com](mailto:purchasocyrix2@gmail.com)>, Anoop <[anoop.karthikeyan@bbraun.com](mailto:anoop.karthikeyan@bbraun.com)>  
Cc: Cyrix Purchase Mgr. <[purchasemgr@cyrix.in](mailto:purchasemgr@cyrix.in)>, Project Manager KLBEMP <[pm.bempkl@cyrix.in](mailto:pm.bempkl@cyrix.in)>, Assistant Manager Purchase <[ampurchase@cyrix.in](mailto:ampurchase@cyrix.in)>, Fathima Service Coordinator <[bempservicekl2@gmail.com](mailto:bempservicekl2@gmail.com)>, Anagha <[bempservicekl3@gmail.com](mailto:bempservicekl3@gmail.com)>, Cyrix Payments <[paymentcyrix@gmail.com](mailto:paymentcyrix@gmail.com)>, Blesson Jose <[zm2.klbemp@cyrix.in](mailto:zm2.klbemp@cyrix.in)>, DM Sarang-KLBEMP <[dm1.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in)>, Cyrix Logistics <[logistics@cyrix.in](mailto:logistics@cyrix.in)>, Kerala GRN <[keralagr@cyrix.in](mailto:keralagr@cyrix.in)>, Izrath K K <[purchasocyrix5@gmail.com](mailto:purchasocyrix5@gmail.com)>, Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>

Dear Anjitha,

Please share payment transfer details.

Thanks & Regards  
**Babu Murugesan**

**B. Braun India is recognized as one of the Best Workplaces to Work in India**



**From:** Anjitha T L <[purchasocyrix2@gmail.com](mailto:purchasocyrix2@gmail.com)>  
**Sent:** Friday, July 5, 2024 11:55 AM  
**To:** Anoop <[anoop.karthikeyan@bbraun.com](mailto:anoop.karthikeyan@bbraun.com)>; Babu Murugesan <[babu.murugesan@bbraun.com](mailto:babu.murugesan@bbraun.com)>  
**Cc:** Cyrix Purchase Mgr. <[purchasemgr@cyrix.in](mailto:purchasemgr@cyrix.in)>; Project Manager KLBEMP <[pm.bempkl@cyrix.in](mailto:pm.bempkl@cyrix.in)>; Assistant Manager Purchase <[ampurchase@cyrix.in](mailto:ampurchase@cyrix.in)>; Fathima Service Coordinator <[bempservicekl2@gmail.com](mailto:bempservicekl2@gmail.com)>; Anagha <[bempservicekl3@gmail.com](mailto:bempservicekl3@gmail.com)>; Cyrix Payments <[paymentcyrix@gmail.com](mailto:paymentcyrix@gmail.com)>; Blesson Jose <[zm2.klbemp@cyrix.in](mailto:zm2.klbemp@cyrix.in)>; DM Sarang-KLBEMP <[dm1.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in)>; Cyrix Logistics <[logistics@cyrix.in](mailto:logistics@cyrix.in)>; Kerala GRN <[keralagr@cyrix.in](mailto:keralagr@cyrix.in)>; Izrath K K <[purchasocyrix5@gmail.com](mailto:purchasocyrix5@gmail.com)>; Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>  
**Subject:** Purchase Order-KL 24 25 101031 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0921333 -THQH OTTAPALAM

**CAUTION:** This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

**Babu Murugesan**

Manager - Technical Services  
+919884450417  
[babu.murugesan@bbraun.com](mailto:babu.murugesan@bbraun.com)

B. Braun Medical (India) Pvt. Ltd. (CIN : U33112MH1984PTC214514)  
WorkeZ - Block-B, Door No.3, 2nd Floor, Hansa Building , RK Swamy Centre, 147, Pathari Rd, Thousand Lights, Chennai, 600006  
Tamil Nadu, IN, India



Chennai, 600006  
India

<http://www.bbraun.co.in>  
[www.linkedin.com/company/bbraun-group](http://www.linkedin.com/company/bbraun-group)

As a true partner  
we develop smart solutions  
and set standards  
to improve health care.

B. Braun – Sharing Expertise

The information contained in this communication is confidential, may be attorney-client privileged, may constitute inside information, and is intended only for the use of the addressee. It is the property of the company of the sender of this e-mail. Unauthorized use, disclosure, or copying of this communication or any part thereof is strictly prohibited and may be unlawful. If you have received this communication in error, please notify us immediately by return e-mail and destroy this communication and all copies thereof, including all attachments.

**Izrath K K** <purchasocyrix5@gmail.com> Mon, 8 Jul, 2024 at 11:40 am  
To: Cyrix Payments <paymentcyrix@gmail.com>  
Cc: Anjitha T L <purchasocyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Akhilesh P <akhileshcyrix@gmail.com>, Babu Murugesan <babu.murugesan@bbraun.com>

Dear tincy ,

Kindly share the UTR Details against the PO.

Thanks & Regards  
Izrath KK  
Jr Purchase Executive  
PH: 8848002238

[Quoted text hidden]

**Cyrix Payments** <paymentcyrix@gmail.com> Mon, 8 Jul, 2024 at 1:59 pm  
To: Anoop <anoop.karthikeyan@bbraun.com>  
Cc: Izrath K K <purchasocyrix5@gmail.com>, Anjitha T L <purchasocyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Akhilesh P <akhileshcyrix@gmail.com>, Babu Murugesan <babu.murugesan@bbraun.com>

Dear Sir/Mam,

Please note that we had transferred Rs.13920 /- as advance against PO 101031 after deducting TDS Rs.240/-

B. Braun Medical (India) Pvt. Ltd - Mumbai	0062620000	13,920.00	AXISP00516133497	PAID	05-JUL-24
---	------------	-----------	------------------	------	-----------

[Quoted text hidden]

Thanks & Regards,

Tincy T Jose

Accounts Department.

## Cyrix Healthcare Pvt Ltd

Office : 30/641 B, First Floor, Petta Junction, Poonithura, Cochin, Kerala - 682 038

Website:- [www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Mob : 7593847123

Babu Murugesan <babu.murugesan@bbraun.com>

Mon, 8 Jul, 2024 at 2:14 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Anjitha,

Please provide machine serial number.

Always mention the same in your PO, so that it will be easy for us track.

Thanks & Regards

Babu Murugesan

B. Braun India is recognized as one of the Best Workplaces to Work in India



From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Friday, July 5, 2024 11:55 AM

To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>

Subject: Purchase Order-KL 24 25 101031 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0921333 -THQH OTTAPALAM

**CAUTION:** This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

[Quoted text hidden]

Anjitha T L <purchasecyrix2@gmail.com>

Mon, 8 Jul, 2024 at 2:17 pm

To: Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-



KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasocyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear sir,

Please find the serial No

Serial No: 106340

[Quoted text hidden]

[Quoted text hidden]

---

Akhilesh P <akhileshcyrix@gmail.com>

Sat, 20 Jul, 2024 at 8:18 am

To: Anjitha T L <purchasocyrix2@gmail.com>

Cc: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasocyrix5@gmail.com>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)  
[Quoted text hidden]

---

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 24 Jul, 2024 at 1:22 pm

To: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>  
Cc: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasocyrix5@gmail.com>, Anjitha T L <purchasocyrix2@gmail.com>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)  
[Quoted text hidden]

---

Babu Murugesan <babu.murugesan@bbraun.com>

Wed, 24 Jul, 2024 at 5:12 pm

To: Anjitha T L <purchasocyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>  
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix

Logistics & Purchasing <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Anjitha,

Please find Service report, Spares estimate & invoice copy in the attached file.

Thanks & Regards  
Babu Murugesan

**B. Braun India is recognized as one of the Best Workplaces to Work in India**



**From:** Anjitha T L <purchasecyrix2@gmail.com>

**Sent:** Friday, July 5, 2024 11:55 AM

**To:** Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

**Cc:** Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>

**Subject:** Purchase Order-KL 24 25 101031 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0921333 -THQH OTTAPALAM

**CAUTION:** This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

[Quoted text hidden]

**Izrath K K** <purchasecyrix5@gmail.com>

Wed, 24 Jul, 2024 at 5:36 pm

To: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Hi,

Kindly take GRN

po:101031

@Akhilesh P kindly give confirmation

Thanks & Regards  
Izrath KK  
Jr Purchase Executive  
PH: 8848002238

[Quoted text hidden]



**Akhilesh P** <akhileshcyrix@gmail.com>  
To: Izrath K K <purchasecyrix5@gmail.com>

Wed, 24 Jul, 2024 at 10:03 pm

Cc: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Dear mam

PO 101031 one time visit done

Thanks & Regards

**AKHILESH P**  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

**CYRIX HEALTHCARE (P) LTD**  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

[Quoted text hidden]

HAEMO DIALYSIS MACHINE.

11


	Tr in from Old Reg P. No.	RD	Issued	Balance	M. Stock
21/7/2013	Rd from Store (B arm)	5		5	
	(man stock B6-PNU 48) in 28,19,250/-				
4/11/2013	Rd from Store (B arm)	2		7	7
	(Install dt 15/7/2013)				
5/11/2013	Used to unit		7	Nil	7
3/8/2017	Rd from Store	2		2	9
8/9/2017	Used to unit		2	Nil	9
	(man stock B6-10) dt 3/8/2017				
	Nipro 2 machine 4,74,600 x 2				
	Install dt on 8/9/17.				
7/5/2018	Rd from Store	10		10	19
	(M-stock B6-10) in 4,04,920 +				
	10,1248 (2 machine) 8 machine				
	NIPRO model 55plus				
	SLNO: 186 x 9247, 9867, 176 x 9045				
	176 x 9041, 8947, 9043, 8957,				
	9037, 9136, 9039.				
	Warranty from 6-2-2018 to 5/4/2021				
8/5/2018	Used to unit		10	Nil	19
30/5/2019	Rd from Store (Previous)	5		5	24
	Warranty - 2019-2020.				
	(entire Old Stock Reg P No. 30)				
30/5/19	Used to unit.		5	Nil	
30/12/2020	Rd from Store	5			29
	(Cost 5,63,360 x 5)				
	Total Bill 4,50,000				
5/1/2021	Used to unit		5	Nil	29
21/7/22	Grant to Unserviceable Register				
	B.No. 7, Page No. 48				
	(67093026, 67093027) 2 machines				
	Inst. dt 15/7/13				
	B. No. - 106619				
	" - 106331				
	" - 106370				



Superintendent  
Taluk Head Quarters Hospital  
Ottapalam

(25/25) 25  
25/25  
25/25





*[Handwritten signature]*

Superintendent  
Taluk Head Quarters Hospital  
Otapalam



# Dialog +

Aparelho para Hemodiálise


REF 710200N  
SN 106340  
Bj 2012  
VA 2500  
V 230~  
Hz 50/60  
IP 21 CE 0123

Manufacturer/Fabricado por:  
**B.Braun Avitum AG**  
34209 Melsungen, Germany

Made in Germany

Importado e Distribuido por: B. Braun  
Laboratórios B. Braun S.A.  
Brasil Registro ANVISA n°.8.01369.90498

Not to be used in presence of  
flammable anaesthetic gases



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)  
0921333



*[Handwritten signature]*  
Superintendent  
Taluk Head Quarters Hospital  
Ottapalam