


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	IDUKKI
2	Name of Hospital	CHC MUTTOM
3	Name of Equipment with Make, Model and Serial Number	Equipment name: Laboratory Microscope Make: LABOMED Model: ATC2000 Serial number:na
4	Equipment ID & Barcode	ID: 124319 Barcode: 0630657
5	Date of purchase / Year of manufacture /Installation Date	6/08/2011
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20-07-2024 (Toll-free)
9	Action taken	Regarding the inspection of the equipment found that the stage assembly, prism, eyepiece,100x lens,40x lens, and 10x lens are defective. Enquired spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Na




13	Asset Value	34440
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	The Cyrix service report and End of life letter are attached.
16	Reasons for recommending the equipment as BER	Checked and found that the stage assembly, prism, eyepiece, 100x lens, 40x lens, and 10x lens are defective. Equipment outlived 12+ years. spares are not available now. OEM declared the end of life in this machine. So recommend RBER as per tender clause 5.3.14.2
17	Name & Signature of CYRIX Authority	Jithu Abraham 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Consultant Biomedical Engineer
National Health Mission
Arogyakeralam
Idukki-685603


6/8/24
Signature of JC BM (NHM)

Date _____ Seal _____


Signature of Medical Officer
സാമൂഹിക ആരോഗ്യകേന്ദ്രം
സാമൂഹിക ആരോഗ്യകേന്ദ്രം
മുട്ടം, തിരുവനന്തപുരം



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
202915

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>CHC</u>	Call Registration Date : <u>20171004</u>
Address <u>Muttum, Theedupuzha</u>	Caller ID : <u>124319</u>
..... <u>Jolankal</u>	Date of Visit : <u>22171004</u>
Ph : <u>8943341400</u>	Asset No. : <u>0630657</u>
	EQPT Name : <u>Laboratory microscope</u>
	Manufacture <u>Larsomed</u> Model : <u>AP10006</u>
	S. No. <u>Nil</u> Dept. <u>LAS</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working

Action Taken : Check the machine. found that stage assembly prism, load, box, box, eye piece, are defective. need to replaced these spares. done further checking working condition of these equipment. → ordered spares with OEM.

Completed Date : 20171004 Time : 10am Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Edison Shine</u>	<u>20171004</u>	<u>9am</u>	<u>10am</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Edison Shine</u>	Customer Name : <u>Smithe, R.S</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>20171004</u>	Date : <u>20171004</u>
Contact Number : <u>7034246213</u>	Contact Number : <u>9496479672</u>
	Designation : <u>Lab Technician</u>
	Hospital Seal : <u>[Seal]</u>



[Signature]
ചെയ്തതും അറിയാതെ ഇന്ന് ചെയ്തതും
സാമൂഹിക ആനുകൂല്യം
2017.10.04



Abijith A <abijithacyrix@gmail.com>

Re: Spare Request for Labomed ATC 2000 Microscope - Reg.- 96938 0731730

2 messages

JIBIN THOMAS GEORGE <jibintcyrix@gmail.com>

Thu, Apr 4, 2024 at 8:25 AM

To: uniage@gmail.com

Cc: es.kmscl@kerala.gov.in, Abijith A <abijithacyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Gentle reminder

On Sun, Mar 31, 2024, 9:24 PM JIBIN THOMAS GEORGE <jibintcyrix@gmail.com> wrote:
Gentle Reminder....

----- Forwarded message -----

From: **JIBIN THOMAS GEORGE** <jibintcyrix@gmail.com>

Date: Fri, Mar 22, 2024, 12:04 PM

Subject: Re: Spare Request for Labomed ATC 2000 Microscope - Reg.

To: <uniage@gmail.com>

Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Abijith A <abijithacyrix@gmail.com>

Gentle reminder...

On Mon, Mar 18, 2024, 1:34 PM JIBIN THOMAS GEORGE <jibintcyrix@gmail.com> wrote:
Gentle reminder....

Dear Sir,

Please provide the estimation of the following spares for repairing the Labomed ATC 2000 Microscope of CHC Varapetty, Ernakulam.

Barcode No 0731730

Caller ID 96938

S: No:120346906

- * 100x , 40x, lens
- * Eyepiece -2Nos.
- * Condenser unit
- * Stage
- * Prism
- * Light with Power supply unit

Kindly send the estimation for the same as soon as possible.

Regards,

Jibin Thomas George
District Incharge
KLBEMMP-ERNAKULAM
Cyrix health care Pvt. Ltd
MOB: 9047958051

On Thu, Mar 14, 2024, 10:23 AM JIBIN THOMAS GEORGE <jibintcyrix@gmail.com> wrote:
Gentle Reminder

Please provide the Spare estimation for Labomed ATC 2000

Regards,

Jibin Thomas George
District Incharge - Ernakulam
KLBEMMP-ERNAKULAM
Cyrix health care Pvt. Ltd
MOB: 9047958051

On Tue, Mar 12, 2024, 11:05 PM JIBIN THOMAS GEORGE <jibintcyrix@gmail.com> wrote:
Dear Sir,

Please provide the estimation of the following spares for repairing the Labomed ATC 2000 Microscope of
CHC Varapetty, Ernakulam.
Barcode No 0731730
Caller ID 96938
S: No:120346906

- * 100x , 40x, lens
- * Eyepiece -2Nos.
- * Condenser unit
- * Stage
- * Prism
- * Light with Power supply unit

Kindly send the estimation for the same as soon as possible.

Regards,

Jibin Thomas George
District Incharge
KLBEMMP-ERNAKULAM
Cyrix health care Pvt. Ltd
MOB: 9047958051

Universal Agencies <uniage@gmail.com>

To: JIBIN THOMAS GEORGE <jibintcyrix@gmail.com>

Thu, Apr 4, 2024 at 10:54 AM

Cc: es.kmscl@kerala.gov.in, Abijith A <abijithacyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sir

Please find the attached Letter.

Warm Regards,
M/S. Universal Agencies
(ISO 9001 : 2008 Certified)
H.O - "Palakkal Angadi, Thrissur -680001 Kerala - India"
Ph No: 0487-2440319/2427031
Cel : 8943341568
GST : 32AAAFU5662Q1ZQ
Email : uniage@gmail.com
uniage.sales@gmail.com

Branch Office: " 2nd Floor, Pride of Aristo,
C.S.I Cemetery Road, Mission Quarters
East Fort, Thrissur- 680005"
Tel: 0487-2427038/8943112277
Email: uniage.etender@gmail.com

 **LETTER ATC.pdf**
149K

UNIVERSAL AGENCIES

ISO 9001 : 2008

Vision for excellence



PALAKKAL ANGADI, THRISSUR - 680 001, KERALA
0487 - 2427038, 2427031, Mob : 93870 211122, 8943341569
Email : uniage@gmail.com, uniage.sales@gmail.com

UA-L. 009/2024

03/04/2024

DISTRICT INCHARGE
KLBEMMP-ERNAKULAM
CYRIX HEALTH CARE PVT. LTD

Dear Sir

Sub: Service of Microscope Model ATC 2000

With reference to above and we wish to inform you that the further repair is not possible on seeing the condition of the equipment. Also due to the non availability of the spare parts in the market as this is an old model. Hence we would recommend buying a new machine.

Sorry for the in convenience caused to you in this regards.

Thanking You



Yours Faithfully,
For Universal Agencies,

[Signature]
Manager.

Keeler memmert BIOWASE
UK GERMANY CHINA



LABOMED
Ideas for vision



SLEE
GERMANY

ELECTROLAB
Your Quality, Our Assurance




21. Ph: 0487 2420367
MES CALIBRATION CENTRE

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0630657



[Handwritten Signature]

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
സാമൂഹിക ആരോഗ്യകേന്ദ്രം
മുട്ടം, തൊടുപുഴ

MICROSCOPE

11/1/15

1) RNECP LABMED 100128018
VISION 2000

Bar code 68062487 (Transferred to CAC. Puzamp
on 7/1/2022)

2) KMSC LABMED NIC-2000 13-3-2010

dt of installation. 6/8/2011

Bar code 68062486

25/1/16

Funstore LABMED LX400 01 Bar code 03
68062487

Model LX 400. Rate 46758-00 S.No. 150782267

Installation start date. 17/2/16

Warranty started 17/2/16 - End. dt. 17/2/19.

25/5/16

complained and reported.

Replace the board on 26/9/16 and working
Satisfactory

25/2/2020

re B. K. Sane's microscope. Type III 01 04.

Model VISION 2000

Receipt 197/12

Make: N. K. Sane Instruments Pvt Ltd.

Warranty 25-2-2020 to 25-2-2022. Re 197/12

(Signature)



(Signature)
 സാമൂഹിക ആരോഗ്യ സേവന വകുപ്പ്
 സാമൂഹിക ആരോഗ്യകേന്ദ്രം
 എ.ടി. മൊട്ടാപ്പള്ളി