



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
233022

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u> <u>Kerala</u></p> <p>Ph : <u>9447665968</u></p>	<p>Call Registration Date : <u>27-7-24</u></p> <p>Caller ID : <u>125897</u></p> <p>Date of Visit : <u>31-7-24</u></p> <p>Asset No. : <u>0110614</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Teuchek</u> Model : <u>N8001</u></p> <p>S. No. <u>B080011709195</u> Dept. <u>General ward</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working.

**Action Taken :** Checked and found Nebulizer motor and piston damaged. Need to replace these spares for further checking and working condition of equipment.


Completed  Date : 31/7/24 Time : 10:30 Am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sooraj Sudheep</u>	<u>31/7/24</u>	<u>10:00 Am</u>	<u>10:30 Am</u>

**Customer Remark** Completed  Pending



**Service Engineer Name :** Sooraj Sudheep  
**Signature :** [Signature]  
**Date :** 31/7/24  
**Contact Number :** 7907406244

**Customer Name :** Jessy Jayhese  
**Signature :** [Signature]  
**Date :** 31/7/24  
**Contact Number :** 9447465968  
**Designation :** SN  
**Hospital Seal :**

Superintendent  
General


578

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULIZER Manufacturer : Truechek Model : N8001 Sl: NA
4	Equipment ID/ Barcode	125897/ 0110614
5	Date of purchase/ Year of manufacture/Installation Date	19/10/2017
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	27/07/2024 (Toll free)
9	Action taken	Checked the unit and identified motor and piston defective .Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor and piston defective. The equipment was installed on 19/10/2017 and covered upto 6+ years. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Soorya Sudheep 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston defective. So it can be RBER as per guidelines.

  
 7.08.2024  
 Signature of JC BM (NHM)

**MANEESHA MOHAN M.R.**  
 Junior Consultant (Biomedical)  
 National Health Mission  
 Thiruvananthapuram


Date:



Signature of  
 Superintendent/Medical Officer (I/c)

Superintendent  
 General Hospital  
 Thiruvananthapuram

		R	L	P	S/F
15/1/17	RF from SBL RFP P-NO 183 (OAP- NOT KNOWN)			2	10
19/11/17	Received from the store 19/11/17 Page No: 8B- 110 Rs 1027,50. Washable oil	1		3	short
28/11	Rd from store. (Page No: 110. IR 813) Rf.	1		4	
5/9/19	(Rs- 1902. 32+Tax) Rd from stock condemnation list on paper order NOC <sub>2</sub> -216/2022 CIVT (2019 June) dated-26/7/22	2		5	3
15/11/2019					
8/4/2021					
23/3/22	79.10 from Ward 5 (stock up) Vol- D Page No 12	1		4	
31/5/22	R.P.S (LP) <del>deposited</del> B. NEB-03-187-d-1/1/22-Exp-1/1/2030	2		6	
28/4/22	RFS (LP) Harden tech vt (K4) mfg No (batch 053e) mfg 1/6/22 Exp -1/6/30	1		7	
8/2/23	Received from store (LP) [Fruchek] Batch No 20220826 (A20) 01/11/2022 Exp date 1/11/32. Rs 1027.50/-	1		8	8
	Condemnation done as per order No: G <sub>2</sub> -5549/23/GHS dated 13/2/24			1	7
20/7/2023	Received from store (L.P) GHS/01318/24-25 dated 20/7/24 800 Batch no: 1.B.500 Supp No: Suv mfg: 1/1/24. Exp: 31/1/29	2		9	



if  
Saw

General Hospital  
Thiruvananthapuram

