



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 229072

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>PHC</u></p> <p>Address ..... <u>PHC Choorikkara</u> <u>Ernakulam</u></p> <p>Ph : ..... <u>9656887284</u></p>	<p>Call Registration Date : <u>18/07/2024</u></p> <p>Caller ID : ..... <u>123981</u></p> <p>Date of Visit : ..... <u>19/07/2024</u></p> <p>Asset No. : ..... <u>0747126</u></p> <p>EQPT Name : ..... <u>BP Apparatus</u></p> <p>Manufacture ..... <u>Omron</u> ..... Model : <u>H6M 7120</u></p> <p>S. No. <u>2017062254646</u> Dept. .... <u>OP</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Bp Apparatus not working

**Action Taken :** checked the machine and found main board, valve and motor gone defective. Need to replace these spare for further checking and working condition of the equipment

Completed  Date : 19/07/2024 Time : 3:00pm Spare Required

Spare Replaced  Requested

No.	Description	Qty.	Part Number	PR Number
1.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Libin Sabu</u>	<u>19/07/2024</u>	<u>1:00PM</u>	<u>3:00PM</u>

Customer Remark  Completed  Pending



*Medical Officer in-charge*

Service Engineer Name : Libin Sabu  
Signature : Libin  
Date : 19/07/2024  
Contact Number : 9562970090

Customer Name : Jancy  
Signature : [Signature]  
Date : 19/07/2024  
Contact Number : 9656887284  
Designation : N/O  
Hospital Seal : [Signature]

For Dr. Sreedha


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC CHOORNIKKARA ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Bp Apparatus Make : Omron Model : HEM-7120 SN :20170622546VGI
4	Equipment ID & Barcode	123981 & 0747126
5	Date of purchase / Year of manufacture /Installation Date	01-09-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18-07-2024 (Toll free)
9	Action taken	Checked and found main board, valve and motor defective . Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA




13	Asset Value	Rs. 1350/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found main board, valve and motor defective. Machine installed on 01-09-2018 and covered up to 5+ years. Quotation not attached since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Libin. 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Installed on 1/9/2018. yearwise BER submitted met.  
 Submitted BER after 30 days.  
 Quot not attached.  
 Multiple spare damage reported.  
 Submitted BER in 2 pages.

  
 Signature of JC BM (NHM)

Medical Officer in-charge  
 A. C. CHOORNIKARA

  
 For Dr. Sreehala

  
 PRIMARY HEALTH CENTRE  
 THAIKATTUKARA P.O.  
 ALUVA  
 CHOORNIKARA

Date 07/08/2024. Seal  
 Signature of Superintendent / Medical Officer (i/c)

# Digital BP apparatus with charger

01/09/18 Digital BP apparatus with charger 1 nos received from the store.

01/09/18 Digital BP apparatus with charger issued to OP.

Hely

Total = 1

Condemnation done

Roby

Roby

Roby mathew S/w

15/09/18 Digital BP apparatus 1 nos received from flood related medical camp

15/09/18 Digital BP apparatus 1 nos issued to OP

Hely

Total = 1+1 = 2

Roby

Roby mathew S/w

20/01/2020 Digital BP apparatus with adapters - 2 nos received from the store

20/01/2020 Digital BP apparatus with adapters - 2 nos issued to OP.

Hely

Total = 2+2 = 4

Roby

Roby mathew S/w

4-1 = 3

Total = 3

Robys/w

4/1/21 → Digital apparatus 4 received from store  
 4/1/21 → Digital apparatus 4 issued to OP 4+3 = 7

Date	From whom received	Receipt	Balance	Signature	Cost/Remarks
26/1/21	(from store)	1	3 2 nos working	<u>Roby</u>	



Medical Officer in-charge  
 • H. C. CHOORNIKARA

Manoj  
 For Dr. Sreedhara

# OMRON

HEM-7120



SYS  
mmHg


DIA  
mmHg

PULSE  
/min



START  
STOP

OP



Medical Officer in-charge  
M. C. CHOORNIBARA

*Maaray*  
Dr. Sreechala



40690

DC6V



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
HEALTH CARE PVT LTD




**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-8906152237) 0747126**



CYRIX HEALTH CARE  
BEMP'S  
PREVENTIVE  
PMS DONE  
NEXT PMS DUE  
PERFORMED



Medical Officer in-charge  
M. C. CHOORNIKARA

*Sreedhara*

Dr. Sreedhara

**OMRON**

AUTOMATIC BLOOD PRESSURE MONITOR

MODEL : HEM-7120

RATING : DC 6V 4W

CE 0197

SERIAL No. :

20170622546VG1


OMRON HEALTHCARE Co., Ltd.  
Kyoto, JAPAN

OMRON HEALTHCARE EUROPE B.V.  
Scorpius 33, 2132 LR Hoofddorp  
THE NETHERLANDS

Asia Pacific HQ :  
OMRON HEALTHCARE SINGAPORE PTE LTD.

IN

OP



Medical Officer M.L.S. &  
W.C. CHOORNIKARA

Manoj  
P. P. K. K. K.