


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC CHITTATTUKARA ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Nebuliser Make : Dr Morpen Model : CN 06 SN : 18010931
4	Equipment ID & Barcode	122786 & 0743635
5	Date of purchase / Year of manufacture /Installation Date	24-08-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	12-07-2024 (Toll free)
9	Action taken	Checked and found compressor motor with piston defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs. 1353/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Machine installed on 24-08-2018 and covered up to 5+ years. Quotation not attached since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Sebin.. 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

yeuvuvu BER criteria met. Quot not attached.  
 more than 5 years old. unable to repair the  
 unit reported by Cyrix.  
 Submitted after 30 days.  
 Submitted BER in 2 pages.

  
 Sevinthy  
 Consultant BM  
 31/8/2018  
 National Health Mission (NHM)  
 Aranyakulam  
 Signature of JC BM (NHM)

Date



  
 MEDICAL OFFICER IN CHARGE  
 Superintendent Medical Officer (I/c)  
 Chittattukara, Vadekkkara P. O.





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 230132

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility : 24x7 PHC</p> <p>Address : Chittattukara</p> <p style="padding-left: 40px;">PKM</p> <p>Ph : 9744854647</p>	<p>Call Registration Date : 12/07/2024</p> <p>Caller ID : 122786</p> <p>Date of Visit : 13/07/2024</p> <p>Asset No. : 0743635</p> <p>EQPT Name : Nebuliser</p> <p>Manufacture : Dr. Marepen Model : Cal 06</p> <p>S. No. : 18010931 Dept. : Observation Room</p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** Fumes not getting.

**Action Taken :** checked and found compressor motor with piston defective. Need to replace these spares for further checking and working condition of the equipment.

Completed  Date : 13/7/2024 Time : 2:00pm Spare Required

Spare Replaced  Requested


No.	Description	Qty.	Part Number	PR Number
1.	-	-	-	-
2.				
3.				

Cyrlix Engineer	Date	Start Time	End Time
Sebin Sby	13/7/24	12:00pm	2:00pm

**Customer Remark** Completed  Pending


<p><b>Service Engineer Name :</b> Sebin Sby</p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> 13/7/2024</p> <p><b>Contact Number :</b> 9746751738</p>	<p><b>Customer Name :</b> SHEEJA</p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> 13/7/2024</p> <p><b>Contact Number :</b> 8907902640</p> <p><b>Designation :</b> Nursing officer</p> <p><b>Hospital Seal :</b></p>
---	--

**MEDICAL OFFICER IN CHARGE**  
PRAMKEY HEALTH CENTRE  
Chittattukara, Vadakkakara P. O.





Date	Received	Cost per Umb.	From stock Page No	Unserviceable article	Condemned	Balence	Sign
1-1-14	B/F From old OP Substock Register						
	page 56						
30/3/2011	Rd from store	Rs 2499	M.S 121	1 (10/11/14)		1	
25/12/2011	Rd from store (Aerocare Sevans)	Rs 2912	M.S 121	1 (11/11/14)		2	
24/3/14	Rd from store (Omicon NEC-28)	Rs 2571	M.S 121			3	SK
Condemned as per order no 835/2011 dated 15.6.2011 from the medical officer CHC Eglistore							
1/2/18	Rd from store Rxv. Fw. body part (33) (safe & touch, Antib. Health care products)	Rs 1562.65/-				2	1 SK
29/8/2018	Rd from store					3	SK
17/1/2020	Rd from the store mettch HANBY MCB. (Floor and off)					4	
16.2.2020	Rd from the store Bill No. 1888 Dated 6/6/2019 (Floor)	Rs 2370/-				5	
21/4/2021	Rd R the store Bill no: 344 Dated Kalam kamuniya	Rs 1232/-				6	
24/3/22	Rd R the store ad/11/21 (10/2)					7	
Condemned as per order no 311/20 dated 9/11/20 from the medical officer							
C/1/3/18 Rs 1562.65/-							
Condemned as per order no 391/20 dated 9/11/20 from the medical officer							
(9/11/20 24/7/14)							



*[Signature]*  
**MEDICAL OFFICER IN CHARGE**  
 PRIMARY HEALTH CENTRE  
 Chittattukara, Vadiakkakara P. O.  
 5



NOT WORKING

Morepen<sup>™</sup>  
bulizer

CYRIX

Dec-03 7:17 AM  
1590931  
12-01-2024  
12-02-2025

CYRIX MEDICAL EQUIPMENT MAINTENANCE PROGRAMME GROUP  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225) 0743636

മരണശിക്ഷ



CYRIX HEALTH CARE PVT LTD  
BEMP SERVICE PROVIDER  
PREVENTIVE MAINTENANCE SERVICE  
NEXT PMS DUE 14/01/2025  
PERFORMED BY

Dr. Morepen Compressor Nebulizer  
MODEL NO. CN 06  
POWER 150 W  
INPUT AC-220V 50Hz  
SERIAL NO. 1801031  
LOT NO. 1801  
Made in operation - continuous  
No. of parts 1 PMS

MEDICAL OFFICER IN CHARGE  
PRIMARY HEALTH CENTRE  
Chittallukara, Vadakkakara P. &