



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 225412

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>UPHC</u>	Call Registration Date : <u>09/7/2024</u>
Address <u>UPHC Perumbaitadu</u>	Caller ID : <u>122150</u>
<u>Kottayam</u>	Date of Visit : <u>12/7/2024</u>
Ph : <u>9847904897</u>	Asset No. : <u>0563079</u>
	EQPT Name : <u>Digital BP Apparatus</u>
	Manufacture <u>microlife</u> Model : <u>clica</u>
	S. No. <u>1005669</u> Dept. <u>DP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : BP Apparatus showing error and not working properly

Action Taken : Checked and found main PCB card and motor had the problem. Checked and found spare parts are not available in the market. It is irreparable.

Completed  Date : ..... Time : ..... Spare Required

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sanish Sivan</u>	<u>12/7/24</u>	<u>10:00am</u>	<u>10:30am</u>

Customer Remark  Completed  Pending   
Machine not working. Neelhumolika

Service Engineer Name : <u>Sanish Sivan</u>	Customer Name : <u>Dr Aswathy S. Nair</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>12/7/2024</u>	Date : <u>7/7/24</u>
Contact Number : <u>9946646571</u>	Contact Number : _____
	Designation : _____
	Hospital Seal : _____



**Medical Officer**  
**UPHC, Perumbaitadu**

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	UPHC PERUMBAIKADU
3	Name of Equipment with Make, Model and Serial Number	Equipment :Digital B P Apparatus Make :Circa Model :microlife ✓ SN:Nil
4	Equipment ID & Barcode	122150& 0543079 ✓
5	Date of purchase / Year of manufacture /Installation Date	18/03/2022
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	09-07-2024 (Toll free)
9	Action taken	Checked and found out the problem with main PCB and Nibp pump
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A



13	Asset Value	Rs .2184/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and NIBP pump. Machine installed in 18/03/2022 and covered up to ? 2+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sanish sivan

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 2184  
 installation Date - 18/3/22  
 Cost of spare - N/A available  
 Life period - 2.4 years  
 As per the RASR guidelines Life period criteria not satisfied

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL ENGINEER  
 URBAN PRIMARY HEALTH CENTRE  
 PERUMBAIKAD  
 28/4/24



Date 19/7/24

Medical Officer  
 UPHC, Perumbaikad  
 Seal

Signature of  
 Superintendent / Medical Officer (i/c)

# STOCK REGISTER

Name

B.P. APPARATUS


27

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
18-4-15	-	-						
20-3-20	22 Issued to				2	2	NIL	
19-12-24	Received. (Digital)	BA-8027 21-22	2184	PAY ONE	2	2	NIL ✓	
					1	1	NIL	



*[Signature]*  
Medical Officer  
UPHC, Perumbaikad





*[Signature]*  
Medical Officer  
UPHC, Perumbakad

