



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 226441

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC</u></p> <p>Address : <u>Udumpanchola</u> <u>Idukki, Kerala</u></p> <p>Ph : <u>9495818470</u></p>	<p>Call Registration Date : <u>8/7/24</u></p> <p>Caller ID : <u>121883</u></p> <p>Date of Visit : <u>10/7/24</u></p> <p>Asset No. : <u>0640067</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture : <u>Life line</u> Model : <u>Mill</u></p> <p>S. No. : <u>Mill</u> Dept. : <u>Sub Center</u></p>
--	---

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : Not Working

Action Taken : Checked the Machine found  
that the Mals board and Pump defective,  
Need to replace the Spare Part Further checking  
working condition of the equipment

Completed  Date : 10/7/24 Time : 2pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>MD</u>			
2.			
3.			

Cyrix Engineer	Medical Officer	Date	Start Time	End Time
<u>Vineesh Vincent</u>	<u>PHC Udumpanchola</u>	<u>10/7/24</u>	<u>12pm</u>	<u>2pm</u>

Customer Remark  Completed  Pending

Service Engineer Name : Vineesh Vincent Name : Mini S  
 Signature : Vineesh Signature : [Signature]  
 Date : 10/7/24 Date : 10/7/24  
 Contact Number : 7593847124 Contact Number : 9495426343  
 Designation : JPHN Gr I  
 Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	IDUKKI
2	Name of Hospital	PHC Udumpanchola
3	Name of Equipment with Make, Model and Serial Number	Equipment name: B.P Apparatus Make: lifeline Model: na Serial number: na
4	Equipment ID & Barcode	ID: 121883 Barcode: 0640067
5	Date of purchase / Year of manufacture /Installation Date	24-9-2019
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	08-7-2024 (Toll-free)
9	Action taken	Regarding the inspection of the equipment found that the main board and pump are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Na



13	Asset Value	2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	The Cyrix service report is attached
16	Reasons for recommending the equipment as BER	Checked and found the main board and pump defective. Equipment was installed on 24/9/2019 and covered up to 4 years. Quotation not submitted since spares are not available in the market. So, recommend for RBER
17	Name & Signature of CYRIX Authority	Jithu Abraham

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Consultant Biomedical Engineer  
National Health Mission  
Arogyakeralam  
Idukki-685603

*Rajitha*  
29/07/2024

Signature of JC BM (NHM)

Date 05/08/24



*[Signature]*  
Signature of  
Superintendent / Medical Officer (i/c)

Medical Officer





STOCK REGISTER

ITEM Sub center kit

Date/No.	Received/issued to whom	Opening Balance	Receipt	Total	Issue	Balance	Initials of receiver	Remarks
24.9.19	Received from N/10							
1	BP apparatus Digital life care RS-1119-15	-	1	1		1	B	
2	(RS 194) Dust Bin black	-	1	1		1	B	
3	(RS) Rack (194)		1	1		1	B	
4	RS 194 Yellow		1	1		1	B	
5	Examination Torch RS 78		1	1		1	B	
6	Fotoscope white		1	1		1	B	
7	Glucometer on call plus 200/200		1	1		1	B	
8	H. Monometer True		1	1		1	B	
9	Hub Culler Many		1	1		1	B	
10	Infantometer Krupp		1	1		1	B	
11	Measuring Tape		1	1		1	B	
12	Nebulizer Life care		1	1		1	B	
13	Stadiometer Life care		1	1		1	B	
14	Ophthalmoscope Super deluxe Topex		1	1		1	B	
15	Thermometer Hospital		1	1		1	B	
16	Weighing machine Digital Pediatric		1	1		1	B	
17	Weighing machine Digital Adult		1	1		1	B	
18	Sterile gloves		25	25		25	B	
					5 for (11)	Nil	B	



*[Signature]*  
Medical Officer  
PHC Hebbal, Bangalore

1  
1  
13  
26  
25  
00  
6