


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	CHC ULLIYERI
3	Name of Equipment with Make, Model and Serial Number	EQUIPMENT: AUTOCLAVE VERTICAL MAKE: RELIANCE INSTRUMENTS CORPORATION MODEL: VT12 SN: NA
4	Equipment ID & Barcode	117130 & 1130999
5	Date of purchase / Year of manufacture /Installation Date	22/04/2011
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-06-2024
9	Action taken	Checked and found that body completely rusted, coil, wirings and control board defective. Enquired service support from OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	49491/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and OEM EOL attached
16	Reasons for recommending the equipment as BER	Checked and found that body completely rusted, coil, wirings and control board defective, machine installed on 22/04/2011 Equipment out lived 13 years 3 month. Oem informed that equipment is in a non repairable condition and issued eol letter , As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHILRAJ MP 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that body rusted, wirings & control board defective. Machine is installed on 22/4/2011. OEM is informed machine is irreparable condition. Report is attached. So recommended for condemnation.


Signature of JC BM (NHM)

Consultant Bio Medical Engineer
National Health Mission
Kozhikode

Date



Seal



Signature of Medical Officer
Community Health Centre
Superintendent / Medical Officer (I/c)
Uthiyer, Kozhikode, Kerala - 673323



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 148466

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHE Ulliyeri</u></p> <p>Address <u>Kozhikode</u></p> <p>Ph :</p>	<p>Call Registration Date : <u>14/06/2024</u></p> <p>Caller ID : <u>117130</u></p> <p>Date of Visit : <u>15/06/2024</u></p> <p>Asset No. : <u>1130999</u></p> <p>EQPT Name : <u>Autoclave vertical</u></p> <p>Manufacture <u>Reliance Instruments Corporation</u> Model : <u>VT.12</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT switching on

Action Taken : checked and found that- body completely rusted, coil, wirings and control board defective. Enquired service support from OEM

Completed Date : 15/06/2024 Time : 1pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jineesh PP</u>	<u>15/6/24</u>	<u>11 AM</u>	<u>1 PM</u>

Customer Remark Completed Pending



Service Engineer Name : Jineesh PP
Signature : [Signature] Sr-BME
Date : 15/6/2024
Contact Number : 8943527662

Customer Name : Bijina . m . P .
Signature : [Signature]
Date : 15/06/24
Contact Number : 9745738140
Designation : Nursing officer gr2
Hospital Seal : [Seal]



RELIANCE INSTRUMENTS CORPORATION

Infection Control Specialist Beside You..

RIC/GEN/KMSCL/086/2024-25

22.07.2024

To
CHC Ulliyeri
Kozhikode (Dist)
Kerala

Dear Sir,



Sub: End of Life Letter for Vertical Autoclave – Reg

Item Name : Autoclave Vertical
Make : Reliance Instruments Corporation
Model: VT12
Equipment Code : EQPT -09- 06/2010
Installation Date : 22/04/2011

With reference to the above Equipment, We would like to inform that the equipment has served its best till date. On Examination the equipment is in a non-repairable condition, hence we would recommend condemning the Autoclave.

Thanking you

Yours Faithfully
For Reliance Instruments Corporation


Authorized Signatory




Required service support-Autoclave Vertical,CHC Ulliyeri, Kozhikode, Kerala - 117130-1130999

3 messages

Jineesh P P <jineeshcyrix@gmail.com>

Fri, 2 Aug, 2024 at 12:35 am

To: reico@relianceinstruments.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir

There is a compliant reported from CHC Ulliyeri, Kozhikode, Kerala , regarding the compliant with Vertical autoclave. Checked and found machine body completely rusted ,coil ,wiring and control board defective.

We required your one time service support for rectification of compliant

Please note the following details and attached photos &stock details for your reference

Autoclave Vertical
Barcode no : 1130999
Machine installed on 22/04/2011
Make : Reliance Instruments Corporation
Model: VT12
Equipment Code : EQPT -09- 06/2010

Kindly do the needful at the earliest

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
Mob-7593847163
jineeshcyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Jineesh P P <jineeshcyrix@gmail.com>

Tue, 6 Aug, 2024 at 1:08 pm

To: reico@relianceinstruments.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

Reminder

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
KLBEMMP
[Quoted text hidden]
[Quoted text hidden]

Reliance Instruments corporation <reico@relianceinstruments.com>

Tue, 6 Aug, 2024 at 2:08 pm

To: Jineesh P P <jineeshcyrix@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir,

PFA - End of life Letter for Vertical Autoclave

Regards
Team Reico



Reliance Instruments Corporation

No.217, SIDCO Industrial Estate, North phase,
9th Street, Ambattur, Chennai 600098

Phone : 044 48572311 , Mobile : 9962567028/34

Website: <http://www.relianceinstruments.com> & <http://www.relianceinstruments.in>

Email : reico@relianceinstruments.com

Support : support@relianceinstruments.com

[Quoted text hidden]



Medical Officer
Community Health Centre
Ulliyeri, Kozhikode, Kerala - 686012



Government Health Centre
Uliyari, Kozhikode, Kerala- 673323



Medical Officer
Community Health Centre
Uliyari, Kozhikode, Kerala- 673323

