


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC MARUTHARODE
3	Name of Equipment with Make, Model and Serial Number	NEBULISER Make : INFINEB Model: CN01WB SN: 043692
4	Equipment ID/ Barcode	0940985 # 125416
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date = Nil (As per stock book equipment installed on 17/12/2009 , stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	25-07-2024 (toll free)
9	Action taken	Checked the machine found not switching on found nebuliser motor burned , piston broken. Need to replace nebuliser motor with piston for further checking of machine
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS. 1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine found not switching on found nebuliser motor burned , piston broken. As per stock book Equipment installed on 17/12/2009 and aged upto 14years 7 months. Quotation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**


* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Nebulizer installed on 17/12/2009 and aged up to 14 years. Quotation not submitted.


06/08/24
Signature of J. S. B. (Biomedical Engg) Consultant (NHM), Palakkad
NHM (Arogyakeralam), Palakkad

Date: _____


Signature of Superintendent/Medical Officer (i/c)
MEDICAL OFFICER
FAMILY HEALTH CENTRE
MARUTHAROAD, KOTTEKAD, P.O.
PALAKKAD-678 732





**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 231794

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>PHC Marutharode</i>	Call Registration Date : <i>25-07-2024</i>
Address <i>Palakkad</i>	Caller ID : <i>125416</i>
<i>Kerala</i>	Date of Visit : <i>26-07-2024</i>
Ph : <i>09778165374</i>	Asset No. : <i>0940985</i>
	EQPT Name : <i>nebuliser</i>
	Manufacture <i>Infi neb</i> Model : <i>EN-01WB</i>
	S. No. <i>043692</i> Dept. <i>OP</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *unit not working*

Action Taken : *checked the machine Found not switching on Found nebuliser motor burned, piston broken. need to replace nebuliser motor with piston for further checking of machine*

Completed Date : *26-07-24* Time : *10:45AM* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <i>will</i>	<i>will</i>	<i>will</i>	<i>will</i>
2. <i>will</i>	<i>will</i>	<i>will</i>	<i>will</i>
3. <i>will</i>	<i>will</i>	<i>will</i>	<i>will</i>

Cyrix Engineer	Date	Start Time	End Time
<i>ABIJITH -B</i>	<i>26-7-24</i>	<i>10:05AM</i>	<i>10:45AM</i>

Customer Remark Completed Pending

Service Engineer Name : <i>ABIJITH -B</i>	Customer Name : <i>PHC Marutharode</i>
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : <i>26-07-2024</i>	Date : <i>26/7/2024</i>
Contact Number : <i>9567164651</i>	Contact Number : <i>09778165374</i>
	Designation : <i>Nursing Officer</i>
	Hospital Seal : <i>[Seal]</i>


MARUTHAROAD, KOTTEKKAD.P.O.
PALAKKAD-678752

Office of NUBOLISER

Receipts

Date	No. of receipt where necessary	Particulars	Pay		Allowances		Contingencies				Miscellaneous	Total		Classification
			Rs.	P.	Rs.	P.	Rs.	P.	Rs.	P.		Rs.	P.	
1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
17/12/09		Recd from stock								1			1	
17/12/09 1/1/2016		S-105 Brought forward to instruments volume 2			Gaken for use to Sub stock Page no 64					1				

[Handwritten signature]



[Handwritten signature]
MEDICAL OFFICER
FAMILY HEALTH CENTRE
MARUTHAROAD, KOTTEKKAD.P.O.
PALAKKAD-678 732

CYRIX
CYRIX MEDICAL EQUIPMENT
TOLL FREE NO: 1800 - 425-7869
SALE CODE: (800A-800915223)
09409885

Compressor Nebulizer
Model: CH-01968
Input: 220V/50Hz
Current: 0.5A
Serial No: 811517
Date of manufacturing: 2007/05/2009
Certified Gbr, Germany
Imported & Marketed by
Infinity Mediquip India
C-11/40, Sec. 3, Rajawada, 110086
e-mail: support@inf-med.com

KTPL
Kerala Tobacco & Cigarette Producers' Limited
KOTTEKKAD
PALAKKAD
Kerala
TOLL FREE NO: 1800-787878
0778846088



[Handwritten signature]

MEDICAL OFFICER
FAMILY HEALTH CENTRE
MARUTHAROAD, KOTTEKKAD, P.O.
PALAKKAD-678 732