



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 231904

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>CHC</u>	Call Registration Date : <u>03/07/2024</u>
Address <u>Manamboor</u>	Caller ID : <u>120903</u>
<u>Thiruvananthapuram</u>	Date of Visit : <u>04/07/2024</u>
Ph : <u>8606264130</u>	Asset No. : <u>0150661</u>
	EQPT Name : <u>Oxygen Concentrator</u>
	Manufacture <u>OLIVE</u> Model : <u>OLV-10</u>
	S. No. <u>2021061043</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not working

Action Taken : checked the unit and found main board, sieve bed and solenoid valve complaint. need to replace these spares for further checking and working condition of the equipment Enquired spares from OEM

Completed Date : 04/07/24 Time : 2:00 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix-Engineer	Date	Start Time	End Time
<u>Vishnu.K</u>	<u>04/07/24</u>	<u>12:45 pm</u>	<u>2:00 pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Vishnu.k</u>	Customer Name : <u>AN S. AV</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>04/07/24</u>	Date : <u>4/7/24</u>
Contact Number : <u>9605884413</u>	Contact Number : <u>8606264130</u>
	Designation : <u>Nursing Officer</u>
	Hospital Seal :




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	CHC MANAMBOOR
3	Name of Equipment with Make, Model and Serial Number	Equipment Name :Oxygen concentrator Make : Olive Model : OLV-10 Serial No : NA
4	Equipment ID/ Barcode	120903/0150661
5	Date of purchase/ Year of manufacture/Installation Date	12/06/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03/07/2024(Toll free)
9	Action taken	Checked and found main board, Sieve Bed and solenoid valve complaint. Need to replace these spares for further checking and working condition of the equipment. Enquired spare from OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found mainboard, sieve bed and Solenoid valve complaint. Enquired spares from OEM. No response from the OEM. The equipment was supplied to the institution through donation at the covid time. Since the manufacturer not exist as per the tender clause 5.3.14.2 recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	VISHNU K 

*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Main board, sieve bed and solenoid valve complaint.


Equipment received through covid. Since manufacturer not exist, this can be BER as per clause.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

 18-08-2024
Signature of JC BM (NHM)

Date: 29/7/2024





Signature of
Superintendent Medical Officer (IC)
Community Health Centre
Manamboor, Trivandrum,

(40)

OXYGEN CONCENTRATOR

Date	KL05256 Reference	Manufacturer / Item	Qty	Used	Balance	Remarks
12/6/21		Drasicon	1		1	
12/6/21	KL001/FP 00630/21-23	Old DDU (TV 0)	1		2	
		Batch No. WTD2101830				
19/6/21		Mkt. TVCUNA				H/11 Inst. No. 19-V011
12/7/21		Issued to Head Nurse		2	Nil	
2/8/21	8/5/20 2/5/21	Old Perical certificate of ophthalmology, Govt Ophthalmic Hospital Triv pore (C-Blomid) / 2728/2021/8/6/21	1		1	Page 112
10/8/21		Issued to Head Nurse		1	Nil	
19/9/21		Old Kaseya medicine Dept KSCOC Building 11/61 plate ophthalmic CIR 39675 / Batch no 111 Exp. 12/2026 Total cost 39675.00 NET cost 42986.00 KUNIG Kaseya Kmic supply records prepared. (Callid. 19)	1		1	Page 113
13/9/21		Issued Head Nurse		1	Nil	



Medical Officer
Health Centre
Mannambur, Trivandrum



REGARDING OXYGEN CONCENTRATOR-QUOTATION REQUIREMENT-CHC MANAMBOOR-OLIVE-V-10 (120903/0150661)

Messages

syep P V <kasyepcyrix@gmail.com>
: info@oliveoxygen.com
: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Sat, 13 Jul 2024 at 9:34 a

Dear Sir,
I'm reaching out to enquire about the quotation for the spare parts for the oxygen concentrator.
Make : Olive
Model : OLV 10

The quotation required for;
1. Seive Bed- 2 nos
2. Solinoid valve
3. MainBoard

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

syep P V <kasyepcyrix@gmail.com>
: info@oliveoxygen.com
: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Mon, 15 Jul 2024 at 10:46 a

Sir,
Consider this mail as gentle reminder
[Quoted text hidden]
[Quoted text hidden]

syep P V <kasyepcyrix@gmail.com>
: info@oliveoxygen.com
: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Thu, 18 Jul 2024 at 11:11 a

Dear sir
Reminder
[Quoted text hidden]

syep P V <kasyepcyrix@gmail.com>
: info@oliveoxygen.com
: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Sat, 20 Jul 2024 at 9:22 a

Dear sir,
Consider this mail as a gentle reminder

[Quoted text hidden]



Medical Officer In -Charge
Community Health Centre
Manamboor, Trivandrum.

