


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	CHC CHALISSERY
3	Name of Equipment with Make, Model and Serial Number	DIALYSIS MACHINE Make : B BRAUN Model: DIALOG+ SN: 107083
4	Equipment ID/ Barcode	0931009 # 95546
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil (as per stock book installation date 30-05-2013, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	12-02-2024 toll free
9	Action taken	Checked the machine Found not switching on. Problem with strips, pressure sensor, motor for gear pump, gear pump ep, spe, lpa, membrane for balance chamber, heater 240v, level sensor complete, pump body, motor for piston pump, blood leak detector, power board, touch screen, 1x800 mother board, compact flash card 1gb, analog board, digital board version 6, pressure monitoring module, retainer tube green, heparin pump defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Switched mode power supply mc2 = 189332/- Pressure sensor = 28332/- Motor for gear pump 3 nos = 118089 - Gear pump ep = 64988 - Gear pump spe = 50413/- Gear pump lpa = 48007/- Membrane for balance chamber = 9323/- Heater 240v /2x900v = 36573/- Level sensor complete = 9432/- Pump body 3 nos = 120573/- Motor for piston pump 3 nos = 52521/- Blood leak detector v4 = 20126/- Power board valves = 52864/- Touch screen for led = 52339/- 1x800 mother board = 115895/- Compact flash card 1gb = 12986/- Analog board = 50230/- Digital board version 6 = 82308/- Pressure monitoring module = 50463/- Retainer tube green = 1405/- Heparin pump = 90123/- Total = 1266322/- GST 18% = 227938/- Total amount = 1494260/-

13	Asset Value	RS. 506667/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	294%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached Quotation attached
16	Reasons for recommending the equipments as BER	Checked the machine Found not switching on. Problem with snips, pressure sensor, motor for gear pump, gear pump ep. spe. spa, membrane for balance chamber, heater 240v, level sensor complete, pump body, motor for piston pump, blood leak detector, power board, touch screen, 1x800 mother board, compact flash card 1gb, analog board, digital board version 6, pressure monitoring module, retainer tube green, heparin pump defective. Enquired OEM for quotation. equipment installation date as per stock book 30-05-2013 aged up to 10 years 9 months. Repairing cost is 294% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine installed on 30-5-2013 and machine aged up to 30+5 10 years. Quotation submitted by OEM is about Rs. 1994260 and the spare cost exceeds actual cost of machine. Here recommended for BER.

Signature of JC BM (NHM) 

26/07/12

Consultant (Biomedical Engg) NHM (Arogyakaram), Palai



Signature of Superintendent/Medical Officer (I/c)

Date:



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-JT/2021-2022/598

CYRIX

HEALTHCARE PVT LTD

No.:

217561

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kechi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bomp.hk@cyrilx.in

<p>Health Facility <u>CHE Chalissery</u></p> <p>Address <u>Palawad.</u></p> <p>Ph : <u>9291713456</u></p>	<p>Call Registration Date : <u>12/02/24</u></p> <p>Caller ID : <u>95546</u></p> <p>Date of Visit : <u>13/02/24</u></p> <p>Asset No. : <u>0921009</u></p> <p>EQPT Name : <u>DIALYSIS MACHINE</u></p> <p>Manufacture <u>B. BRAUN</u> Model : <u>DIALOCAT</u></p> <p>S. No. <u>107083</u> Dept. <u>DIALYSIS</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not working.

Action Taken : Checked the Machine found not working on Problem with SMPS. Pressure Sensor, Motor for gas pump, fpc, fpa, membrane for balance chamber, heater 240V, level sensor, Comptie, Pump body, Motor for Astor Pump, blood leak detector, Power board, touch Screen, LV 500 Mother board, Compact flash card 1gb, analog board, digital board Version C, Pressure Monitor Module, tubing tube green, heparin pump defective, Enquired OEM for quotation

Completed Date : 13/02/24 Time : 4:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Tajav</u>	<u>13/02/24</u>	<u>11:00 AM</u>	<u>4:00 PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Tajav</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>13/02/24</u></p> <p>Contact Number : <u>8948098422</u></p>	<p>Customer Name : <u>Dr. Saji</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>13/02/24</u></p> <p>Contact Number : <u>9447309994</u></p> <p>Designation : <u>Medical officer</u></p> <p>Hospital Seal :</p>
---	---



SUPER INDEPENDENT
COMMUNITY HEALTH CENTRE
CHALISSERY, PALAWAD, PIN



Service report

Service ETS Request No **49969**
 Date of Visit : **11-7-2024**

Installation

PMS / TSC

Break Down

Customer Name : community health center		Service Request Category:	
Customer Address : chaliserry		Installation	
Contact Number : _____		User Training checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N	
Division: BA <input checked="" type="checkbox"/> HC <input type="checkbox"/> AE <input type="checkbox"/>		Installation checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N	
Service Place: Onsite <input checked="" type="checkbox"/> Workshop <input type="checkbox"/> Office <input type="checkbox"/>		Warranty End Date: (YYYY / MM / DD)	
Model : Dialog + EVO		TSC cum PMS	
Serial Number: 107083		If Yes PM KIT replacement date: (YYYY / MM / DD)	
Software version: 9.12 Working hours: _____		User Training: <input type="checkbox"/> Y <input type="checkbox"/> N	
Observations:		User Training checklist signed: <input type="checkbox"/> Y <input type="checkbox"/> N	
RO TDS: ___ (0-10PPM Acceptable limits) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Breakdown call	
Neutral to Earth Voltage: ___ (0-2 Volt)		Software: <input type="checkbox"/> Y <input type="checkbox"/> N	
Use of UPS / Stabilizer: ___ (recommended 3KVA per machine in Dialog model) <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Electrical Hardware: <input type="checkbox"/> Y <input type="checkbox"/> N	
Bicarb mixer: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA		Mechanical Hardware: <input type="checkbox"/> Y <input type="checkbox"/> N	
Reported Problem: machine not switching on		Trend backup: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	
Observed Problem: _____		50% Citric Disinfection after each Therapy: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	
Service Resolution: check the machine need to replace mention spare		50% Citric Disinfection after end of day: <input type="checkbox"/> Y <input type="checkbox"/> N <input type="checkbox"/> NA	
Following spareparts are: touch screen		Service Type	
Replaced: <input type="checkbox"/> Y <input type="checkbox"/> N		Warranty / Extended warranty: <input type="checkbox"/>	
To be replaced: <input type="checkbox"/> Y <input type="checkbox"/> N		CAMC: <input type="checkbox"/>	
To be quoted: <input type="checkbox"/> Y <input type="checkbox"/> N		AMC: <input type="checkbox"/>	
		Oncall & Chargeable: <input checked="" type="checkbox"/>	

Reported Problem: **machine not switching on**

Observed Problem: _____

Service Resolution: **check the machine need to replace mention spare**
SMPS-1, Heater-1, level sensor-1, Pressure sensor-2
Gear Pump EP-1, Gear Pump FPE-1, Gear Pump FPA1, motor for
Gear Pump-3 & Pump body-3, motor for piston pump-3
retainer tube Green-1, Blood leak sensor-1, membran-1
Digital board-1, Analog board-1, Power board valve-1, mother
board for 800-1, CP card-1, PA/PV modul-1, Heparin Pump

SN	Article number	Article Name	Quantity
		SMPS, mother board, Analog, digital	
		Gear Pump - EP, FPA, FPE, motor,	
		Pump body, Pressure sensor, PA/PV	
		modul, Blood leak sensor, membran	

Date (YYYY / MM / DD) _____ Start time (Hrs : Min) _____ End time (Hrs : Min) _____ Labour Hours Spent: (Hrs : Min) _____

This is to certify that above mentioned spareparts of B. Braun machine are consumed / to be consumed on our in hospital for satisfactory working of machine.

Status of Service request:	Completed: <input type="checkbox"/> Y <input type="checkbox"/> N	Followup Date: (YYYY / MM / DD)
Contact person: PREETHA K V	Signature with Date & Time (YYYY / MM / DD): Preetha	



Customer remark / Feedback: _____

Engineer name & Signature: **[Signature]** Engineer Contact No.: _____

To,
 Cyrix Healthcare
 C/O Chalisery

Sub: - Quotation for Spare Parts

Our Ref.BB/BA-W-221/24 Date: 17.07.2024

Dear Sir/ Madam,

Greetings from BBraun Medical India Pvt. Ltd.

Kindly find our most competitive rate for Spare part for Hemodialysis machine.

Art. No.	Description	Qty	Unit price (Excl. Tax)	Total Price (Excl. Tax)
3461076	Switch mode power supplyMc2	1	189,332	189,332
3456124A	Pressure Sensor	2	14,166	28,332
34560718	Motor for gear pump	3	39,363	118,089
3461097	Gear pump Ep	1	64,988	64,988
34562331	Gear pump Fpe	1	50,413	50,413
34562330	Gear pump Fpa	1	48,007	48,007
3451615A	Membrane for balance chamber	1	9,323	9,323
3456020B	HEATER 240V / 2X900W	1	36,573	36,573
34560459	LEVEL SENSOR COMPLETE	1	9,432	9,432
34560602	PUMP BODY	3	40,191	120,573
3456055B	MOTOR FOR PISTON PUMP	3	17,507	52,521
34561692	BLOOD LEAK DETECTOR V4	1	20,126	20,126
34562150	POWER BOARD (VALVES)	1	52,864	52,864
34514260	TOUCH SCREEN FOR LCD+	1	52,339	52,339
34560610	LX800 motherboard	1	115,895	115,895
3456088A	Compact flash card 1 GB	1	12,986	12,986
34560734	Analog board	1	60,230	60,230
34562303	Digital board, version 6	1	82,308	82,308
34561641	Pressure monitoring module	1	50,463	50,463
3456077A	Retainer tubes (green)	1	1,405	1,405
3451432D	HEPARIN PUMP COMPACT	1	90,123	90,123
			Total:	1,266,322
			GST: 18%	227,938
			Total Amount:	1,494,260



**Purchase Order -KL 24 25 101011 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0311421 ,0931009
DH KOZHENCHERRY ,CHC CHALISSERY**

11 messages

Anjitha T L <purchasecyrix2@gmail.com> Thu, 4 Jul, 2024 at 11:07 am
To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, dm2.klbemp@cyrix.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear Sir,

Please find the attached Purchase Order and do the needful. Kindly send a confirmation mail with details of material delivery (Docket details) along with the invoice Copy mentioning our PO No. (For reference) with Seal and Signature.

EDD as per payment term:08/07/24

For Queries and Clarifications

1. Technical: dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in
2. Accounts: paymentcyrix@gmail.com
3. Logistics: logistics@cyrix.in
4. PO Clarification: ampurchase@cyrix.in/purchasemgr@cyrix.in
5. Delivery address: **Cyrix Healthcare PVT LTD**

Methanam Road
Door No: 312/a Eloor North,
Kerala – 683501- India
Nikesh V
Mob: 9072522014

@ Cyrix Accounts

Payment Term: 100% adv

Thanks & Regards

Anjitha T L
Sr. Purchase Executive
PH:- 9947719111
purchasercyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Cyrix Payments <paymentcyrix@gmail.com>

Thu, 4 Jul, 2024 at 5:37 pm

To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>

Cc: Anjitha T L <purchasercyrix2@gmail.com>, Babu Murugesan <babu.murugesan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, dm2.klbemp@cyrix.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasercyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear Sir/Mam,

Please note that we had transferred Rs.27840 /- as advance against PO101011 after deducting TDS Rs.480/-

B. Braun Medical (India) Pvt. Ltd - Mumbai	0062620000	27,840.00	AXISP00515309209	PAID	04-JUL-24
---	------------	-----------	------------------	------	-----------

[Quoted text hidden]

--

Thanks & Regards,

Tincy T Jose
Accounts Department.
Cyrix Healthcare Pvt Ltd

Office : 30/641 B, First Floor, Petta Junction, Poonithura, Cochin, Kerala - 682 038

Website:- www.cyrixhealthcare.com

Mob : 7593847123

Babu Murugesan <babu.murugesan@bbraun.com>

Fri, 5 Jul, 2024 at 11:27 am

To: Anjitha T L <purchasercyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, dm2.klbemp@cyrix.in <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasercyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear Anjitha,

Kindly provide machine serial numbers.

Thanks & Regards

Babu Murugesan

B. Braun India is recognized as one of the Best Workplaces to Work in India



From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Thursday, July 4, 2024 11:07 AM

To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>; dm2.klbemp@cyrix.in; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>; Maneesh R Nair <maneeshr@cyrix@gmail.com>

Subject: Purchase Order -KL 24 25 101011 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0311421 ,0931009 DH KOZHENCHERRY ,CHC CHALISSERY

CAUTION: This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

Babu Murugesan

Manager - Technical Services

+919884450417

babu.murugesan@bbraun.com

B. Braun Medical (India) Pvt. Ltd. (CIN : U33112MH1984PTC214514)

WorkEZ - Block-B, Door No.3, 2nd Floor, Hansa Building , RK Swamy Centre, 147, Pathari Rd, Thousand Lights, Chennai, 600006

Tamil Nadu, IN, India

Chennai, 600006

,India

<http://www.bbraun.co.in>

www.linkedin.com/company/bbraun-group

As a true partner
we develop smart solutions
and set standards
to improve health care.

B. Braun – Sharing Expertise

The information contained in this communication is confidential, may be attorney-client privileged, may constitute inside information, and is intended only for the use of the addressee. It is the property of the company of the sender of this e-mail. Unauthorized use, disclosure, or copying of this communication or any part thereof is strictly prohibited and may be unlawful. If you have received this communication in error, please notify us immediately by return e-mail and destroy this communication and all copies thereof, including all attachments.

Anjitha T L <purchasecyrix2@gmail.com>

To: Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP

<om.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>,

Fri, 5 Jul, 2024 at 11:36 am

Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, dm2.klbemp@cyrix.in <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear sir,

Please find the serial No.

DH KOZHENCHERRY - Serial No: 106333

CHC CHALISSERY - Serial No: 107083

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 10 Jul, 2024 at 11:10 a

To: Anjitha T L <purchasecyrix2@gmail.com>

Cc: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, sherin <purchasecyrix5@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038

www.cyrixhealthcare.com

[Quoted text hidden]

DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Wed, 17 Jul, 2024 at 1:22 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan

<babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear Izrath,

Kindly update the status.

Thanks & Regards,

Abijith.A

Divisional Manager | South Zone

dm2.klbemp@cyrix.in

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: 04 July 2024 11:06 AM

To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>; DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>; Maneesh R Nair <maneeshrcyrix@gmail.com>

Subject: Purchase Order -KL 24 25 101011 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0311421 ,0931009 DH KOZHENCHERRY ,CHC CHALISSERY

Dear Sir,

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Sat, 20 Jul, 2024 at 8:17 am

To: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 24 Jul, 2024 at 1:21 pm

To: Babu Murugesan <babu.murugesan@bbraun.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura

[Quoted text hidden]

Wed, 24 Jul, 2024 at 5:08 pm

Babu Murugesan <babu.murugesan@bbraun.com>
To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>, dm2.klbemp@cyrix.in <dm2.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>

Dear Anjitha,

Please find Service report, Spares estimate & invoice copy in the attached files.

Thanks & Regards
Babu Murugesan

B. Braun India is recognized as one of the Best Workplaces to Work in India



From: Anjitha T L <purchasecyrix2@gmail.com>
Sent: Thursday, July 4, 2024 11:07 AM
To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>; dm2.klbemp@cyrix.in; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>; Maneesh R Nair <maneeshrcyrix@gmail.com>
Subject: Purchase Order -KL 24 25 101011 -B. Braun Medical (India) Pvt. Ltd - Mumbai -0311421 ,0931009 DH KOZHENCHERRY ,CHC CHALISSERY

CAUTION: This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Dear Sir,

[Quoted text hidden]

[Quoted text hidden]

Wed, 24 Jul, 2024 at 5:34 pm

Izrath K K <purchasecyrix5@gmail.com>
To: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>
Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Hi,

po:101011

@Akhilesh P @Maneesh R Nair kindly give confirmation

Thanks & Regards
Izrath KK
Jr Purchase Executive
PH: 8848002238

On Thu, Jul 4, 2024 at 11:07 AM Anjitha T L <purchsecyrix2@gmail.com> wrote:

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 24 Jul, 2024 at 10:03 pm

To: Izrath K K <purchsecyrix5@gmail.com>

Cc: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Maneesh R Nair <maneeshrcyrix@gmail.com>, Anjitha T L <purchsecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Dear mam
PO 101011 one time visit done.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

HAEMODIALYSIS MACHINE

117

Date	From whom Received	Receipt	Issue	Balance	Remarks
27/6/14	<p>Received from pharmacy NV NO: 9310001027 dt 30/5/2013 Purchased from B. Bran medical (India) Pvt. Ltd Gandhi Nagar Hyderabad.</p>				
mfg	<p>B- Braun Medical Surgeries Germany Model - Dialog + EVO Warranty 3 year. Serial no. 107083 Serial no. 107085 Rs - 11,27,700.00 /- INV. NO. 93,1000,1055 dt 11/6/2013 purchased from B. Braun - I Pvt Ltd mfg by B. Braun Batch no. 107086 " 107087 Rs - 11,27,700.00</p>	2 NO.		2 NO.	
		2 NO.		2 NO.	
			Total HNo.		
27/6/14	Issued for use.		2nd	NV 9	4
19/10/20	Received from store Issued for use.	1	1		5
	<p>Dialysis machine Transactin type (R) (Nipro medical India) Invt in of Recipient - A job done, through Grand total 5,65,000/- Batch no 20C18823 Invoice no - 5101637, Surgical SS plus (CF - Flushing unit 30 plug)</p>				



SUPERINTENDENT
 COMMUNITY HEALTH CENTRE
 CHALISSERY, PALAKKAD DISTRICT
 18/12

