


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC MALA
3	Name of Equipment with Make, Model and Serial Number	DIALYSIS UNIT Make: Nipro Model: Diamax Sr,No- 16J11726E
4	Equipment ID & Barcode	#120075 – 0830274
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date -22-06-2017
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29-06-2024
9	Action taken	Checked the machine and found 2way valve, DP sensor type 2 assy, Blood leak sensor, cond.sensor pc board and magnetic gear pump assy are complaint. Need to replace these spares for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	2 Way Valve - 17,542.00/- Dp Sensor Type 2 Assy - 52,472.00/- Blood Leak Sensor - 1,11,022.80/- Cond. Sensor Pc Board - 38,640.00/- Magnetic Gear Pump Assy -1,29,360.00/- TOTAL - 3,49,036.80/-

13	Asset Value	435000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	80.21%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report, and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found 2way valve, DP sensor type 2 assy, Blood leak sensor, cond. sensor pc board, and magnetic gear pump assy are complaint. As per stock book Equipment installed on 22/06/2017 and aged up to 7 years, Spare cost is 80.21% of asset value and both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spall cost is high


 Jr. Consultant (Biomedical)
 IN.M. Prissur

Signature of JC BM (NHM)



Superintendent / Medical Officer (i/c)



SUPERINTENDENT
K. KARUNAKARAN SMARAKA
COMMUNITY HEALTH CENTRE
MALA

Signature of Super Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 227970

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>C.H.C Mala</u></p> <p>Address <u>Tharissar</u></p> <p>Ph : <u>9606644825</u></p>	<p>Call Registration Date : <u>29/6/24</u></p> <p>Caller ID : <u>120075</u></p> <p>Date of Visit : <u>30/6/24</u></p> <p>Asset No. : <u>0830274</u></p> <p>EQPT Name : <u>Dialysis Machine</u></p> <p>Manufacture <u>Nipro</u> Model : <u>Diamax</u></p> <p>S. No. <u>16117266</u> Dept. <u>Dialysis unit</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : IMP error, conductivity issue, blood leak alarm

Action Taken : checked the equipment and found a way valve, De. sensor Type 2 assy, Blood leak sensor, Cond. sensor PC board, magnetic gear pump assy are defective. need to replace these spares for further checking.

Completed Date : 30/6/24 Time : 1:20 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ramdas</u>	<u>30/6/24</u>	<u>9:30 AM</u>	<u>1:20 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Ramdas
Signature :
Date : 30/6/24
Contact Number : 9745520163

Customer Name : Al...
Signature :
Date : 30/6/24
Contact Number : 8...
Designation : Staff
Hospital Seal :





Quotation

Quotation No.SW [433] 07-2024

Date:16-07-2024

To

Cyrix Healthcare Pvt Ltd,
Poonithura , Ernakulam

Dear Sir,

Thank you for considering us and the opportunity to submit a quotation. We are delighted to present our quote as follows.

Sl no	Name of the Items	Unit Price	Qty	Amount	GST	GST Amount	Total
1	0-401-114 2 WAY VALVE	₹ 15,662.50	1	₹ 15,662.50	12%	₹ 1,879.50	₹ 17,542.00
2	0-401-071 DP SENSOR TYPE 2 ASSY	₹ 46,850.00	1	₹ 46,850.00	12%	₹ 5,622.00	₹ 52,472.00
3	0-401-231 BLOOD LEAK SENSOR	₹ 99,127.50	1	₹ 99,127.50	12%	₹ 11,895.30	₹ 1,11,022.80
4	COND.SENSOR PC BOARD VER 4.03	₹ 34,500.00	1	₹ 34,500.00	12%	₹ 4,140.00	₹ 38,640.00
5	MAGNETIC GEAR PUMP ASSY	₹ 1,15,500.00	1	₹ 1,15,500.00	12%	₹ 13,860.00	₹ 1,29,360.00
							₹ 3,49,036.80

Terms & Conditions

- 100% advance payment along with your Purchase Order.
- Cheque in favor of: SmartWay Healthcare
A/c No: 0224073000005622
IFSC: SIBL0000224
Bank: South Indian Bank
Branch: Kalamassery

Thanking You

ayas

SmartWay Healthcare



GSTIN: 32AERFS9757B12M

Shop No: 21/497 | Ground Floor, Kaipadamugal
Thrikkakara North | Ernakulam - 682 021

+91 8921172300 | smartwayhealthcareekm@gmail.com



TAX INVOICE

Name of the Assessee or Supplier: Nipro Medical (India) Pvt Ltd
 Address of the Assessee or Supplier: 16-1-D, Anjikkath Arcade, Kaipadamugal, Vadakode P O,
 Seaport Airport Road Kalamassery, Umakulam, 682021 Cochin, Kerala India
 Ph: 0484-6062123 Fax:
 VAT/TIN No.: 32071279746
 CST/TIN No.: 32071279746C
 Service Tax No.:
 CIN No.: [U]5071N2006PTC070002
 Drug License: KL-1KM-110877
 E-mail ID: Niprocare@nipromedical.com

Tax Invoice No.: Coc/2017/580001240
 Dated: 22.06.2017
 Range:
 Division:
 Commissionerate:
 PAN No:
 Excise Reg No:
 ECC.No.:

Bill To Party:
 Party Code: 1303089_KSB CARE
 CHARITABLE TRUST
 C/o KSB PUMPS LIMITED, 126 MARKET
 CHAMBERS III, NARIMAN POINT, NARIMAN,
 MUMBAI, 400021, Maharashtra, India.
 Tel: 0480-2695766 Fax: Mob: 9446005167

VAT/TIN No. :
 CST/TIN No. :
 Range :
 Division :
 Commissionerate :
 PAN No :
 Excise Reg No :
 ECC.No. :
 Drug License :

Ship To Party:
 Party Code: 1303091_Community Health
 Centre
 Mula Govt Hospital, Thrissur Thrissur -KERALA
 680732, Kerala, India
 Tel: 0480 2695766 Fax: Mob: 09446005167

VAT/TIN No. :
 CST/TIN No. :
 Range :
 Division :
 Commissionerate :
 PAN No :
 Excise Reg No :
 ECC.No. :
 Drug License :

Contract No.: 4018948
 Despatch Doc No.: 5042274
 Despatch Date: 22.06.2017
 Despatched through: Truck
 Destination: NARIMAN, MUMBAI
 L.R No.:
 L.R Date: 22.06.2017
 Vehicle No.: OTHERS
 Customer P.O.No.: Email Date 16.6.2017
 P.O Date: 16.06.2017
 BOF No.: RAJEEV G
 L.C and /B.G No.:
 L.C and/B.G Date:
 No. of Boxes:
 Gross Weight: 110 000 KG
 Mode of Shipment: Others

Sr. No	Item Code	Description	Tariff Heading No	Batch No.	Expiry Date	Qty	UOM	Per Unit Rate	Amount (INR)
1	5100245	Dialysis Machines (Diamax)		16J11726E		1.00	EA	476,190.5000	476,190.50
Grand Total						1.00			476190.50

Stock entered to Dialysis Registers page
 no: 3 on 22/06/2017



Total Central Excise Duty Payable (In Words)- Rupee Zero And Paise Zero Only
 Date of Issue of Invoice : 22.06.2017 Time of Issue : 13:07:21
 Date of Removal of Invoice : 22.06.2017 Time of Removal : 13:07:21

Grand total (In Words) - Rupee Five Lakh And Paise Three Only
 Payment To:-
 Cheque/Demand Draft to be prepared in favour of "NIPRO MEDICAL INDIA PVT LTD."
 Remarks:

Total		: 476190.50
Excise Duty	%	:
Ed Cess	%	:
SE Cess	%	:
Sub Total		: 0.00
Freight	%	:
Discount		:
Sub Total		: 0
VAT/CST	5%	: 23809.53
Additional Tax	%	:
Service Tax	%	: 0
Grand Total		: 500000.03

Certified that the particular given above are true, correct and the amount indicated represents the price actually charged and that there is no additional consideration flowing directly or indirectly from the buyer.

Terms & Conditions:
 1. If due amount is not paid as per agreed payment terms, interest will be charged at 1.75% per month for delayed payment.
 2. Claim of any nature whatsoever will lapse unless raised in writing within 3 days from date of invoice.
 3. We reserve to ourselves the right to demand payment of this bill at any time before due date.
 4. Subject to Jurisdiction

VAT Declaration: I/We hereby certify that my/our registration under the Kerala Value Added Tax, 2002 is in force on the date on which the sale of the goods specify in this tax invoice is made by me/us and the transaction of sale covered by this tax invoice has been effected by me/us and it shall be accounted for in the number of sales while filing of return and the due tax, if any, payable on the sale has been or shall be paid.



For Nipro Medical India Pvt Ltd.
 Authorized Signatory



Single Patient Dialysis Machine

MODEL	DIAMAX
SERIAL No.	16J11726E
POWER	AC230V 50/60Hz 2.2kVA
CLASS	CLASS I
MANUFACTURER	NIPRO CORPORATION 3-9-3, Honjo-Nishi, Kita-ku, Osaka, Japan



IPX1



2018-06



NIPRO

NIPRO CORPORATION

Made In Japan

EMC

EN60601-1-2:2001

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0830274



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BMMP)



QR/2D MATRIX

