


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRISSUR
2	Name of Hospital	PHC VETILAPARA
3	Name of Equipment with Make, Model and Serial Number	Semi automatic biochemistry analyser Make - Robonik Model - Prietest touch S/N - AT0440315RBK
4	Equipment ID/ Barcode	0841705- #87542
5	Date of purchase/ Year of manufacture/ installation Date	Ticket master installation date - 22-04-2015 Actual installation date - 19-03-2015 (stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	27-12-2023
9	Action taken	Checked the equipment and found contrast board and primary board are defective. Spare Enquired to OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	64760
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report, EOL from OEM attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found contrast board and primary board are defective. Equipment installed on 19/03/2015 aged up to 08 years and 8 months, oem informed that this machine is obsolete model, end of life letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Obsolete model


 Signature of JC BM (NHM)




 Signature of

സെനിയർ മെഡിക്കൽ ഓഫീസർ (ജനറൽ)
 വെറ്റിലടാന പ്രാഥമികാരോഗ്യകേന്ദ്രം
 തൃശ്ശൂർ ജില്ല - 680 721



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
227046

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Phc vettilapuzha</u></p> <p>Address <u>Thirissar</u></p> <p>Ph :</p>	<p>Call Registration Date : <u>27/12/2023</u></p> <p>Caller ID : <u>87552</u></p> <p>Date of Visit : <u>28/12/2023</u></p> <p>Asset No. : <u>0841705</u></p> <p>EQPT Name : <u>Semi auto analyser</u></p> <p>Manufacture <u>Robonik</u> Model : <u>Prietest touch</u></p> <p>S. No. <u>AT0410315RBK</u> Dept. <u>Laboratory</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Semi auto analyser not working

Action Taken : checked the equipment and found contrast board and primary board are defective spare enguired to D.E.M.

Completed Date : 28/12/2023 Time : 3:30pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Suhas</u>	<u>28/12/2023</u>	<u>1:30pm</u>	<u>3:50pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Suhas. p Rafter</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>28/12/2023</u></p> <p>Contact Number : <u>7593847135</u></p>	<p>Customer Name : <u>Twinkle george</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>28/12/2023</u></p> <p>Contact Number : <u>9078409856</u></p> <p>Designation : <u>LAB TECHNICIAN</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--



തൃശ്ശൂർ ജില്ല - 680 721



Date : 12th January 2024

Dear Valued Customer,



Thank you for choosing **Robonik (India) Pvt Ltd**. We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model: Prietest Touch) having **Serial No. AT0440315RBK / AT0280315RBK**.

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model: Prietest Touch) having **Serial No. AT0440315RBK / AT0280315RBK** as **End of Life (EOL) Notified**. (Duration from installation is 08 Years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK INDIA PVT LTD**


 

AUTHORIZED SIGNATORY

Sir any one get 9:20 am ✓✓

No 9:51 am


➡ Forwarded



EOL - AT0440315RBK & AT0280315RBK.pdf
1 page · 393 kB · PDF

2:26 pm


➡ Forwarded




EOL - AT2301214RBK.pdf
1 page · 391 kB · PDF

2:26 pm

➡ Forwarded



EOL - AT0440315RBK & AT0280315RBK.pdf

 Message








Handwritten signature
Kerala Public Health Department
Kerala State Government
Ph: 680 724

**BIO CHEMISTRY SEMI AUTO
ANALYSER**



prifetest

TOUCH
Control Panel




Consilio Medical

CYRIX HEALTH CARE PVT.LTD
BEMP SERVICE PROVIDER
PREVENTIVE MAINTENANCE SERVICE
 PMS DONE ON: 30/12/23
 NEXT PMS DUE: 29/12/24
 PERFORMED BY: SOMAN

Preventive Maintenance Schedule (PMS)
 01 29/12/20
 02

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
0841705



CYRIX HEALTHCARE PVT.LTD
 ISO 9001:2008 Certified No. 14555073
 PMS Done On: 30/12/23
 Next PMS Due: 29/12/24
 Performed by: SOMAN

Consultant (Biomedical)
 Anand, Vinay