



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 232712

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

**Service Report**

Health Facility General Hospital  
Address Miruvananthapuram  
Kerala  
Ph : 9012896198

Call Registration Date : 21/6/24  
Caller ID : 118594  
Date of Visit : 5/8/24  
Asset No. : 0150924  
EQPT Name : Electrical Suction Unit  
Manufacture Lucky Surgical Model : Nil  
S. No. LSA/2022/603 Dept. Dialysis Unit

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Not working

Action Taken : Checked and found motor defective. Found that bottle not damaged and wheel rusted and tubings defective. Need to replace these spares for further checking and working condition of the equipment

Completed  Date : 5/8/24 Time : 2:00PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer

Soojye Sudheep

Date

5/8/24

Start Time

1:30PM

End Time

2:00PM

Customer Remark



Pending

Service Engineer Name : Soojye Sudheep  
Signature : Soojye  
Date : 5/8/24  
Contact Number : 7907406244

Customer Name : NITHIN  
Signature : Nithin  
Date : 05/08/24  
Contact Number : 9012995543  
Designation : Tech  
Hospital Seal :

Sub-Independent General Hospital Thiruvananthapuram




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Suction Apparatus Make : Lucky Surgicals Model : NA Serial No : NA
4	Equipment ID/ Barcode	118594/0150924
5	Date of purchase/ Year of manufacture/Installation Date	24/10/2017
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	21/06/2024(Toll free)
9	Action taken	Checked and found suction motor, jars, wheels rusted and tubings defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare with Vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Suction Motor – 4650/- Suction Jars–3000/- Wheels – 1200/- Suction tube set – 1100/- Total – 9950/-

13	Asset Value	11525/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	86.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND REFERENCE QUOTATION ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found suction motor and suction jars, wheels rusted and tubings defective. The unit is installed on 24/10/2017 and covered up to 4+ years. The repairing cost is 86.3%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	SOORIYA SUDHEEP 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Suction motor and jars defective. Verified the estimate. So can be RBER as per criteria

Call date : 24-06-2024  
 RBER submission date : 5-8-2024.  
 Late submission

**MANEESHA MOHAN M.R.**  
 Junior Consultant (Biomedical)  
 National Health Mission  
 Thiruvananthapuram

  
 5-8-2024  
 Signature of JC BM (NHM)

Date:




Signature of  
 Superintendent/Medical Officer (i/c)

Superintendent  
 General Hospital  
 Thiruvananthapuram



11.05.17	B/F Form P.No: <sup>128</sup> dt: 11.05.17 LPR GAT IN. NO: 03A Target Surgicals Rs: 10710/- Issued to ward 7 (SSR 28)	01	01	90%
24.10.17	B/F P.No: 225 dt 24.10.17 LPR GAT IN. NO: 40 Target Surgicals Rs: 11525/- Issued to FSW (SSR 99)	01	01	Full
7.3.19	KMSCL as per ord. no: SO/1117/2018-19/119, purchased from Monarch Surgical Industries, 62, DSIDC, Scheme-1, Phase II, Okhla Industrial area, New Delhi. Tel: 26389237, as per inv. no. MST-0064 dt: 06.2.2019 Installed on: 13.6.2019 Warranty from 13.6.2019 to 12/6/2023. Amount ₹. 9900 + tax/enc.			
	1. Suction Apparatus Issued Model RG-02 ISI Marked	4 pc	4 pc	
28.3.19	Issued to OT	2	2	
25.10.19	Issued to Ward 8 down (SSR 75)	1	1	Full
25.10.19	Issued to Ward I (SSR 205)	1	NIL	Full
10.02.2020	Invoice No: MST-00169 Supplied by Monarch Surgical Industries, Scheme-1, Phase II, Okhla Industrial Area, New Delhi - Purchase Order No: SO/1160/2019-2020/119 Model: RG-02 ISI Marked. Warranty start date: 4/3/2020 End date: 8/3/2021 Engineer: Komsal Singh 9717374579. Serial No: 2020029, 2020032			
4/3/2020	Used to Operator theatre General	2	2	
4/6/2020	Scheme: DHS 2019-20 Strengthening of emergency medical care intended in Sub block Phase Vol II	Signature: Shankar Singh Superintendent Okhla Hospital Gurgaon		



No. Page No 2020/119  
 Vol 69  
 Signature: Shankar Singh  
 Superintendent  
 Okhla Hospital  
 Gurgaon





9744473043 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333  
Room No: 22  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

## CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0  
State: 32 - Kerala

MSME NO.: UDYAM-KL-07--0030714  
DL NO.: KL/KTM/MD42/2023/000015

## Estimate

Estimate For:

### KASHYEP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-40

Date: 05/08/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION JAR 2 LTRS	90189099	2	Nos	₹ 1,339.29	₹ 321.43 (12%)	₹ 3,000.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
3	2" NYLON ICU COT WHEEL		4	Nos	₹ 300.00	₹ 0.00 (0%)	₹ 1,200.00
4	SUCTION TUBE SET 3 MTR		1	Nos	₹ 1,100.00	₹ 0.00 (0%)	₹ 1,100.00
Total						₹ 1,030.75	₹ 9,950.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 8,919.25
SGST@6%	₹ 160.71
CGST@6%	₹ 160.71
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 9,950.00
You Saved	₹ 0.01

Description

GH TRIVANDRUM  
EQUIPMENT: SUCTION MACHINE  
BAR CODE: 0150924  
TICKET ID: 118594

Estimate Amount In Words

Nine Thousand Nine Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



---

## SUCTION MACHINE Estimate invoice details

1 message

---

<cloud99biotech@gmail.com>

Mon, 5 Aug 2024 at 6:19 pm

To: KASYEPCYRIX@gmail.com

Dear KASHYEP,

Thanks for doing business with us. Below are your invoice details.

Also attached are the PDF of invoice for your convenience.


Estimate:

Amount: 9950

Thanks & Regards




**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)**  
**CYRIX**  
 HEALTHCARE SERVICES  
 TOTAL TECH. INC. 1992 - 988-1700  
 BAR CODE - (800)-800-6152  
**0150924**



**Suction Apparatus**  
 FILTER POWER ON/OFF FUSE FUSE  
 SACTION  
 N. CONTR.